

DIESEL RANGE ORGANICS CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Chemtech Contract: CAMP02
ProjectID: South River WM Replacement
Lab Code: CHEM Case No.: Q1982 SAS No.: Q1982 SDG No.: Q1982
DataFile: FF015835.D Analyst Name: YP\AJ Analyst Date: 05-13-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
500	59789393	119579	119463	0.097