

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB042325

Review By	yogesh	Review On	4/23/2025 1:56:12 PM
Supervise By	mohammad	Supervise On	4/26/2025 2:19:22 AM
SubDirectory	FB042325	HP Acquire Method	HP Processing Method
			FB042325
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24110,PP24485,PP24486,PP24487,PP24488,PP24489		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24111,PP24490		

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	5 GRO STD		FB031638.D	23 Apr 2025 12:11		YP/AJ	Ok
2	10 GRO STD		FB031639.D	23 Apr 2025 12:38		YP/AJ	Ok
3	20 GRO STD		FB031640.D	23 Apr 2025 13:05		YP/AJ	Ok
4	50 GRO STD		FB031641.D	23 Apr 2025 13:32		YP/AJ	Ok,M
5	100 GRO STD		FB031642.D	23 Apr 2025 14:00		YP/AJ	Ok,M
6	FB042325GROICV		FB031643.D	23 Apr 2025 14:39		YP/AJ	Ok,M

M : Manual Integration

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB051325

Review By	yogesh	Review On	5/13/2025 11:36:14 AM
Supervise By	mohammad	Supervise On	5/15/2025 3:42:45 AM
SubDirectory	FB051325	HP Acquire Method	HP Processing Method
			FB042325

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24110,PP24485,PP24486,PP24487,PP24488,PP24489
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24567,PP24568,PP24569,PP24570 PP24111,PP24490

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	20 PPB GRO STD		FB031704.D	13 May 2025 8:44		YP/AJ	Ok,M
2	VBF0513S1		FB031705.D	13 May 2025 9:24		YP/AJ	Ok
3	VBF0513S2		FB031706.D	13 May 2025 9:51		YP/AJ	Ok
4	BSF0513S1		FB031707.D	13 May 2025 10:19		YP/AJ	Ok
5	Q1982-02		FB031708.D	13 May 2025 10:45	Vial-A, Not Purged	YP/AJ	Not Ok
6	Q1982-03		FB031709.D	13 May 2025 11:26	Vial-A, Not Purged	YP/AJ	Not Ok
7	Q1982-04		FB031710.D	13 May 2025 11:54	Vial-A	YP/AJ	Ok,M
8	Q1982-05		FB031711.D	13 May 2025 12:21	Vial-A	YP/AJ	Ok
9	Q1982-06		FB031712.D	13 May 2025 12:59	Vial-A, Not Purged	YP/AJ	Not Ok
10	Q1982-07		FB031713.D	13 May 2025 13:27	Vial-A, Not Purged	YP/AJ	Not Ok
11	BSF0513S2		FB031714.D	13 May 2025 13:54		YP/AJ	Ok
12	20 PPB GRO STD		FB031715.D	13 May 2025 14:22		YP/AJ	Ok
13	Q1982-08		FB031716.D	13 May 2025 15:53	Vial-A	YP/AJ	Ok
14	Q2010-01		FB031717.D	13 May 2025 16:21	Vial-A	YP/AJ	Ok
15	Q2010-02		FB031718.D	13 May 2025 16:49	Vial-A, Not Purged	YP/AJ	Not Ok
16	Q2010-03		FB031719.D	13 May 2025 17:17	Vial-A, Not Purged	YP/AJ	Not Ok
17	Q2010-04		FB031720.D	13 May 2025 17:44	Vial-A	YP/AJ	Ok
18	Q1982-01		FB031721.D	13 May 2025 18:12	Vial-B, Surr Fail	YP/AJ	Ok

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB051325

Review By	yogesh	Review On	5/13/2025 11:36:14 AM
Supervise By	mohammad	Supervise On	5/15/2025 3:42:45 AM
SubDirectory	FB051325	HP Acquire Method	HP Processing Method
			FB042325
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24110,PP24485,PP24486,PP24487,PP24488,PP24489		
CCC Internal Standard/PEM	PP24567,PP24568,PP24569,PP24570		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24111,PP24490		

19	Q1982-02		FB031722.D	13 May 2025 18:40	Vial-B, Surrogate failed	YP/AJ	Ok, M
20	Q1982-03		FB031723.D	13 May 2025 19:07	Vial-B	YP/AJ	Ok
21	Q1982-06		FB031724.D	13 May 2025 19:35	Vial-B	YP/AJ	Ok
22	BSF0513S3		FB031725.D	13 May 2025 20:02		YP/AJ	Ok
23	20 PPB GRO STD		FB031726.D	13 May 2025 20:30		YP/AJ	Ok
24	Q1982-07		FB031727.D	13 May 2025 21:51	Vial-B, Not Purged	YP/AJ	Not Ok
25	Q1982-08		FB031728.D	13 May 2025 22:19	Not required	YP/AJ	Not Ok
26	Q2010-01		FB031729.D	13 May 2025 22:46	Not required	YP/AJ	Not Ok
27	Q2010-02		FB031730.D	13 May 2025 23:13	Vial-B, Not Purged	YP/AJ	Not Ok
28	Q2010-03		FB031731.D	13 May 2025 23:40	Vial-B	YP/AJ	Ok
29	Q2010-04		FB031732.D	14 May 2025 00:08	Not required	YP/AJ	Not Ok
30	BSF0513S4		FB031733.D	14 May 2025 00:35		YP/AJ	Ok
31	20 PPB GRO STD		FB031734.D	14 May 2025 2:24		YP/AJ	Ok

M : Manual Integration

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB051425

Review By	yogesh	Review On	5/14/2025 11:41:14 AM
Supervise By	mohammad	Supervise On	5/19/2025 2:35:24 AM
SubDirectory	FB051425	HP Acquire Method	HP Processing Method
			FB042325
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24110,PP24485,PP24486,PP24487,PP24488,PP24489		
CCC	PP24571,PP24572,PP24574		
Internal Standard/PEM			
ICV/I.BLK	PP24111,PP24490		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	20 PPB GRO STD		FB031735.D	14 May 2025 9:26		YP/AJ	Ok
2	VBF0514S1		FB031736.D	14 May 2025 10:36		YP/AJ	Ok
3	VBF0514S2		FB031737.D	14 May 2025 11:04		YP/AJ	Ok
4	BSF0514S1		FB031738.D	14 May 2025 11:31		YP/AJ	Ok,M
5	Q1982-07		FB031739.D	14 May 2025 11:58	Vial-C	YP/AJ	Ok
6	Q2010-02		FB031740.D	14 May 2025 12:26	Vial-C	YP/AJ	Ok
7	BSF0514S2		FB031741.D	14 May 2025 12:53		YP/AJ	Ok
8	20 PPB GRO STD		FB031742.D	14 May 2025 13:20		YP/AJ	Ok
9	VBF0514W1		FB031743.D	14 May 2025 14:13		YP/AJ	Ok
10	BSF0514W1		FB031744.D	14 May 2025 14:40		YP/AJ	Ok
11	Q2032-11		FB031745.D	14 May 2025 15:07	Vial-A	YP/AJ	Ok
12	BSF0514W2		FB031746.D	14 May 2025 15:35		YP/AJ	Ok,M
13	20 PPB GRO STD		FB031747.D	14 May 2025 16:02		YP/AJ	Ok

M : Manual Integration