

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1982 SAS No.: Q1982 SDG NO.: Q1982

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL01 Date Analyzed: 05/15/2025

Lab Sample No.: HSTDCCC750 Data File : PS030224.D Time Analyzed: 10:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.132	9.033	9.233	669.850	712.500	-6.0
2,4,5-TP(Silvex)	8.851	8.750	8.950	671.550	712.500	-5.7
2,4-D	8.010	7.910	8.110	666.560	705.000	-5.5
2,4-DB	9.690	9.589	9.789	698.480	712.500	-2.0
2,4-DCAA	6.934	6.834	7.034	694.260	750.000	-7.4
DICAMBA	7.109	7.010	7.210	661.320	705.000	-6.2
DICHLORPROP	7.790	7.689	7.889	650.860	705.000	-7.7
Dinoseb	10.851	10.751	10.951	654.470	705.000	-7.2

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1982 SAS No.: Q1982 SDG NO.: Q1982

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL01 Date Analyzed: 05/15/2025

Lab Sample No.: HSTDCCC750 Data File : PS030224.D Time Analyzed: 10:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.923	9.826	10.026	669.280	712.500	-6.1
2,4,5-TP(Silvex)	9.519	9.423	9.623	674.040	712.500	-5.4
2,4-D	8.641	8.544	8.744	668.710	705.000	-5.1
2,4-DB	10.481	10.385	10.585	633.760	712.500	-11.1
2,4-DCAA	7.451	7.354	7.554	698.470	750.000	-6.9
DICAMBA	7.636	7.539	7.739	667.200	705.000	-5.4
DICHLORPROP	8.328	8.232	8.432	648.370	705.000	-8.0
Dinoseb	10.854	10.758	10.958	655.130	705.000	-7.1