

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1982 SAS No.: Q1982 SDG NO.: Q1982

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL04 Date Analyzed: 05/19/2025

Lab Sample No.: HSTDCCC750 Data File : PS030276.D Time Analyzed: 16:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.130	9.033	9.233	740.040	712.500	3.9
2,4,5-TP(Silvex)	8.849	8.750	8.950	734.250	712.500	3.1
2,4-D	8.008	7.910	8.110	722.650	705.000	2.5
2,4-DB	9.688	9.589	9.789	792.670	712.500	11.3
2,4-DCAA	6.932	6.834	7.034	761.130	750.000	1.5
DICAMBA	7.108	7.010	7.210	724.110	705.000	2.7
DICHLORPROP	7.788	7.689	7.889	713.170	705.000	1.2
Dinoseb	10.849	10.751	10.951	718.030	705.000	1.8

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1982 SAS No.: Q1982 SDG NO.: Q1982

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL04 Date Analyzed: 05/19/2025

Lab Sample No.: HSTDCCC750 Data File : PS030276.D Time Analyzed: 16:37

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.922	9.826	10.026	735.560	712.500	3.2
2,4,5-TP(Silvex)	9.518	9.423	9.623	740.370	712.500	3.9
2,4-D	8.641	8.544	8.744	720.730	705.000	2.2
2,4-DB	10.481	10.385	10.585	672.490	712.500	-5.6
2,4-DCAA	7.451	7.354	7.554	751.730	750.000	0.2
DICAMBA	7.636	7.539	7.739	729.140	705.000	3.4
DICHLORPROP	8.329	8.232	8.432	702.130	705.000	-0.4
Dinoseb	10.854	10.758	10.958	727.930	705.000	3.3