

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q1982</u>	Analytical Method:	<u>8151A</u>
Client:	<u>CDM Smith</u>	DataFile :	<u>PS030238.D</u>

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q1982-08MSD (Column 1)	TP-9MSD											
	DICAMBA	204.5	0	174	ug/Kg	88		16		10	112	20
	DICHLORPROP	204.5	0	198	ug/Kg	97		14		10	113	20
	2,4-D	204.5	0	207	ug/Kg	101		14		10	144	20
	2,4,5-TP(Silvex)	204.5	0	182	ug/Kg	96		16		10	114	20
	2,4,5-T	204.5	0	175	ug/Kg	92		14		10	115	20
	2,4-DB	204.5	0	145	ug/Kg	91		17		10	140	20
Client Sample ID: Q1982-08MSD (Column 2)	Dinoseb	204.5	0	0	ug/Kg	52		21	*	10	118	20
	TP-9MSD											
	DICAMBA	204.5	0	180	ug/Kg	0		16		10	112	20
	DICHLORPROP	204.5	0	195	ug/Kg	0		14		10	113	20
	2,4-D	204.5	0	204	ug/Kg	0		14		10	144	20
	2,4,5-TP(Silvex)	204.5	0	196	ug/Kg	0		16		10	114	20
	2,4,5-T	204.5	0	189	ug/Kg	0		14		10	115	20
	2,4-DB	204.5	0	187	ug/Kg	0		17		10	140	20
	Dinoseb	204.5	0	107	ug/Kg	0		21	*	10	118	20