

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1982

Analytical Method: 8151A

Client: CDM Smith

Datafile : PS030280.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits	
								Qual	Low	High
PB167996BS (Column 1)	DICAMBA	166.5	173	ug/Kg	104				72	129
	DICHLORPROP	166.5	172	ug/Kg	103				77	135
	2,4-D	166.5	174	ug/Kg	105				65	144
	2,4,5-TP(Silvex)	166.5	177	ug/Kg	106				74	146
	2,4,5-T	166.5	178	ug/Kg	107				77	134
	2,4-DB	166.5	188	ug/Kg	113				72	122
	Dinoseb	166.5	175	ug/Kg	105				74	132
PB167996BS (Column 2)	DICAMBA	166.5	169	ug/Kg	102				72	129
	DICHLORPROP	166.5	167	ug/Kg	100				77	135
	2,4-D	166.5	172	ug/Kg	103				65	144
	2,4,5-TP(Silvex)	166.5	175	ug/Kg	105				74	146
	2,4,5-T	166.5	174	ug/Kg	105				77	134
	2,4-DB	166.5	152	ug/Kg	91				72	122
	Dinoseb	166.5	171	ug/Kg	103				74	132