

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1982 SAS No.: Q1982 SDG NO.: Q1982

Continuing Calib Date: 05/13/2025 Initial Calibration Date(s): 05/02/2025 05/02/2025

Continuing Calib Time: 09:33 Initial Calibration Time(s): 09:45 18:29

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.78	4.78	4.68	4.88	0.00
Aroclor-1242-2	(2)	4.80	4.80	4.70	4.90	0.00
Aroclor-1242-3	(3)	4.85	4.85	4.75	4.95	0.00
Aroclor-1242-4	(4)	4.97	4.97	4.87	5.07	0.00
Aroclor-1242-5	(5)	5.63	5.62	5.52	5.72	-0.01
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.73	8.73	8.63	8.83	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1982 SAS No.: Q1982 SDG NO.: Q1982

Continuing Calib Date: 05/13/2025 Initial Calibration Date(s): 05/02/2025 05/02/2025

Continuing Calib Time: 09:33 Initial Calibration Time(s): 09:45 18:29

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.76	4.76	4.66	4.86	0.00
Aroclor-1242-2	(2)	4.78	4.78	4.68	4.88	0.00
Aroclor-1242-3	(3)	4.95	4.96	4.86	5.06	0.01
Aroclor-1242-4	(4)	5.04	5.04	4.94	5.14	0.00
Aroclor-1242-5	(5)	5.56	5.56	5.46	5.66	0.00
Tetrachloro-m-xylene		3.68	3.68	3.58	3.78	0.00
Decachlorobiphenyl		8.68	8.68	8.58	8.78	0.00