

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO050225

Review By	yogesh	Review On	5/2/2025 11:26:03 AM
Supervise By	mohammad	Supervise On	5/6/2025 4:49:19 AM
SubDirectory	PO050225	HP Acquire Method	HP Processing Method PO050225

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO110890.D	02 May 2025 09:08		YP/AJ	Ok
2	I.BLK	I.BLK	PO110891.D	02 May 2025 09:27		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO110892.D	02 May 2025 09:45		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO110893.D	02 May 2025 10:04		YP/AJ	Ok
5	AR1660ICC500	AR1660ICC500	PO110894.D	02 May 2025 10:21		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO110895.D	02 May 2025 10:39		YP/AJ	Ok
7	AR1660ICC050	AR1660ICC050	PO110896.D	02 May 2025 10:58		YP/AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PO110897.D	02 May 2025 11:16		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO110898.D	02 May 2025 11:34		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO110899.D	02 May 2025 11:53		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO110900.D	02 May 2025 12:11		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO110901.D	02 May 2025 12:29		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO110902.D	02 May 2025 12:48		YP/AJ	Ok
14	AR1242ICC050	AR1242ICC050	PO110903.D	02 May 2025 13:06		YP/AJ	Ok
15	AR1248ICC1000	AR1248ICC1000	PO110904.D	02 May 2025 13:25		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO110905.D	02 May 2025 13:43		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO110906.D	02 May 2025 14:00		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO110907.D	02 May 2025 14:35		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO050225

Review By	yogesh	Review On	5/2/2025 11:26:03 AM
Supervise By	mohammad	Supervise On	5/6/2025 4:49:19 AM
SubDirectory	PO050225	HP Acquire Method	HP Processing Method PO050225
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		

19	AR1248ICC050	AR1248ICC050	PO110908.D	02 May 2025 14:53		YP/AJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PO110909.D	02 May 2025 15:12		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO110910.D	02 May 2025 15:30		YP/AJ	Ok
22	AR1254ICC500	AR1254ICC500	PO110911.D	02 May 2025 15:48		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO110912.D	02 May 2025 16:07		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO110913.D	02 May 2025 16:25		YP/AJ	Ok
25	AR1262ICC500	AR1262ICC500	PO110914.D	02 May 2025 17:01		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO110915.D	02 May 2025 17:19		YP/AJ	Ok
27	AR1268ICC750	AR1268ICC750	PO110916.D	02 May 2025 17:38		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO110917.D	02 May 2025 17:55		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO110918.D	02 May 2025 18:12		YP/AJ	Ok
30	AR1268ICC050	AR1268ICC050	PO110919.D	02 May 2025 18:29		YP/AJ	Ok,M
31	PO050225ICV500	ICVPO050225	PO110920.D	02 May 2025 19:41		YP/AJ	Ok
32	AR1242ICV500	ICVPO050225AR1242	PO110921.D	02 May 2025 20:18		YP/AJ	Ok
33	AR1248ICV500	ICVPO050225AR1248	PO110922.D	02 May 2025 20:54		YP/AJ	Ok
34	AR1254ICV500	ICVPO050225AR1254	PO110923.D	02 May 2025 21:31		YP/AJ	Ok
35	AR1268ICV500	ICVPO050225AR1268	PO110924.D	02 May 2025 22:08		YP/AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO051325

Review By	yogesh	Review On	5/13/2025 10:46:19 AM
Supervise By	mohammad	Supervise On	5/15/2025 3:42:21 AM
SubDirectory	PO051325	HP Acquire Method	HP Processing Method PO050225

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO111025.D	13 May 2025 08:31		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO111026.D	13 May 2025 09:08		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO111027.D	13 May 2025 09:33	DCB high in 2nd column	YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO111028.D	13 May 2025 09:51	AR1248 high in 2nd column , TCMX high in 2nd column	YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO111029.D	13 May 2025 10:10		YP/AJ	Ok,M
6	I.BLK	I.BLK	PO111030.D	13 May 2025 10:28		YP/AJ	Ok
7	Q1984-09	OU4-TS-25-050725	PO111031.D	13 May 2025 10:46		YP/AJ	Ok
8	PB167972BL	PB167972BL	PO111032.D	13 May 2025 13:24		YP/AJ	Ok
9	PB167972BS	PB167972BS	PO111033.D	13 May 2025 13:42		YP/AJ	Ok
10	Q1982-01	TP-1	PO111034.D	13 May 2025 14:01		YP/AJ	Ok
11	Q1982-02	TP-2B	PO111035.D	13 May 2025 14:19		YP/AJ	Ok
12	Q1982-03	TP-3	PO111036.D	13 May 2025 14:37		YP/AJ	Ok
13	Q2017-01	MH-I	PO111037.D	13 May 2025 14:56		YP/AJ	Ok
14	Q2017-01MS	MH-IMS	PO111038.D	13 May 2025 15:14		YP/AJ	Ok
15	Q2017-01MSD	MH-IMSD	PO111039.D	13 May 2025 15:32		YP/AJ	Ok
16	Q2017-05	MH-J	PO111040.D	13 May 2025 15:51		YP/AJ	Ok
17	AR1660CCC500	AR1660CCC500	PO111041.D	13 May 2025 17:33		YP/AJ	Ok,M

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO051325

Review By	yogesh	Review On	5/13/2025 10:46:19 AM
Supervise By	mohammad	Supervise On	5/15/2025 3:42:21 AM
SubDirectory	PO051325	HP Acquire Method	HP Processing Method PO050225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1242CCC500	AR1242CCC500	PO111042.D	13 May 2025 18:28	DCB high in 2nd column	YP/AJ	Ok,M
19	AR1248CCC500	AR1248CCC500	PO111043.D	13 May 2025 18:46		YP/AJ	Ok
20	AR1254CCC500	AR1254CCC500	PO111044.D	13 May 2025 19:04		YP/AJ	Ok,M
21	I.BLK	I.BLK	PO111045.D	13 May 2025 19:23		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042225

Review By	yogesh	Review On	4/22/2025 1:34:56 PM
Supervise By	mohammad	Supervise On	4/24/2025 6:33:42 AM
SubDirectory	PP042225	HP Acquire Method	HP Processing Method PP042225

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP071387.D	22 Apr 2025 09:56		YP\AJ	Ok
2	I.BLK	I.BLK	PP071388.D	22 Apr 2025 10:13		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP071389.D	22 Apr 2025 10:29		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP071390.D	22 Apr 2025 10:45		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP071391.D	22 Apr 2025 11:02		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP071392.D	22 Apr 2025 11:18		YP\AJ	Ok
7	AR1660ICC050	AR1660ICC050	PP071393.D	22 Apr 2025 11:34		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP071394.D	22 Apr 2025 11:51		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PP071395.D	22 Apr 2025 12:07		YP\AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PP071396.D	22 Apr 2025 12:23		YP\AJ	Ok
11	AR1242ICC750	AR1242ICC750	PP071397.D	22 Apr 2025 12:39		YP\AJ	Ok
12	AR1242ICC500	AR1242ICC500	PP071398.D	22 Apr 2025 12:56		YP\AJ	Ok
13	AR1242ICC250	AR1242ICC250	PP071399.D	22 Apr 2025 13:12		YP\AJ	Ok
14	AR1242ICC050	AR1242ICC050	PP071400.D	22 Apr 2025 13:28		YP\AJ	Ok
15	AR1248ICC1000	AR1248ICC1000	PP071401.D	22 Apr 2025 13:45		YP\AJ	Ok
16	AR1248ICC750	AR1248ICC750	PP071402.D	22 Apr 2025 14:01		YP\AJ	Ok
17	AR1248ICC500	AR1248ICC500	PP071403.D	22 Apr 2025 14:17		YP\AJ	Ok
18	AR1248ICC250	AR1248ICC250	PP071404.D	22 Apr 2025 14:33		YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042225

Review By	yogesh	Review On	4/22/2025 1:34:56 PM
Supervise By	mohammad	Supervise On	4/24/2025 6:33:42 AM
SubDirectory	PP042225	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PP071405.D	22 Apr 2025 14:50		YPIAJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PP071406.D	22 Apr 2025 15:06		YPIAJ	Ok
21	AR1254ICC750	AR1254ICC750	PP071407.D	22 Apr 2025 15:22		YPIAJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PP071408.D	22 Apr 2025 15:38		YPIAJ	Ok
23	AR1254ICC250	AR1254ICC250	PP071409.D	22 Apr 2025 15:55		YPIAJ	Ok
24	AR1254ICC050	AR1254ICC050	PP071410.D	22 Apr 2025 16:11		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP071411.D	22 Apr 2025 16:27		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP071412.D	22 Apr 2025 16:44		YPIAJ	Ok,M
27	AR1268ICC750	AR1268ICC750	PP071413.D	22 Apr 2025 17:00		YPIAJ	Ok,M
28	AR1268ICC500	AR1268ICC500	PP071414.D	22 Apr 2025 17:16		YPIAJ	Ok,M
29	AR1268ICC250	AR1268ICC250	PP071415.D	22 Apr 2025 17:33		YPIAJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PP071416.D	22 Apr 2025 17:49		YPIAJ	Ok,M
31	PP042225ICV500	ICVPP042225	PP071417.D	22 Apr 2025 18:05		YPIAJ	Ok
32	AR1242ICV500	ICVPP042225AR1242	PP071418.D	22 Apr 2025 18:38		YPIAJ	Ok
33	AR1248ICV500	ICVPP042225AR1248	PP071419.D	22 Apr 2025 19:27		YPIAJ	Ok
34	AR1254ICV500	ICVPP042225AR1254	PP071420.D	22 Apr 2025 19:43		YPIAJ	Ok
35	AR1268ICV500	ICVPP042225AR1268	PP071421.D	22 Apr 2025 20:16		YPIAJ	Ok,M

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051325

Review By	yogesh	Review On	5/13/2025 10:45:34 AM
Supervise By	mohammad	Supervise On	5/16/2025 1:54:35 AM
SubDirectory	PP051325	HP Acquire Method	HP Processing Method PP042225

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP072002.D	13 May 2025 08:29		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP072003.D	13 May 2025 08:46	DCB high in 2nd column	YPIAJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PP072004.D	13 May 2025 09:02		YPIAJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP072005.D	13 May 2025 09:19	DCB high in 2nd column	YPIAJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP072006.D	13 May 2025 09:36	DCB high in 2nd column	YPIAJ	Ok,M
6	I.BLK	I.BLK	PP072007.D	13 May 2025 09:52		YPIAJ	Ok,M
7	Q1984-13	OU4-TS-27-050725	PP072008.D	13 May 2025 10:08		YPIAJ	Ok,M
8	Q1982-04	TP-4	PP072009.D	13 May 2025 13:02		YPIAJ	Ok,M
9	Q1982-05	TP-5	PP072010.D	13 May 2025 13:18		YPIAJ	Ok,M
10	Q1982-06	TP-6	PP072011.D	13 May 2025 13:35		YPIAJ	Ok,M
11	Q1982-07	TP-8	PP072012.D	13 May 2025 13:51		YPIAJ	Ok,M
12	Q1982-08	TP-9	PP072013.D	13 May 2025 14:07		YPIAJ	Ok,M
13	Q1984-15	OU4-TS-28-050725	PP072014.D	13 May 2025 14:23	DCB Low in both column	YPIAJ	ReRun
14	Q2001-02	WC-A4-03-C	PP072015.D	13 May 2025 14:40		YPIAJ	Ok,M
15	Q2001-06	WC-A1-05-C	PP072016.D	13 May 2025 14:56		YPIAJ	Ok,M
16	Q2001-10	WC-A1-06-C	PP072017.D	13 May 2025 15:13		YPIAJ	Ok,M
17	AR1660CCC500	AR1660CCC500	PP072018.D	13 May 2025 16:35		YPIAJ	Ok,M
18	AR1242CCC500	AR1242CCC500	PP072019.D	13 May 2025 17:24		YPIAJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051325

Review By	yogesh	Review On	5/13/2025 10:45:34 AM
Supervise By	mohammad	Supervise On	5/16/2025 1:54:35 AM
SubDirectory	PP051325	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PP072020.D	13 May 2025 17:40		YPIAJ	Ok,M
20	AR1254CCC500	AR1254CCC500	PP072021.D	13 May 2025 17:56		YPIAJ	Ok,M
21	I.BLK	I.BLK	PP072022.D	13 May 2025 18:13		YPIAJ	Ok,M
22	Q2001-14	WC-A1-07-C	PP072023.D	13 May 2025 18:29		YPIAJ	Ok,M
23	Q2014-01	MOO-25-0134	PP072024.D	13 May 2025 18:45	DCB high in 1st column	YPIAJ	Ok,M
24	Q2014-03	MOO-25-0145	PP072025.D	13 May 2025 19:01		YPIAJ	Ok,M
25	Q2014-05	MOO-25-0148	PP072026.D	13 May 2025 19:18	F Flag in TCMX in 2nd column	YPIAJ	Not Ok
26	AR1660CCC500	AR1660CCC500	PP072027.D	13 May 2025 20:39	DCB high in 2nd column	YPIAJ	Ok,M
27	AR1242CCC500	AR1242CCC500	PP072028.D	13 May 2025 21:28	DCB high in 2nd column	YPIAJ	Ok,M
28	AR1248CCC500	AR1248CCC500	PP072029.D	13 May 2025 21:45	DCB high in 2nd column	YPIAJ	Ok,M
29	AR1254CCC500	AR1254CCC500	PP072030.D	13 May 2025 22:01		YPIAJ	Ok,M
30	I.BLK	I.BLK	PP072031.D	13 May 2025 22:17		YPIAJ	Ok,M

M : Manual Integration