

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1982 SAS No.: Q1982 SDG NO.: Q1982

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL01 Date Analyzed: 05/12/2025

Lab Sample No.: PSTDCCC050 Data File : PD088464.D Time Analyzed: 13:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.713	6.606	6.806	51.220	50.000	2.4
4,4'-DDE	6.204	6.097	6.297	50.250	50.000	0.5
4,4'-DDT	7.029	6.922	7.122	51.370	50.000	2.7
Aldrin	5.280	5.173	5.373	50.870	50.000	1.7
alpha-BHC	4.007	3.901	4.101	50.980	50.000	2.0
alpha-Chlordane	6.035	5.929	6.129	51.070	50.000	2.1
beta-BHC	4.523	4.416	4.616	49.630	50.000	-0.7
Decachlorobiphenyl	9.082	8.975	9.175	48.460	50.000	-3.1
delta-BHC	4.772	4.665	4.865	50.220	50.000	0.4
Dieldrin	6.355	6.249	6.449	51.520	50.000	3.0
Endosulfan I	6.083	5.976	6.176	51.000	50.000	2.0
Endosulfan II	6.794	6.687	6.887	49.630	50.000	-0.7
Endosulfan sulfate	7.157	7.050	7.250	50.840	50.000	1.7
Endrin	6.582	6.476	6.676	50.670	50.000	1.3
Endrin aldehyde	6.922	6.816	7.016	49.550	50.000	-0.9
Endrin ketone	7.637	7.531	7.731	50.860	50.000	1.7
gamma-BHC (Lindane)	4.338	4.232	4.432	49.850	50.000	-0.3
gamma-Chlordane	5.954	5.847	6.047	51.220	50.000	2.4
Heptachlor	4.938	4.831	5.031	51.810	50.000	3.6
Heptachlor epoxide	5.699	5.592	5.792	50.870	50.000	1.7
Methoxychlor	7.501	7.395	7.595	50.660	50.000	1.3
Tetrachloro-m-xylene	3.558	3.452	3.652	48.550	50.000	-2.9

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1982 SAS No.: Q1982 SDG NO.: Q1982

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL01 Date Analyzed: 05/12/2025

Lab Sample No.: PSTDCCC050 Data File : PD088464.D Time Analyzed: 13:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.932	5.834	6.034	43.530	50.000	-12.9
4,4'-DDE	5.377	5.280	5.480	44.480	50.000	-11.0
4,4'-DDT	6.186	6.088	6.288	44.750	50.000	-10.5
Aldrin	4.370	4.273	4.473	44.510	50.000	-11.0
alpha-BHC	3.394	3.296	3.496	44.150	50.000	-11.7
alpha-Chlordane	5.192	5.094	5.294	44.190	50.000	-11.6
beta-BHC	4.026	3.928	4.128	44.270	50.000	-11.5
Decachlorobiphenyl	8.075	7.977	8.177	42.270	50.000	-15.5
delta-BHC	4.263	4.165	4.365	44.220	50.000	-11.6
Dieldrin	5.515	5.417	5.617	43.880	50.000	-12.2
Endosulfan I	5.249	5.151	5.351	43.800	50.000	-12.4
Endosulfan II	6.083	5.984	6.184	43.930	50.000	-12.1
Endosulfan sulfate	6.484	6.386	6.586	43.640	50.000	-12.7
Endrin	5.791	5.693	5.893	42.740	50.000	-14.5
Endrin aldehyde	6.261	6.163	6.363	42.610	50.000	-14.8
Endrin ketone	6.994	6.895	7.095	43.520	50.000	-13.0
gamma-BHC (Lindane)	3.730	3.633	3.833	43.630	50.000	-12.7
gamma-Chlordane	5.128	5.030	5.230	44.250	50.000	-11.5
Heptachlor	4.084	3.986	4.186	42.940	50.000	-14.1
Heptachlor epoxide	4.874	4.777	4.977	44.120	50.000	-11.8
Methoxychlor	6.757	6.659	6.859	40.810	50.000	-18.4
Tetrachloro-m-xylene	2.881	2.783	2.983	43.390	50.000	-13.2