

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1982 SAS No.: Q1982 SDG NO.: Q1982

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL02 Date Analyzed: 05/12/2025

Lab Sample No.: PSTDCCC050 Data File : PD088475.D Time Analyzed: 17:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.705	6.606	6.806	51.540	50.000	3.1
4,4'-DDE	6.196	6.097	6.297	50.920	50.000	1.8
4,4'-DDT	7.021	6.922	7.122	51.670	50.000	3.3
Aldrin	5.272	5.173	5.373	52.260	50.000	4.5
alpha-BHC	4.000	3.901	4.101	51.650	50.000	3.3
alpha-Chlordane	6.028	5.929	6.129	52.350	50.000	4.7
beta-BHC	4.516	4.416	4.616	51.010	50.000	2.0
Decachlorobiphenyl	9.073	8.975	9.175	48.050	50.000	-3.9
delta-BHC	4.764	4.665	4.865	51.550	50.000	3.1
Dieldrin	6.347	6.249	6.449	52.270	50.000	4.5
Endosulfan I	6.075	5.976	6.176	52.110	50.000	4.2
Endosulfan II	6.786	6.687	6.887	50.230	50.000	0.5
Endosulfan sulfate	7.150	7.050	7.250	51.110	50.000	2.2
Endrin	6.575	6.476	6.676	50.990	50.000	2.0
Endrin aldehyde	6.915	6.816	7.016	49.870	50.000	-0.3
Endrin ketone	7.630	7.531	7.731	51.250	50.000	2.5
gamma-BHC (Lindane)	4.331	4.232	4.432	51.150	50.000	2.3
gamma-Chlordane	5.946	5.847	6.047	52.840	50.000	5.7
Heptachlor	4.930	4.831	5.031	52.960	50.000	5.9
Heptachlor epoxide	5.692	5.592	5.792	51.650	50.000	3.3
Methoxychlor	7.494	7.395	7.595	50.650	50.000	1.3
Tetrachloro-m-xylene	3.551	3.452	3.652	49.440	50.000	-1.1

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1982 SAS No.: Q1982 SDG NO.: Q1982

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL02 Date Analyzed: 05/12/2025

Lab Sample No.: PSTDCCC050 Data File : PD088475.D Time Analyzed: 17:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.932	5.834	6.034	44.970	50.000	-10.1
4,4'-DDE	5.377	5.280	5.480	45.840	50.000	-8.3
4,4'-DDT	6.186	6.088	6.288	45.760	50.000	-8.5
Aldrin	4.371	4.273	4.473	46.200	50.000	-7.6
alpha-BHC	3.395	3.296	3.496	45.490	50.000	-9.0
alpha-Chlordane	5.193	5.094	5.294	45.100	50.000	-9.8
beta-BHC	4.027	3.928	4.128	45.860	50.000	-8.3
Decachlorobiphenyl	8.075	7.977	8.177	41.060	50.000	-17.9
delta-BHC	4.264	4.165	4.365	45.690	50.000	-8.6
Dieldrin	5.515	5.417	5.617	45.230	50.000	-9.5
Endosulfan I	5.249	5.151	5.351	44.100	50.000	-11.8
Endosulfan II	6.083	5.984	6.184	44.610	50.000	-10.8
Endosulfan sulfate	6.484	6.386	6.586	43.910	50.000	-12.2
Endrin	5.791	5.693	5.893	44.180	50.000	-11.6
Endrin aldehyde	6.260	6.163	6.363	43.450	50.000	-13.1
Endrin ketone	6.994	6.895	7.095	43.780	50.000	-12.4
gamma-BHC (Lindane)	3.732	3.633	3.833	45.060	50.000	-9.9
gamma-Chlordane	5.128	5.030	5.230	45.140	50.000	-9.7
Heptachlor	4.085	3.986	4.186	45.340	50.000	-9.3
Heptachlor epoxide	4.875	4.777	4.977	45.290	50.000	-9.4
Methoxychlor	6.757	6.659	6.859	43.610	50.000	-12.8
Tetrachloro-m-xylene	2.882	2.783	2.983	44.840	50.000	-10.3