

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1982 SAS No.: Q1982 SDG NO.: Q1982

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL03 Date Analyzed: 05/12/2025

Lab Sample No.: PSTDCCC050 Data File : PD088484.D Time Analyzed: 19:35

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| 4,4'-DDD             | 6.705 | 6.606     | 6.806 | 52.060             | 50.000            | 4.1  |
| 4,4'-DDE             | 6.196 | 6.097     | 6.297 | 50.920             | 50.000            | 1.8  |
| 4,4'-DDT             | 7.021 | 6.922     | 7.122 | 51.630             | 50.000            | 3.3  |
| Aldrin               | 5.272 | 5.173     | 5.373 | 52.200             | 50.000            | 4.4  |
| alpha-BHC            | 4.000 | 3.901     | 4.101 | 51.480             | 50.000            | 3.0  |
| alpha-Chlordane      | 6.028 | 5.929     | 6.129 | 52.580             | 50.000            | 5.2  |
| beta-BHC             | 4.516 | 4.416     | 4.616 | 50.890             | 50.000            | 1.8  |
| Decachlorobiphenyl   | 9.073 | 8.975     | 9.175 | 48.550             | 50.000            | -2.9 |
| delta-BHC            | 4.764 | 4.665     | 4.865 | 51.480             | 50.000            | 3.0  |
| Dieldrin             | 6.348 | 6.249     | 6.449 | 52.570             | 50.000            | 5.1  |
| Endosulfan I         | 6.075 | 5.976     | 6.176 | 52.310             | 50.000            | 4.6  |
| Endosulfan II        | 6.786 | 6.687     | 6.887 | 50.510             | 50.000            | 1.0  |
| Endosulfan sulfate   | 7.149 | 7.050     | 7.250 | 51.510             | 50.000            | 3.0  |
| Endrin               | 6.575 | 6.476     | 6.676 | 50.660             | 50.000            | 1.3  |
| Endrin aldehyde      | 6.915 | 6.816     | 7.016 | 50.390             | 50.000            | 0.8  |
| Endrin ketone        | 7.630 | 7.531     | 7.731 | 51.460             | 50.000            | 2.9  |
| gamma-BHC (Lindane)  | 4.331 | 4.232     | 4.432 | 51.000             | 50.000            | 2.0  |
| gamma-Chlordane      | 5.946 | 5.847     | 6.047 | 53.040             | 50.000            | 6.1  |
| Heptachlor           | 4.931 | 4.831     | 5.031 | 52.610             | 50.000            | 5.2  |
| Heptachlor epoxide   | 5.692 | 5.592     | 5.792 | 51.820             | 50.000            | 3.6  |
| Methoxychlor         | 7.493 | 7.395     | 7.595 | 50.480             | 50.000            | 1.0  |
| Tetrachloro-m-xylene | 3.551 | 3.452     | 3.652 | 49.460             | 50.000            | -1.1 |

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1982 SAS No.: Q1982 SDG NO.: Q1982

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL03 Date Analyzed: 05/12/2025

Lab Sample No.: PSTDCCC050 Data File : PD088484.D Time Analyzed: 19:35

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| 4,4'-DDD             | 5.931 | 5.834     | 6.034 | 45.140             | 50.000            | -9.7  |
| 4,4'-DDE             | 5.377 | 5.280     | 5.480 | 45.770             | 50.000            | -8.5  |
| 4,4'-DDT             | 6.185 | 6.088     | 6.288 | 45.360             | 50.000            | -9.3  |
| Aldrin               | 4.371 | 4.273     | 4.473 | 45.780             | 50.000            | -8.4  |
| alpha-BHC            | 3.395 | 3.296     | 3.496 | 45.270             | 50.000            | -9.5  |
| alpha-Chlordane      | 5.192 | 5.094     | 5.294 | 44.840             | 50.000            | -10.3 |
| beta-BHC             | 4.027 | 3.928     | 4.128 | 45.380             | 50.000            | -9.2  |
| Decachlorobiphenyl   | 8.074 | 7.977     | 8.177 | 41.590             | 50.000            | -16.8 |
| delta-BHC            | 4.264 | 4.165     | 4.365 | 45.350             | 50.000            | -9.3  |
| Dieldrin             | 5.515 | 5.417     | 5.617 | 45.110             | 50.000            | -9.8  |
| Endosulfan I         | 5.249 | 5.151     | 5.351 | 43.900             | 50.000            | -12.2 |
| Endosulfan II        | 6.082 | 5.984     | 6.184 | 44.400             | 50.000            | -11.2 |
| Endosulfan sulfate   | 6.483 | 6.386     | 6.586 | 44.010             | 50.000            | -12.0 |
| Endrin               | 5.791 | 5.693     | 5.893 | 44.040             | 50.000            | -11.9 |
| Endrin aldehyde      | 6.260 | 6.163     | 6.363 | 43.430             | 50.000            | -13.1 |
| Endrin ketone        | 6.993 | 6.895     | 7.095 | 44.410             | 50.000            | -11.2 |
| gamma-BHC (Lindane)  | 3.732 | 3.633     | 3.833 | 44.790             | 50.000            | -10.4 |
| gamma-Chlordane      | 5.127 | 5.030     | 5.230 | 45.010             | 50.000            | -10.0 |
| Heptachlor           | 4.085 | 3.986     | 4.186 | 45.020             | 50.000            | -10.0 |
| Heptachlor epoxide   | 4.875 | 4.777     | 4.977 | 45.110             | 50.000            | -9.8  |
| Methoxychlor         | 6.756 | 6.659     | 6.859 | 43.950             | 50.000            | -12.1 |
| Tetrachloro-m-xylene | 2.883 | 2.783     | 2.983 | 44.650             | 50.000            | -10.7 |