

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1982 SAS No.: Q1982 SDG NO.: Q1982

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL04 Date Analyzed: 05/13/2025

Lab Sample No.: PSTDCCC050 Data File : PD088495.D Time Analyzed: 02:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.706	6.606	6.806	53.680	50.000	7.4
4,4'-DDE	6.197	6.097	6.297	51.500	50.000	3.0
4,4'-DDT	7.022	6.922	7.122	52.040	50.000	4.1
Aldrin	5.273	5.173	5.373	54.330	50.000	8.7
alpha-BHC	4.001	3.901	4.101	53.850	50.000	7.7
alpha-Chlordane	6.029	5.929	6.129	54.070	50.000	8.1
beta-BHC	4.517	4.416	4.616	52.920	50.000	5.8
Decachlorobiphenyl	9.074	8.975	9.175	47.810	50.000	-4.4
delta-BHC	4.765	4.665	4.865	53.340	50.000	6.7
Dieldrin	6.349	6.249	6.449	53.650	50.000	7.3
Endosulfan I	6.076	5.976	6.176	53.640	50.000	7.3
Endosulfan II	6.787	6.687	6.887	51.580	50.000	3.2
Endosulfan sulfate	7.151	7.050	7.250	51.810	50.000	3.6
Endrin	6.576	6.476	6.676	52.100	50.000	4.2
Endrin aldehyde	6.916	6.816	7.016	49.710	50.000	-0.6
Endrin ketone	7.632	7.531	7.731	51.460	50.000	2.9
gamma-BHC (Lindane)	4.332	4.232	4.432	53.200	50.000	6.4
gamma-Chlordane	5.948	5.847	6.047	54.370	50.000	8.7
Heptachlor	4.932	4.831	5.031	54.940	50.000	9.9
Heptachlor epoxide	5.692	5.592	5.792	54.170	50.000	8.3
Methoxychlor	7.494	7.395	7.595	50.720	50.000	1.4
Tetrachloro-m-xylene	3.552	3.452	3.652	51.360	50.000	2.7

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1982 SAS No.: Q1982 SDG NO.: Q1982

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL04 Date Analyzed: 05/13/2025

Lab Sample No.: PSTDCCC050 Data File : PD088495.D Time Analyzed: 02:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.932	5.834	6.034	46.760	50.000	-6.5
4,4'-DDE	5.376	5.280	5.480	47.000	50.000	-6.0
4,4'-DDT	6.186	6.088	6.288	46.120	50.000	-7.8
Aldrin	4.370	4.273	4.473	47.650	50.000	-4.7
alpha-BHC	3.395	3.296	3.496	47.610	50.000	-4.8
alpha-Chlordane	5.193	5.094	5.294	46.780	50.000	-6.4
beta-BHC	4.027	3.928	4.128	47.270	50.000	-5.5
Decachlorobiphenyl	8.075	7.977	8.177	41.040	50.000	-17.9
delta-BHC	4.264	4.165	4.365	47.760	50.000	-4.5
Dieldrin	5.515	5.417	5.617	46.920	50.000	-6.2
Endosulfan I	5.249	5.151	5.351	42.220	50.000	-15.6
Endosulfan II	6.083	5.984	6.184	46.210	50.000	-7.6
Endosulfan sulfate	6.484	6.386	6.586	45.230	50.000	-9.5
Endrin	5.791	5.693	5.893	46.260	50.000	-7.5
Endrin aldehyde	6.261	6.163	6.363	43.860	50.000	-12.3
Endrin ketone	6.994	6.895	7.095	44.320	50.000	-11.4
gamma-BHC (Lindane)	3.732	3.633	3.833	46.970	50.000	-6.1
gamma-Chlordane	5.128	5.030	5.230	46.970	50.000	-6.1
Heptachlor	4.085	3.986	4.186	47.270	50.000	-5.5
Heptachlor epoxide	4.875	4.777	4.977	47.200	50.000	-5.6
Methoxychlor	6.757	6.659	6.859	44.620	50.000	-10.8
Tetrachloro-m-xylene	2.883	2.783	2.983	46.920	50.000	-6.2