

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1982 SAS No.: Q1982 SDG NO.: Q1982

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL05 Date Analyzed: 05/13/2025

Lab Sample No.: PSTDCCC050 Data File : PD088499.D Time Analyzed: 09:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.706	6.606	6.806	55.230	50.000	10.5
4,4'-DDE	6.197	6.097	6.297	52.340	50.000	4.7
4,4'-DDT	7.023	6.922	7.122	53.040	50.000	6.1
Aldrin	5.273	5.173	5.373	54.770	50.000	9.5
alpha-BHC	4.001	3.901	4.101	54.670	50.000	9.3
alpha-Chlordane	6.028	5.929	6.129	54.970	50.000	9.9
beta-BHC	4.516	4.416	4.616	53.370	50.000	6.7
Decachlorobiphenyl	9.074	8.975	9.175	48.950	50.000	-2.1
delta-BHC	4.765	4.665	4.865	54.030	50.000	8.1
Dieldrin	6.349	6.249	6.449	54.680	50.000	9.4
Endosulfan I	6.075	5.976	6.176	54.560	50.000	9.1
Endosulfan II	6.787	6.687	6.887	52.760	50.000	5.5
Endosulfan sulfate	7.151	7.050	7.250	52.790	50.000	5.6
Endrin	6.576	6.476	6.676	53.150	50.000	6.3
Endrin aldehyde	6.917	6.816	7.016	50.890	50.000	1.8
Endrin ketone	7.632	7.531	7.731	52.520	50.000	5.0
gamma-BHC (Lindane)	4.332	4.232	4.432	53.830	50.000	7.7
gamma-Chlordane	5.948	5.847	6.047	55.410	50.000	10.8
Heptachlor	4.931	4.831	5.031	55.670	50.000	11.3
Heptachlor epoxide	5.692	5.592	5.792	54.860	50.000	9.7
Methoxychlor	7.494	7.395	7.595	51.790	50.000	3.6
Tetrachloro-m-xylene	3.552	3.452	3.652	51.960	50.000	3.9

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1982 SAS No.: Q1982 SDG NO.: Q1982

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL05 Date Analyzed: 05/13/2025

Lab Sample No.: PSTDCCC050 Data File : PD088499.D Time Analyzed: 09:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.932	5.834	6.034	47.100	50.000	-5.8
4,4'-DDE	5.376	5.280	5.480	47.200	50.000	-5.6
4,4'-DDT	6.186	6.088	6.288	46.910	50.000	-6.2
Aldrin	4.369	4.273	4.473	47.620	50.000	-4.8
alpha-BHC	3.395	3.296	3.496	47.880	50.000	-4.2
alpha-Chlordane	5.192	5.094	5.294	47.010	50.000	-6.0
beta-BHC	4.027	3.928	4.128	47.190	50.000	-5.6
Decachlorobiphenyl	8.075	7.977	8.177	42.160	50.000	-15.7
delta-BHC	4.263	4.165	4.365	47.870	50.000	-4.3
Dieldrin	5.515	5.417	5.617	47.160	50.000	-5.7
Endosulfan I	5.248	5.151	5.351	41.720	50.000	-16.6
Endosulfan II	6.083	5.984	6.184	46.750	50.000	-6.5
Endosulfan sulfate	6.485	6.386	6.586	45.680	50.000	-8.6
Endrin	5.791	5.693	5.893	46.530	50.000	-6.9
Endrin aldehyde	6.261	6.163	6.363	44.600	50.000	-10.8
Endrin ketone	6.994	6.895	7.095	45.040	50.000	-9.9
gamma-BHC (Lindane)	3.731	3.633	3.833	47.220	50.000	-5.6
gamma-Chlordane	5.128	5.030	5.230	47.240	50.000	-5.5
Heptachlor	4.085	3.986	4.186	47.280	50.000	-5.4
Heptachlor epoxide	4.875	4.777	4.977	47.370	50.000	-5.3
Methoxychlor	6.757	6.659	6.859	45.070	50.000	-9.9
Tetrachloro-m-xylene	2.882	2.783	2.983	47.090	50.000	-5.8