

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1982 SAS No.: Q1982 SDG NO.: Q1982

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL06 Date Analyzed: 05/13/2025

Lab Sample No.: PSTDCCC050 Data File : PD088515.D Time Analyzed: 14:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.712	6.606	6.806	57.230	50.000	14.5
4,4'-DDE	6.203	6.097	6.297	52.410	50.000	4.8
4,4'-DDT	7.028	6.922	7.122	52.550	50.000	5.1
Aldrin	5.279	5.173	5.373	55.590	50.000	11.2
alpha-BHC	4.007	3.901	4.101	55.650	50.000	11.3
alpha-Chlordane	6.035	5.929	6.129	56.450	50.000	12.9
beta-BHC	4.523	4.416	4.616	54.330	50.000	8.7
Decachlorobiphenyl	9.080	8.975	9.175	48.470	50.000	-3.1
delta-BHC	4.771	4.665	4.865	56.620	50.000	13.2
Dieldrin	6.354	6.249	6.449	54.840	50.000	9.7
Endosulfan I	6.082	5.976	6.176	55.440	50.000	10.9
Endosulfan II	6.793	6.687	6.887	53.040	50.000	6.1
Endosulfan sulfate	7.156	7.050	7.250	53.000	50.000	6.0
Endrin	6.582	6.476	6.676	53.290	50.000	6.6
Endrin aldehyde	6.922	6.816	7.016	51.380	50.000	2.8
Endrin ketone	7.637	7.531	7.731	52.450	50.000	4.9
gamma-BHC (Lindane)	4.338	4.232	4.432	55.030	50.000	10.1
gamma-Chlordane	5.952	5.847	6.047	54.280	50.000	8.6
Heptachlor	4.937	4.831	5.031	55.970	50.000	11.9
Heptachlor epoxide	5.698	5.592	5.792	54.730	50.000	9.5
Methoxychlor	7.500	7.395	7.595	51.030	50.000	2.1
Tetrachloro-m-xylene	3.558	3.452	3.652	52.540	50.000	5.1

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q1982 SAS No.: Q1982 SDG NO.: Q1982

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL06 Date Analyzed: 05/13/2025

Lab Sample No.: PSTDCCC050 Data File : PD088515.D Time Analyzed: 14:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.933	5.834	6.034	47.380	50.000	-5.2
4,4'-DDE	5.377	5.280	5.480	47.480	50.000	-5.0
4,4'-DDT	6.188	6.088	6.288	46.650	50.000	-6.7
Aldrin	4.370	4.273	4.473	48.250	50.000	-3.5
alpha-BHC	3.395	3.296	3.496	48.460	50.000	-3.1
alpha-Chlordane	5.194	5.094	5.294	47.400	50.000	-5.2
beta-BHC	4.028	3.928	4.128	48.190	50.000	-3.6
Decachlorobiphenyl	8.076	7.977	8.177	41.330	50.000	-17.3
delta-BHC	4.265	4.165	4.365	48.660	50.000	-2.7
Dieldrin	5.516	5.417	5.617	47.580	50.000	-4.8
Endosulfan I	5.249	5.151	5.351	41.650	50.000	-16.7
Endosulfan II	6.084	5.984	6.184	46.980	50.000	-6.0
Endosulfan sulfate	6.486	6.386	6.586	45.710	50.000	-8.6
Endrin	5.792	5.693	5.893	47.220	50.000	-5.6
Endrin aldehyde	6.262	6.163	6.363	44.930	50.000	-10.1
Endrin ketone	6.995	6.895	7.095	45.010	50.000	-10.0
gamma-BHC (Lindane)	3.732	3.633	3.833	47.870	50.000	-4.3
gamma-Chlordane	5.129	5.030	5.230	47.660	50.000	-4.7
Heptachlor	4.086	3.986	4.186	47.490	50.000	-5.0
Heptachlor epoxide	4.876	4.777	4.977	47.790	50.000	-4.4
Methoxychlor	6.758	6.659	6.859	45.030	50.000	-9.9
Tetrachloro-m-xylene	2.882	2.783	2.983	47.460	50.000	-5.1