284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1982 SAS No.: Q1982 SDG NO.: Q1982

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/15/2025

Lab File ID: BP024634.D Time Analyzed: 12:25

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

		IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
	12 HOUR STD	130531	7.658	590490	10.43	383835	14.30
	UPPER LIMIT	261062	8.158	1180980	10.928	767670	14.804
	LOWER LIMIT	65265.5	7.158	295245	9.928	191918	13.804
	EPA SAMPLE NO.						
01	PB167973BL	146276	7.66	564318	10.43	357828	14.29
02	PB167973BS	131081	7.65	537517	10.42	347277	14.30
03	TP-9	165553	7.66	643582	10.43	391444	14.29
04	TP-13MS	165529	7.66	634360	10.43	378565	14.30
05	TP-13MSD	159501	7.66	628820	10.43	375530	14.30

IS1 (DCB) = 1,4-Dichlorobenzene-d4
IS2 (NPT) = Naphthalene-d8
IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.