

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1982

Client: CDM Smith

Analytical Method: SW8260D

Datafile : VY022263.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0515SBSD01	Dichlorodifluoromethane	20	21.3	ug/Kg	106	6		64	136	20
	Chloromethane	20	23.0	ug/Kg	115	6		70	130	20
	Vinyl chloride	20	22.8	ug/Kg	114	3		72	129	20
	Bromomethane	20	24.3	ug/Kg	121	2		58	141	20
	Chloroethane	20	22.4	ug/Kg	112	2		69	130	20
	Trichlorofluoromethane	20	22.1	ug/Kg	111	4		69	134	20
	1,1,2-Trichlorotrifluoroethane	20	20.7	ug/Kg	104	2		81	123	20
	1,1-Dichloroethene	20	20.0	ug/Kg	100	7		79	121	20
	Acetone	100	100	ug/Kg	100	0		60	131	20
	Carbon disulfide	20	20.5	ug/Kg	103	4		45	154	20
	Methyl tert-butyl Ether	20	21.0	ug/Kg	105	3		77	129	20
	Methyl Acetate	20	21.3	ug/Kg	106	10		69	149	20
	Methylene Chloride	20	21.3	ug/Kg	106	0		56	174	20
	trans-1,2-Dichloroethene	20	20.7	ug/Kg	104	1		80	123	20
	1,1-Dichloroethane	20	20.3	ug/Kg	102	2		82	123	20
	Cyclohexane	20	19.9	ug/Kg	100	5		76	122	20
	2-Butanone	100	100	ug/Kg	100	2		69	131	20
	Carbon Tetrachloride	20	19.8	ug/Kg	99	3		76	129	20
	cis-1,2-Dichloroethene	20	20.1	ug/Kg	101	3		82	123	20
	Bromochloromethane	20	20.3	ug/Kg	102	2		80	127	20
	Chloroform	20	20.5	ug/Kg	103	3		82	125	20
	1,1,1-Trichloroethane	20	20.1	ug/Kg	101	5		80	126	20
	Methylcyclohexane	20	20.0	ug/Kg	100	6		77	123	20
	Benzene	20	20.0	ug/Kg	100	4		84	121	20
	1,2-Dichloroethane	20	20.4	ug/Kg	102	0		81	126	20
	Trichloroethene	20	19.8	ug/Kg	99	4		83	122	20
	1,2-Dichloropropane	20	19.9	ug/Kg	100	2		83	122	20
	Bromodichloromethane	20	20.3	ug/Kg	102	2		82	123	20
	4-Methyl-2-Pentanone	100	110	ug/Kg	110	12		70	135	20
	Toluene	20	19.5	ug/Kg	98	3		83	122	20
	t-1,3-Dichloropropene	20	19.9	ug/Kg	100	0		78	124	20
	cis-1,3-Dichloropropene	20	19.9	ug/Kg	100	3		81	122	20
	1,1,2-Trichloroethane	20	20.6	ug/Kg	103	2		82	125	20
	2-Hexanone	100	100	ug/Kg	100	4		66	138	20
	Dibromochloromethane	20	20.4	ug/Kg	102	2		79	125	20
	1,2-Dibromoethane	20	20.5	ug/Kg	103	2		80	125	20
	Tetrachloroethene	20	19.9	ug/Kg	100	0		83	125	20
	Chlorobenzene	20	20.0	ug/Kg	100	2		84	122	20
	Ethyl Benzene	20	19.6	ug/Kg	98	3		82	124	20
	m/p-Xylenes	40	39.1	ug/Kg	98	2		83	124	20
	o-Xylene	20	19.7	ug/Kg	99	1		83	123	20
	Styrene	20	19.4	ug/Kg	97	2		82	124	20
	Bromoform	20	20.6	ug/Kg	103	4		75	127	20
	Isopropylbenzene	20	19.6	ug/Kg	98	3		82	124	20
	1,1,2,2-Tetrachloroethane	20	21.3	ug/Kg	106	5		77	127	20
	1,3-Dichlorobenzene	20	19.3	ug/Kg	97	1		83	122	20
	1,4-Dichlorobenzene	20	20.0	ug/Kg	100	0		84	121	20
	1,2-Dichlorobenzene	20	20.0	ug/Kg	100	1		83	124	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1982
Client: CDM Smith
Analytical Method: SW8260D **Datafile :** VY022263.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0515SBSD01	1,2-Dibromo-3-Chloropropane	20	21.4	ug/Kg	107	3		66	134	20
	1,2,4-Trichlorobenzene	20	20.6	ug/Kg	103	4		78	127	20
	1,2,3-Trichlorobenzene	20	21.0	ug/Kg	105	8		70	137	20