

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q1982 SAS No.: Q1982 SDG NO.: Q1982
 Lab File ID: VY022256.D Date Analyzed: 05/15/2025
 Instrument ID: MSVOA_Y Time Analyzed: 11:02
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	267598	7.71	439323	8.62	377571	11.41
UPPER LIMIT	535196	8.207	878646	9.116	755142	11.914
LOWER LIMIT	133799	7.207	219662	8.116	188786	10.914
EPA SAMPLE NO.						
TP-4	303248	7.71	528458	8.61	395320	11.41
TP-5	345998	7.71	610490	8.62	492184	11.41
TP-6	302935	7.71	529384	8.62	407189	11.41
VY0515SBL01	357935	7.71	617938	8.62	477752	11.41
VY0515SBS01	240556	7.71	405751	8.62	344872	11.41
VY0515SBSD01	238919	7.71	402031	8.62	341987	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.