

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q1982 SAS No.: Q1982 SDG NO.: Q1982
 Lab File ID: VY022256.D Date Analyzed: 05/15/2025
 Instrument ID: MSVOA_Y Time Analyzed: 11:02
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	180358	13.346				
UPPER LIMIT	360716	13.846				
LOWER LIMIT	90179	12.846				
EPA SAMPLE NO.						
TP-4	123883	13.35				
TP-5	188552	13.35				
TP-6	138768	13.35				
VY0515SBL01	173912	13.35				
VY0515SBS01	163448	13.35				
VY0515SBSD01	161126	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.