



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

### **Nobis Group**

**Project Name: Raymark Superfund Site**

**Project # N/A**

**Order ID # Q1984**

**Test Name: Mercury, Metals ICP-TAL**

### **A. Number of Samples and Date of Receipt:**

17 Solid samples were received on 05/08/2025.

### **B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: Cyanide, Herbicide Group1, Mercury, Metals ICP-TAL, METALS-TAL, PCB, Pesticide-TCL, SPLP Extraction, SPLP Mercury, SPLP MetalGroup3, SVOCMS Group3, VOCMS Group1 and VOCMS Group3. This data package contains results for Mercury, Metals ICP-TAL.

### **C. Analytical Techniques:**

The analysis of Metals ICP-TAL was based on method 6010D, digestion based on method 3050 (soils). The analysis and digestion of Mercury was based on method 7471B.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate (OU4-TS-28-050725DUP) analysis met criteria for all samples except for Calcium due to matrix interference.

The Matrix Spike (OU4-TS-28-050725MS) analysis met criteria for all samples except for Antimony, Arsenic, Beryllium, Chromium, Cobalt, Copper, Selenium, Silver, Sodium, Vanadium due to matrix interference.

The Matrix Spike Duplicate (OU4-TS-28-050725MSD) analysis met criteria for all samples except for Antimony, Arsenic, Beryllium, Chromium, Cobalt, Copper, Lead, Nickel, Selenium, Silver, Sodium, Vanadium due to matrix interference.

The Post Digest Spike (OU4-TS-28-050725A) analysis met criteria for all samples except for Antimony, Arsenic, Beryllium, Chromium, Cobalt, Copper, Nickel, Selenium, Silver, Vanadium due to unknown chemical interference of matrix with the addition of spike amount after digestion and before analysis, matrix has suppression effect during addition of spike.

The Blank analysis did not indicate the presence of lab contamination.



The Calibration met the requirements.

The Serial Dilution (OU4-TS-28-050725L) met criteria for all samples except for Arsenic, Calcium, Chromium, Copper, Iron, Magnesium, Manganese, Zinc due to unknown interference.

The Serial Dilution (VNJ-218L) met criteria for all samples except for Mercury due to unknown interference.

**E. Additional Comments:**

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_