

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135733

Review By	MOHAN	Review On	5/13/2025 11:37:48 AM
Supervise By	jaswal	Supervise On	5/13/2025 11:15:56 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85626,MP85627,MP85628,MP85629,MP85630,MP85631		
ICV Standard	MP85632		
CCV Standard	MP85634		
ICSA Standard			
CRI Standard	MP85636		
LCS Standard			
Chk Standard	MP85633,MP85634,MP837,MP85646		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	05/12/25 09:36		MOHAN	OK
2	S0.2	S0.2	CAL2	05/12/25 09:38		MOHAN	OK
3	S2.5	S2.5	CAL3	05/12/25 09:40		MOHAN	OK
4	S5	S5	CAL4	05/12/25 09:45		MOHAN	OK
5	S7.5	S7.5	CAL5	05/12/25 09:47		MOHAN	OK
6	S10	S10	CAL6	05/12/25 09:53		MOHAN	OK
7	ICV05	ICV05	ICV	05/12/25 09:56		MOHAN	OK
8	ICB05	ICB05	ICB	05/12/25 09:58		MOHAN	OK
9	CCV07	CCV07	CCV	05/12/25 10:06		MOHAN	OK
10	CCB07	CCB07	CCB	05/12/25 10:08		MOHAN	OK
11	CRA	CRA	CRDL	05/12/25 10:11		MOHAN	OK
12	HighStd	HighStd	HIGH STD	05/12/25 10:13		MOHAN	OK
13	ChkStd	ChkStd	SAM	05/12/25 10:15		MOHAN	OK
14	PB167937BL	PB167937BL	MB	05/12/25 10:18		MOHAN	OK
15	PB167937BS	PB167937BS	LCS	05/12/25 10:22		MOHAN	OK
16	Q1982-01	TP-1	SAM	05/12/25 10:25		MOHAN	OK
17	Q1982-02	TP-2B	SAM	05/12/25 10:27		MOHAN	OK
18	Q1982-03	TP-3	SAM	05/12/25 10:29		MOHAN	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135733

Review By	MOHAN	Review On	5/13/2025 11:37:48 AM
Supervise By	jaswal	Supervise On	5/13/2025 11:15:56 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85626,MP85627,MP85628,MP85629,MP85630,MP85631		
ICV Standard	MP85632		
CCV Standard	MP85634		
ICSA Standard			
CRI Standard	MP85636		
LCS Standard			
Chk Standard	MP85633,MP85634,MP837,MP85646		

19	Q1982-04	TP-4	SAM	05/12/25 10:32		MOHAN	OK
20	Q1982-05	TP-5	SAM	05/12/25 10:34		MOHAN	OK
21	CCV08	CCV08	CCV	05/12/25 10:36		MOHAN	OK
22	CCB08	CCB08	CCB	05/12/25 10:39		MOHAN	OK
23	Q1982-06	TP-6	SAM	05/12/25 10:41		MOHAN	OK
24	Q1982-07	TP-8	SAM	05/12/25 10:43		MOHAN	OK
25	Q1982-08	TP-9	SAM	05/12/25 10:45		MOHAN	OK
26	Q1983-01	OR-636-COMP-01	SAM	05/12/25 10:48		MOHAN	OK
27	Q1983-01DUP	OR-636-COMP-01DUP	DUP	05/12/25 10:50		MOHAN	OK
28	Q1983-01MS	OR-636-COMP-01MS	MS	05/12/25 10:52		MOHAN	OK
29	Q1983-01MSD	OR-636-COMP-01MSD	MSD	05/12/25 10:55		MOHAN	OK
30	Q1983-07	OR-636-COMP-02	SAM	05/12/25 10:57		MOHAN	OK
31	Q1983-13	OR-636-COMP-03	SAM	05/12/25 10:59		MOHAN	OK
32	Q1983-19	OR-636-COMP-04	SAM	05/12/25 11:01		MOHAN	OK
33	CCV09	CCV09	CCV	05/12/25 11:09		MOHAN	OK
34	CCB09	CCB09	CCB	05/12/25 11:11		MOHAN	OK
35	Q1983-25	OR-636-COMP-05	SAM	05/12/25 11:14		MOHAN	OK
36	Q1983-31	OR-636-COMP-06	SAM	05/12/25 11:16		MOHAN	OK
37	Q1983-37	OR-636-COMP-07	SAM	05/12/25 11:18		MOHAN	OK
38	Q1983-43	OR-636-COMP-08	SAM	05/12/25 11:20		MOHAN	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135733

Review By	MOHAN	Review On	5/13/2025 11:37:48 AM
Supervise By	jaswal	Supervise On	5/13/2025 11:15:56 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85626,MP85627,MP85628,MP85629,MP85630,MP85631		
ICV Standard	MP85632		
CCV Standard	MP85634		
ICSA Standard			
CRI Standard	MP85636		
LCS Standard			
Chk Standard	MP85633,MP85634,MP837,MP85646		

39	Q1983-49	OR-636-COMP-09	SAM	05/12/25 11:23		MOHAN	OK
40	PB167938BL	PB167938BL	MB	05/12/25 11:25		MOHAN	OK
41	PB167938BS	PB167938BS	LCS	05/12/25 11:27		MOHAN	OK
42	Q1984-01	OU4-PCS-TC-33-050	SAM	05/12/25 11:30		MOHAN	OK
43	Q1984-03	OU4-PCS-TC-34-050	SAM	05/12/25 11:32		MOHAN	OK
44	Q1984-05	OU4-PCS-TC-35-050	SAM	05/12/25 11:34		MOHAN	OK
45	CCV10	CCV10	CCV	05/12/25 11:36		MOHAN	OK
46	CCB10	CCB10	CCB	05/12/25 11:39		MOHAN	OK
47	Q1984-07	OU4-TS-24-050725	SAM	05/12/25 11:41		MOHAN	OK
48	Q1984-09	OU4-TS-25-050725	SAM	05/12/25 11:43		MOHAN	OK
49	Q1984-11	OU4-TS-26-050725	SAM	05/12/25 11:46		MOHAN	OK
50	Q1984-13	OU4-TS-27-050725	SAM	05/12/25 11:48		MOHAN	OK
51	Q1984-15	OU4-TS-28-050725	SAM	05/12/25 11:50		MOHAN	OK
52	Q1986-01	COMP-8	SAM	05/12/25 11:56		MOHAN	OK
53	Q1986-03	COMP-9	SAM	05/12/25 12:00		MOHAN	OK
54	Q1986-05	COMP-10	SAM	05/12/25 12:02		MOHAN	OK
55	Q1986-07	COMP-11	SAM	05/12/25 12:05		MOHAN	OK
56	Q1986-09	VNJ-218	SAM	05/12/25 12:07		MOHAN	OK
57	CCV11	CCV11	CCV	05/12/25 12:42		MOHAN	OK
58	CCB11	CCB11	CCB	05/12/25 12:53		MOHAN	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135733

Review By	MOHAN	Review On	5/13/2025 11:37:48 AM
Supervise By	jaswal	Supervise On	5/13/2025 11:15:56 PM

STD. NAME	STD REF.#
ICAL Standard	MP85626,MP85627,MP85628,MP85629,MP85630,MP85631
ICV Standard	MP85632
CCV Standard	MP85634
ICSA Standard	
CRI Standard	MP85636
LCS Standard	
Chk Standard	MP85633,MP85634,MP837,MP85646

59	Q1986-09DUP	VNJ-218DUP	DUP	05/12/25 12:58		MOHAN	OK
60	Q1986-09MS	VNJ-218MS	MS	05/12/25 13:00		MOHAN	OK
61	Q1986-09MSD	VNJ-218MSD	MSD	05/12/25 13:03		MOHAN	OK
62	Q2002-01	EO-02-05092025	SAM	05/12/25 13:05		MOHAN	OK
63	Q1983-01L	OR-636-COMP-01L	SD	05/12/25 13:08		MOHAN	OK
64	Q1983-01A	OR-636-COMP-01A	PS	05/12/25 13:10		MOHAN	OK
65	Q1986-09L	VNJ-218L	SD	05/12/25 13:12		MOHAN	OK
66	Q1986-09A	VNJ-218A	PS	05/12/25 13:15		MOHAN	OK
67	CCV12	CCV12	CCV	05/12/25 13:17		MOHAN	OK
68	CCB12	CCB12	CCB	05/12/25 13:27		MOHAN	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135855

Review By	Janvi	Review On	5/22/2025 10:50:55 AM
Supervise By	jaswal	Supervise On	5/22/2025 2:47:40 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546		
ICV Standard	MP85553		
CCV Standard	MP85556		
ICSA Standard	MP85554,MP85555		
CRI Standard	MP85552		
LCS Standard			
Chk Standard	MP85557,MP85558		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	05/20/25 12:07		Kareem	OK
2	S1	S1	CAL2	05/20/25 12:12		Kareem	OK
3	S2	S2	CAL3	05/20/25 12:16		Kareem	OK
4	S3	S3	CAL4	05/20/25 12:20		Kareem	OK
5	S4	S4	CAL5	05/20/25 12:24		Kareem	OK
6	S5	S5	CAL6	05/20/25 12:29		Kareem	OK
7	ICV01	ICV01	ICV	05/20/25 12:33		Kareem	OK
8	LLICV01	LLICV01	LLICV	05/20/25 12:40		Kareem	OK
9	ICB01	ICB01	ICB	05/20/25 12:44		Kareem	OK
10	CRI01	CRI01	CRDL	05/20/25 12:48		Kareem	OK
11	ICSA01	ICSA01	ICSA	05/20/25 12:53		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	05/20/25 13:00		Kareem	OK
13	ICSADL	ICSADL	ICSA	05/20/25 13:04		Kareem	OK
14	ICSABDL	ICSABDL	ICSAB	05/20/25 13:09		Kareem	OK
15	CCV01	CCV01	CCV	05/20/25 13:14		Kareem	OK
16	CCB01	CCB01	CCB	05/20/25 13:18		Kareem	OK
17	Q2034-01	L3-WC-1	SAM	05/20/25 13:22		Kareem	OK
18	Q2034-05	L3-WC-2	SAM	05/20/25 13:26		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135855

Review By	Janvi	Review On	5/22/2025 10:50:55 AM
Supervise By	jaswal	Supervise On	5/22/2025 2:47:40 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546		
ICV Standard	MP85553		
CCV Standard	MP85556		
ICSA Standard	MP85554,MP85555		
CRI Standard	MP85552		
LCS Standard			
Chk Standard	MP85557,MP85558		

19	Q2034-09	L3-WC-3	SAM	05/20/25 13:30		Kareem	OK
20	Q2034-13	L3-WC-4	SAM	05/20/25 13:34		Kareem	OK
21	Q2034-17	L3-WC-5	SAM	05/20/25 13:39		Kareem	OK
22	Q2034-21	L3-WC-6	SAM	05/20/25 13:43		Kareem	OK
23	Q2048-01	L2-WC-1	SAM	05/20/25 13:47		Kareem	OK
24	Q2048-05	L2-WC-2	SAM	05/20/25 13:51		Kareem	OK
25	Q2048-09	L2-WC-3	SAM	05/20/25 13:55		Kareem	OK
26	CCV02	CCV02	CCV	05/20/25 13:59		Kareem	OK
27	CCB02	CCB02	CCB	05/20/25 14:04		Kareem	OK
28	Q2048-13	L2-WC-4	SAM	05/20/25 14:08		Kareem	OK
29	PB168023BL	PB168023BL	MB	05/20/25 14:12		Kareem	OK
30	PB168023BS	PB168023BS	LCS	05/20/25 14:16		Kareem	OK
31	PB168063BL	PB168063BL	MB	05/20/25 14:20		Kareem	OK
32	PB168063BS	PB168063BS	LCS	05/20/25 14:25		Kareem	OK
33	PB167931BL	PB167931BL	MB	05/20/25 14:29		Kareem	OK
34	PB167931BS	PB167931BS	LCS	05/20/25 14:33		Kareem	OK
35	Q2032-01	TP-11	SAM	05/20/25 14:37		Kareem	OK
36	Q2032-02	TP-29	SAM	05/20/25 14:41		Kareem	OK
37	CCV03	CCV03	CCV	05/20/25 14:46		Kareem	OK
38	CCB03	CCB03	CCB	05/20/25 14:50		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135855

Review By	Janvi	Review On	5/22/2025 10:50:55 AM
Supervise By	jaswal	Supervise On	5/22/2025 2:47:40 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546		
ICV Standard	MP85553		
CCV Standard	MP85556		
ICSA Standard	MP85554,MP85555		
CRI Standard	MP85552		
LCS Standard			
Chk Standard	MP85557,MP85558		

39	Q2032-03	TP-29-99	SAM	05/20/25 14:54		Kareem	OK
40	Q2032-04	TP-24	SAM	05/20/25 14:58		Kareem	OK
41	Q2032-04DUP	TP-24DUP	DUP	05/20/25 15:02		Kareem	OK
42	Q2032-04L	TP-24L	SD	05/20/25 15:07		Kareem	OK
43	Q2032-05	TP-24MS	MS	05/20/25 15:11		Kareem	OK
44	Q2032-06	TP-24MSD	MSD	05/20/25 15:15		Kareem	OK
45	Q2032-04A	TP-24A	PS	05/20/25 15:19		Kareem	OK
46	Q2032-07	TP-37	SAM	05/20/25 15:23		Kareem	OK
47	Q2032-08	TP-32	SAM	05/20/25 15:27		Kareem	OK
48	CCV04	CCV04	CCV	05/20/25 15:44		Kareem	OK
49	CCB04	CCB04	CCB	05/20/25 15:50		Kareem	OK
50	PB168033TB	PB168033TB	MB	05/20/25 15:54		Kareem	OK
51	Q2052-04	TP-B	SAM	05/20/25 15:59		Kareem	OK
52	Q2053-01	SB-1	SAM	05/20/25 16:03		Kareem	OK
53	Q2053-02	SB-2	SAM	05/20/25 16:08		Kareem	OK
54	Q2053-03	SB-3	SAM	05/20/25 16:12		Kareem	OK
55	Q2053-04	SB-4	SAM	05/20/25 16:16		Kareem	OK
56	Q2053-05	SB-5	SAM	05/20/25 16:21		Kareem	OK
57	Q2053-06	SB-6	SAM	05/20/25 16:25		Kareem	OK
58	Q2053-07	SB-7	SAM	05/20/25 16:29		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135855

Review By	Janvi	Review On	5/22/2025 10:50:55 AM
Supervise By	jaswal	Supervise On	5/22/2025 2:47:40 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546		
ICV Standard	MP85553		
CCV Standard	MP85556		
ICSA Standard	MP85554,MP85555		
CRI Standard	MP85552		
LCS Standard			
Chk Standard	MP85557,MP85558		

59	CCV05	CCV05	CCV	05/20/25 16:34		Kareem	OK
60	CCB05	CCB05	CCB	05/20/25 16:38		Kareem	OK
61	Q2053-08	SB-8	SAM	05/20/25 16:42		Kareem	OK
62	Q2053-09	SB-9	SAM	05/20/25 16:47		Kareem	OK
63	Q2053-10	SB-10	SAM	05/20/25 16:51		Kareem	OK
64	Q2053-11	SB-11	SAM	05/20/25 16:56		Kareem	OK
65	Q2057-04	MH-L	SAM	05/20/25 17:00		Kareem	OK
66	Q2062-04	L1-WC-1	SAM	05/20/25 17:04		Kareem	OK
67	Q2062-08	L1-WC-2	SAM	05/20/25 17:09		Kareem	OK
68	Q2062-12	L1-WC-3	SAM	05/20/25 17:13		Kareem	OK
69	Q2062-16	L1-WC-4	SAM	05/20/25 17:18		Kareem	OK
70	CCV06	CCV06	CCV	05/20/25 17:22		Kareem	OK
71	CCB06	CCB06	CCB	05/20/25 17:26		Kareem	OK
72	Q2062-20	L1-WC-5	SAM	05/20/25 17:31		Kareem	OK
73	Q2062-24	L1-WC-6	SAM	05/20/25 17:35		Kareem	OK
74	Q2062-24DUP	L1-WC-6DUP	DUP	05/20/25 17:39		Kareem	OK
75	Q2062-24L	L1-WC-6L	SD	05/20/25 17:44		Kareem	OK
76	Q2062-24MS	L1-WC-6MS	MS	05/20/25 17:48		Kareem	OK
77	Q2062-24MSD	L1-WC-6MSD	MSD	05/20/25 17:52		Kareem	OK
78	Q2062-24A	L1-WC-6A	PS	05/20/25 17:57		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135855

Review By	Janvi	Review On	5/22/2025 10:50:55 AM
Supervise By	jaswal	Supervise On	5/22/2025 2:47:40 PM

STD. NAME	STD REF.#
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546
ICV Standard	MP85553
CCV Standard	MP85556
ICSA Standard	MP85554,MP85555
CRI Standard	MP85552
LCS Standard	
Chk Standard	MP85557,MP85558

79	Q1984-01	OU4-PCS-TC-33-050	SAM	05/20/25 18:01		Kareem	OK
80	Q1984-03	OU4-PCS-TC-34-050	SAM	05/20/25 18:05		Kareem	OK
81	CCV07	CCV07	CCV	05/20/25 18:09		Kareem	OK
82	CCB07	CCB07	CCB	05/20/25 18:13		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135868

Review By	Janvi	Review On	5/22/2025 2:51:30 PM
Supervise By	jaswal	Supervise On	5/22/2025 2:52:47 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546		
ICV Standard	MP85553		
CCV Standard	MP85556		
ICSA Standard	MP85554,MP85555		
CRI Standard	MP85552		
LCS Standard			
Chk Standard	MP85557,MP85558		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	05/21/25 12:05		Kareem	OK
2	S1	S1	CAL2	05/21/25 12:09		Kareem	OK
3	S2	S2	CAL3	05/21/25 12:13		Kareem	OK
4	S3	S3	CAL4	05/21/25 12:17		Kareem	OK
5	S4	S4	CAL5	05/21/25 12:22		Kareem	OK
6	S5	S5	CAL6	05/21/25 12:26		Kareem	OK
7	ICV01	ICV01	ICV	05/21/25 13:19		Kareem	OK
8	LLICV01	LLICV01	LLICV	05/21/25 13:29		Kareem	OK
9	ICB01	ICB01	ICB	05/21/25 13:36		Kareem	OK
10	CRI01	CRI01	CRDL	05/21/25 13:43		Kareem	OK
11	ICSA01	ICSA01	ICSA	05/21/25 13:52		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	05/21/25 13:56		Kareem	OK
13	ICSADL	ICSADL	ICSA	05/21/25 14:00		Kareem	OK
14	ICSABDL	ICSABDL	ICSAB	05/21/25 14:05		Kareem	OK
15	CCV01	CCV01	CCV	05/21/25 14:09		Kareem	OK
16	CCB01	CCB01	CCB	05/21/25 14:13		Kareem	OK
17	Q2068-01	LAW-25-0075	SAM	05/21/25 14:17		Kareem	OK
18	Q2067-01	303-PPR-FRAC	SAM	05/21/25 14:22		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135868

Review By	Janvi	Review On	5/22/2025 2:51:30 PM
Supervise By	jaswal	Supervise On	5/22/2025 2:52:47 PM

STD. NAME	STD REF.#
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546
ICV Standard	MP85553
CCV Standard	MP85556
ICSA Standard	MP85554,MP85555
CRI Standard	MP85552
LCS Standard	
Chk Standard	MP85557,MP85558

19	Q2068-03DL	LAW-25-0074DL	SAM	05/21/25 14:26		Kareem	OK
20	Q2065-01	VACTRUCK-728068	SAM	05/21/25 14:30	Ca High	Kareem	Dilution
21	Q2090-01	001-WILLETS-PT-BL	SAM	05/21/25 14:35		Kareem	OK
22	Q2090-02	002-35TH-AVE(APR)	SAM	05/21/25 14:39		Kareem	OK
23	Q2090-02DUP	002-35TH-AVE(APR)	DUP	05/21/25 14:44		Kareem	OK
24	Q2090-02L	002-35TH-AVE(APR)	SD	05/21/25 14:48		Kareem	OK
25	Q2090-02MS	002-35TH-AVE(APR)	MS	05/21/25 14:52		Kareem	OK
26	Q2090-02MSD	002-35TH-AVE(APR)	MSD	05/21/25 14:56		Kareem	OK
27	CCV02	CCV02	CCV	05/21/25 15:00		Kareem	OK
28	CCB02	CCB02	CCB	05/21/25 15:04		Kareem	OK
29	Q2065-01DL	VACTRUCK-728068D	SAM	05/21/25 15:09	2X For Still High	Kareem	Dilution
30	Q2090-02A	002-35TH-AVE(APR)	PS	05/21/25 15:13		Kareem	OK
31	Q2065-01DL2	VACTRUCK-728068D	SAM	05/21/25 15:18	Bad Injection	Kareem	Not Ok
32	Q1984-05	OU4-PCS-TC-35-050	SAM	05/21/25 15:24		Kareem	OK
33	Q1984-07	OU4-TS-24-050725	SAM	05/21/25 15:28		Kareem	OK
34	Q1984-09	OU4-TS-25-050725	SAM	05/21/25 15:37		Kareem	OK
35	Q1984-11	OU4-TS-26-050725	SAM	05/21/25 15:41		Kareem	OK
36	Q1984-13	OU4-TS-27-050725	SAM	05/21/25 15:45		Kareem	OK
37	Q2065-01DL3	VACTRUCK-728068D	SAM	05/21/25 15:53	20X For Ca	Kareem	Confirms
38	Q1984-15	OU4-TS-28-050725	SAM	05/21/25 15:58		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135868

Review By	Janvi	Review On	5/22/2025 2:51:30 PM
Supervise By	jaswal	Supervise On	5/22/2025 2:52:47 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546		
ICV Standard	MP85553		
CCV Standard	MP85556		
ICSA Standard	MP85554,MP85555		
CRI Standard	MP85552		
LCS Standard			
Chk Standard	MP85557,MP85558		

39	CCV03	CCV03	CCV	05/21/25 16:02		Kareem	OK
40	CCB03	CCB03	CCB	05/21/25 16:25		Kareem	OK
41	Q1984-15DUP	OU4-TS-28-050725D	DUP	05/21/25 16:29		Kareem	OK
42	Q1984-15L	OU4-TS-28-050725L	SD	05/21/25 16:34		Kareem	OK
43	Q1984-15MS	OU4-TS-28-050725M	MS	05/21/25 16:38		Kareem	OK
44	Q1984-15MSD	OU4-TS-28-050725M	MSD	05/21/25 16:42		Kareem	OK
45	Q1984-15A	OU4-TS-28-050725A	PS	05/21/25 16:46		Kareem	OK
46	PB168085BL	PB168085BL	MB	05/21/25 16:51		Kareem	OK
47	PB168085BS	PB168085BS	LCS	05/21/25 17:05		Kareem	OK
48	PB168087BL	PB168087BL	MB	05/21/25 17:10		Kareem	OK
49	PB168087BS	PB168087BS	LCS	05/21/25 17:15		Kareem	OK
50	PB168086BL	PB168086BL	MB	05/21/25 17:22		Kareem	OK
51	CCV04	CCV04	CCV	05/21/25 17:27		Kareem	OK
52	CCB04	CCB04	CCB	05/21/25 17:32		Kareem	OK
53	PB168086BS	PB168086BS	LCS	05/21/25 17:37		Kareem	OK
54	PB168107BL	PB168107BL	MB	05/21/25 17:41		Kareem	OK
55	PB168107BS	PB168107BS	LCS	05/21/25 17:45		Kareem	OK
56	PB168067TB	PB168067TB	MB	05/21/25 17:49		Kareem	OK
57	Q2071-04	L1-WC-7	SAM	05/21/25 17:53		Kareem	OK
58	Q2071-08	L1-WC-8	SAM	05/21/25 17:58		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135868

Review By	Janvi	Review On	5/22/2025 2:51:30 PM
Supervise By	jaswal	Supervise On	5/22/2025 2:52:47 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546		
ICV Standard	MP85553		
CCV Standard	MP85556		
ICSA Standard	MP85554,MP85555		
CRI Standard	MP85552		
LCS Standard			
Chk Standard	MP85557,MP85558		

59	Q2071-12	L1-WC-9	SAM	05/21/25 18:02		Kareem	OK
60	Q2071-16	L2-WC-5	SAM	05/21/25 18:07		Kareem	OK
61	Q2071-20	L2-WC-6	SAM	05/21/25 18:11		Kareem	OK
62	CCV05	CCV05	CCV	05/21/25 18:16		Kareem	OK
63	CCB05	CCB05	CCB	05/21/25 18:20		Kareem	OK
64	Q2080-06	PL-HRH-COMP-01	SAM	05/21/25 18:24		Kareem	OK
65	Q2080-12	PL-HRH-COMP-02	SAM	05/21/25 18:29		Kareem	OK
66	Q2080-12DUP	PL-HRH-COMP-02DUP	DUP	05/21/25 18:33		Kareem	OK
67	Q2080-12L	PL-HRH-COMP-02L	SD	05/21/25 18:37		Kareem	OK
68	Q2080-12MS	PL-HRH-COMP-02MS	MS	05/21/25 18:42		Kareem	OK
69	Q2080-12MSD	PL-HRH-COMP-02MSD	MSD	05/21/25 18:46		Kareem	OK
70	Q2080-12A	PL-HRH-COMP-02A	PS	05/21/25 18:50		Kareem	OK
71	Q2070-01	CONCRETE-PAD-051	SAM	05/21/25 18:54		Kareem	OK
72	Q2071-01	L1-WC-7	SAM	05/21/25 18:59		Kareem	OK
73	CCV06	CCV06	CCV	05/21/25 19:08		Kareem	OK
74	CCB06	CCB06	CCB	05/21/25 19:16		Kareem	OK
75	Q2071-05	L1-WC-8	SAM	05/21/25 19:20		Kareem	OK
76	Q2071-09	L1-WC-9	SAM	05/21/25 19:24		Kareem	OK
77	Q2071-13	L2-WC-5	SAM	05/21/25 19:28		Kareem	OK
78	Q2071-17	L2-WC-6	SAM	05/21/25 19:32		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135868

Review By	Janvi	Review On	5/22/2025 2:51:30 PM
Supervise By	jaswal	Supervise On	5/22/2025 2:52:47 PM

STD. NAME	STD REF.#
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546
ICV Standard	MP85553
CCV Standard	MP85556
ICSA Standard	MP85554,MP85555
CRI Standard	MP85552
LCS Standard	
Chk Standard	MP85557,MP85558

79	Q2074-01	TP-12	SAM	05/21/25 19:36		Kareem	OK
80	Q2074-02	TP-7	SAM	05/21/25 19:41		Kareem	OK
81	Q2074-03	TP-15	SAM	05/21/25 19:45		Kareem	OK
82	Q2074-04	TP-20	SAM	05/21/25 19:49		Kareem	OK
83	Q2074-05	TP-38	SAM	05/21/25 19:53		Kareem	OK
84	CCV07	CCV07	CCV	05/21/25 19:57		Kareem	OK
85	CCB07	CCB07	CCB	05/21/25 20:02		Kareem	OK
86	Q2074-06	TP-19	SAM	05/21/25 20:06		Kareem	OK
87	Q2074-07	TP-40	SAM	05/21/25 20:10		Kareem	OK
88	Q2074-08	TP-18	SAM	05/21/25 20:14		Kareem	OK
89	Q2080-01	PL-HRH-COMP-01	SAM	05/21/25 20:19	Qc Missing	Kareem	Not Ok
90	Q2080-07	PL-HRH-COMP-02	SAM	05/21/25 20:23		Kareem	OK
91	Q2084-01	OR-640-COMP-66	SAM	05/21/25 20:27		Kareem	OK
92	Q2084-01DUP	OR-640-COMP-66DU	DUP	05/21/25 20:31	Wrong Qc Analyzed	Kareem	Not Ok
93	Q2084-01L	OR-640-COMP-66L	SD	05/21/25 20:35	Wrong Qc Analyzed	Kareem	Not Ok
94	Q2084-01MS	OR-640-COMP-66MS	MS	05/21/25 20:39	Wrong Qc Analyzed	Kareem	Not Ok
95	CCV08	CCV08	CCV	05/21/25 20:43		Kareem	OK
96	CCB08	CCB08	CCB	05/21/25 20:48		Kareem	OK
97	Q2084-01MSD	OR-640-COMP-66MS	MSD	05/21/25 20:52	CCV09 Na Fail , Wrong Qc Analyzed	Kareem	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135868

Review By	Janvi	Review On	5/22/2025 2:51:30 PM
Supervise By	jaswal	Supervise On	5/22/2025 2:52:47 PM

STD. NAME	STD REF.#
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546
ICV Standard	MP85553
CCV Standard	MP85556
ICSA Standard	MP85554,MP85555
CRI Standard	MP85552
LCS Standard	
Chk Standard	MP85557,MP85558

98	Q2084-01A	OR-640-COMP-66A	PS	05/21/25 20:56	CCV09 Na Fail , Wrong Qc Analyzed	Kareem	Not Ok
99	Q2072-01	MW-2	SAM	05/21/25 21:00	CCV09 Na Fail	Kareem	Not Ok
100	Q2072-02	MW-3	SAM	05/21/25 21:04	CCV09 Na Fail	Kareem	Not Ok
101	Q2072-03	MW-4	SAM	05/21/25 21:08	CCV09 Na Fail	Kareem	Not Ok
102	Q2072-04	MW-6	SAM	05/21/25 21:13	CCV09 Na Fail	Kareem	Not Ok
103	Q2072-04DUP	MW-6DUP	DUP	05/21/25 21:17	CCV09 Na Fail	Kareem	Not Ok
104	Q2072-04L	MW-6L	SD	05/21/25 21:21	CCV09 Na Fail	Kareem	Not Ok
105	Q2072-04MS	MW-6MS	MS	05/21/25 21:25	CCV09 Na Fail	Kareem	Not Ok
106	CCV09	CCV09	CCV	05/21/25 21:29	Na Fail	Kareem	OK
107	CCB09	CCB09	CCB	05/21/25 21:34		Kareem	OK
108	Q2072-04MSD	MW-6MSD	MSD	05/21/25 21:38	CCV09 Na Fail	Kareem	Not Ok
109	Q2072-04A	MW-6A	PS	05/21/25 21:42	CCV09 Na Fail	Kareem	Not Ok
110	Q2072-05	FB	SAM	05/21/25 21:46	CCV09 Na Fail	Kareem	Not Ok
111	CCV10	CCV10	CCV	05/21/25 21:50		Kareem	OK
112	CCB10	CCB10	CCB	05/21/25 21:54		Kareem	OK