



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group **SDG No.:** Q1984
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1984 **SAS No.:** Q1984

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB05	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	05/12/2025	09:58	LB135733

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>Q1984</u>								
Contract: <u>NOBI03</u>		Lab Code: <u>CHEM</u>		Case No.: <u>Q1984</u>			SAS No.: <u>Q1984</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB07	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	05/12/2025	10:08	LB135733
CCB08	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	05/12/2025	10:39	LB135733
CCB09	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	05/12/2025	11:11	LB135733
CCB10	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	05/12/2025	11:39	LB135733
CCB11	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	05/12/2025	12:53	LB135733
CCB12	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	05/12/2025	13:27	LB135733

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>Q1984</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1984</u>	SAS No.: <u>Q1984</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	80.0	100	P	05/20/2025	12:44	Lb135855
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	05/20/2025	12:44	Lb135855
	Arsenic	20.0	+/-20.0	U	15.0	20.0	P	05/20/2025	12:44	Lb135855
	Barium	100	+/-100	U	25.0	100	P	05/20/2025	12:44	Lb135855
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	05/20/2025	12:44	Lb135855
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	05/20/2025	12:44	Lb135855
	Calcium	2000	+/-2000	U	500	2000	P	05/20/2025	12:44	Lb135855
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	05/20/2025	12:44	Lb135855
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	05/20/2025	12:44	Lb135855
	Copper	20.0	+/-20.0	U	16.0	20.0	P	05/20/2025	12:44	Lb135855
	Iron	100	+/-100	U	80.0	100	P	05/20/2025	12:44	Lb135855
	Lead	12.0	+/-12.0	U	9.60	12.0	P	05/20/2025	12:44	Lb135855
	Magnesium	2000	+/-2000	U	500	2000	P	05/20/2025	12:44	Lb135855
	Manganese	20.0	+/-20.0	U	15.0	20.0	P	05/20/2025	12:44	Lb135855
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	05/20/2025	12:44	Lb135855
	Potassium	2000	+/-2000	U	1600	2000	P	05/20/2025	12:44	Lb135855
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	05/20/2025	12:44	Lb135855
	Silver	10.0	+/-10.0	U	5.00	10.0	P	05/20/2025	12:44	Lb135855
	Sodium	2000	+/-2000	U	1000	2000	P	05/20/2025	12:44	Lb135855
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	05/20/2025	12:44	Lb135855
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	05/20/2025	12:44	Lb135855
	Zinc	40.0	+/-40.0	U	15.0	40.0	P	05/20/2025	12:44	Lb135855

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>Q1984</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1984</u>	SAS No.: <u>Q1984</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	80.0	100	P	05/20/2025	13:18	Lb135855
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	05/20/2025	13:18	Lb135855
	Arsenic	20.0	+/-20.0	U	15.0	20.0	P	05/20/2025	13:18	Lb135855
	Barium	100	+/-100	U	25.0	100	P	05/20/2025	13:18	Lb135855
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	05/20/2025	13:18	Lb135855
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	05/20/2025	13:18	Lb135855
	Calcium	2000	+/-2000	U	500	2000	P	05/20/2025	13:18	Lb135855
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	05/20/2025	13:18	Lb135855
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	05/20/2025	13:18	Lb135855
	Copper	20.0	+/-20.0	U	16.0	20.0	P	05/20/2025	13:18	Lb135855
	Iron	100	+/-100	U	80.0	100	P	05/20/2025	13:18	Lb135855
	Lead	12.0	+/-12.0	U	9.60	12.0	P	05/20/2025	13:18	Lb135855
	Magnesium	2000	+/-2000	U	500	2000	P	05/20/2025	13:18	Lb135855
	Manganese	20.0	+/-20.0	U	15.0	20.0	P	05/20/2025	13:18	Lb135855
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	05/20/2025	13:18	Lb135855
	Potassium	2000	+/-2000	U	1600	2000	P	05/20/2025	13:18	Lb135855
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	05/20/2025	13:18	Lb135855
	Silver	10.0	+/-10.0	U	5.00	10.0	P	05/20/2025	13:18	Lb135855
	Sodium	2000	+/-2000	U	1000	2000	P	05/20/2025	13:18	Lb135855
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	05/20/2025	13:18	Lb135855
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	05/20/2025	13:18	Lb135855
	Zinc	40.0	+/-40.0	U	15.0	40.0	P	05/20/2025	13:18	Lb135855
CCB02	Aluminum	100	+/-100	U	80.0	100	P	05/20/2025	14:04	Lb135855
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	05/20/2025	14:04	Lb135855
	Arsenic	20.0	+/-20.0	U	15.0	20.0	P	05/20/2025	14:04	Lb135855
	Barium	100	+/-100	U	25.0	100	P	05/20/2025	14:04	Lb135855
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	05/20/2025	14:04	Lb135855
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	05/20/2025	14:04	Lb135855
	Calcium	2000	+/-2000	U	500	2000	P	05/20/2025	14:04	Lb135855
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	05/20/2025	14:04	Lb135855
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	05/20/2025	14:04	Lb135855
	Copper	20.0	+/-20.0	U	16.0	20.0	P	05/20/2025	14:04	Lb135855
	Iron	100	+/-100	U	80.0	100	P	05/20/2025	14:04	Lb135855
	Lead	12.0	+/-12.0	U	9.60	12.0	P	05/20/2025	14:04	Lb135855
	Magnesium	2000	+/-2000	U	500	2000	P	05/20/2025	14:04	Lb135855
	Manganese	20.0	+/-20.0	U	15.0	20.0	P	05/20/2025	14:04	Lb135855
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	05/20/2025	14:04	Lb135855
	Potassium	2000	+/-2000	U	1600	2000	P	05/20/2025	14:04	Lb135855
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	05/20/2025	14:04	Lb135855

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>Q1984</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1984</u>	SAS No.: <u>Q1984</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	5.00	10.0	P	05/20/2025	14:04	Lb135855
	Sodium	2000	+/-2000	U	1000	2000	P	05/20/2025	14:04	Lb135855
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	05/20/2025	14:04	Lb135855
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	05/20/2025	14:04	Lb135855
	Zinc	40.0	+/-40.0	U	15.0	40.0	P	05/20/2025	14:04	Lb135855
CCB03	Aluminum	100	+/-100	U	80.0	100	P	05/20/2025	14:50	Lb135855
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	05/20/2025	14:50	Lb135855
	Arsenic	20.0	+/-20.0	U	15.0	20.0	P	05/20/2025	14:50	Lb135855
	Barium	100	+/-100	U	25.0	100	P	05/20/2025	14:50	Lb135855
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	05/20/2025	14:50	Lb135855
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	05/20/2025	14:50	Lb135855
	Calcium	2000	+/-2000	U	500	2000	P	05/20/2025	14:50	Lb135855
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	05/20/2025	14:50	Lb135855
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	05/20/2025	14:50	Lb135855
	Copper	20.0	+/-20.0	U	16.0	20.0	P	05/20/2025	14:50	Lb135855
	Iron	100	+/-100	U	80.0	100	P	05/20/2025	14:50	Lb135855
	Lead	12.0	+/-12.0	U	9.60	12.0	P	05/20/2025	14:50	Lb135855
	Magnesium	2000	+/-2000	U	500	2000	P	05/20/2025	14:50	Lb135855
	Manganese	20.0	+/-20.0	U	15.0	20.0	P	05/20/2025	14:50	Lb135855
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	05/20/2025	14:50	Lb135855
	Potassium	2000	+/-2000	U	1600	2000	P	05/20/2025	14:50	Lb135855
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	05/20/2025	14:50	Lb135855
	Silver	10.0	+/-10.0	U	5.00	10.0	P	05/20/2025	14:50	Lb135855
	Sodium	2000	+/-2000	U	1000	2000	P	05/20/2025	14:50	Lb135855
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	05/20/2025	14:50	Lb135855
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	05/20/2025	14:50	Lb135855
	Zinc	40.0	+/-40.0	U	15.0	40.0	P	05/20/2025	14:50	Lb135855
CCB04	Aluminum	100	+/-100	U	80.0	100	P	05/20/2025	15:50	Lb135855
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	05/20/2025	15:50	Lb135855
	Arsenic	20.0	+/-20.0	U	15.0	20.0	P	05/20/2025	15:50	Lb135855
	Barium	100	+/-100	U	25.0	100	P	05/20/2025	15:50	Lb135855
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	05/20/2025	15:50	Lb135855
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	05/20/2025	15:50	Lb135855
	Calcium	2000	+/-2000	U	500	2000	P	05/20/2025	15:50	Lb135855
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	05/20/2025	15:50	Lb135855
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	05/20/2025	15:50	Lb135855
	Copper	20.0	+/-20.0	U	16.0	20.0	P	05/20/2025	15:50	Lb135855
	Iron	100	+/-100	U	80.0	100	P	05/20/2025	15:50	Lb135855
	Lead	12.0	+/-12.0	U	9.60	12.0	P	05/20/2025	15:50	Lb135855



- 3a -

Client:	Nobis Group				SDG No.:	Q1984				
Contract:	NOBI03		Lab Code:	CHEM		Case No.:	Q1984		SAS No.:	Q1984
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	500	2000	P	05/20/2025	15:50	Lb135855
	Manganese	20.0	+/-20.0	U	15.0	20.0	P	05/20/2025	15:50	Lb135855
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	05/20/2025	15:50	Lb135855
	Potassium	2000	+/-2000	U	1600	2000	P	05/20/2025	15:50	Lb135855
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	05/20/2025	15:50	Lb135855
	Silver	10.0	+/-10.0	U	5.00	10.0	P	05/20/2025	15:50	Lb135855
	Sodium	2000	+/-2000	U	1000	2000	P	05/20/2025	15:50	Lb135855
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	05/20/2025	15:50	Lb135855
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	05/20/2025	15:50	Lb135855
	Zinc	40.0	+/-40.0	U	15.0	40.0	P	05/20/2025	15:50	Lb135855
CCB05	Aluminum	100	+/-100	U	80.0	100	P	05/20/2025	16:38	Lb135855
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	05/20/2025	16:38	Lb135855
	Arsenic	20.0	+/-20.0	U	15.0	20.0	P	05/20/2025	16:38	Lb135855
	Barium	100	+/-100	U	25.0	100	P	05/20/2025	16:38	Lb135855
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	05/20/2025	16:38	Lb135855
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	05/20/2025	16:38	Lb135855
	Calcium	2000	+/-2000	U	500	2000	P	05/20/2025	16:38	Lb135855
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	05/20/2025	16:38	Lb135855
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	05/20/2025	16:38	Lb135855
	Copper	20.0	+/-20.0	U	16.0	20.0	P	05/20/2025	16:38	Lb135855
CCB06	Iron	100	+/-100	U	80.0	100	P	05/20/2025	16:38	Lb135855
	Lead	12.0	+/-12.0	U	9.60	12.0	P	05/20/2025	16:38	Lb135855
	Magnesium	2000	+/-2000	U	500	2000	P	05/20/2025	16:38	Lb135855
	Manganese	20.0	+/-20.0	U	15.0	20.0	P	05/20/2025	16:38	Lb135855
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	05/20/2025	16:38	Lb135855
	Potassium	2000	+/-2000	U	1600	2000	P	05/20/2025	16:38	Lb135855
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	05/20/2025	16:38	Lb135855
	Silver	10.0	+/-10.0	U	5.00	10.0	P	05/20/2025	16:38	Lb135855
	Sodium	2000	+/-2000	U	1000	2000	P	05/20/2025	16:38	Lb135855
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	05/20/2025	16:38	Lb135855
CCB06	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	05/20/2025	16:38	Lb135855
	Zinc	40.0	+/-40.0	U	15.0	40.0	P	05/20/2025	16:38	Lb135855
	Aluminum	100	+/-100	U	80.0	100	P	05/20/2025	17:26	Lb135855
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	05/20/2025	17:26	Lb135855
	Arsenic	20.0	+/-20.0	U	15.0	20.0	P	05/20/2025	17:26	Lb135855
	Barium	100	+/-100	U	25.0	100	P	05/20/2025	17:26	Lb135855
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	05/20/2025	17:26	Lb135855
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	05/20/2025	17:26	Lb135855
	Calcium	2000	+/-2000	U	500	2000	P	05/20/2025	17:26	Lb135855

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>Q1984</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1984</u>	SAS No.: <u>Q1984</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	5.00	10.0	P	05/20/2025	17:26	Lb135855
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	05/20/2025	17:26	Lb135855
	Copper	20.0	+/-20.0	U	16.0	20.0	P	05/20/2025	17:26	Lb135855
	Iron	100	+/-100	U	80.0	100	P	05/20/2025	17:26	Lb135855
	Lead	12.0	+/-12.0	U	9.60	12.0	P	05/20/2025	17:26	Lb135855
	Magnesium	2000	+/-2000	U	500	2000	P	05/20/2025	17:26	Lb135855
	Manganese	20.0	+/-20.0	U	15.0	20.0	P	05/20/2025	17:26	Lb135855
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	05/20/2025	17:26	Lb135855
	Potassium	2000	+/-2000	U	1600	2000	P	05/20/2025	17:26	Lb135855
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	05/20/2025	17:26	Lb135855
	Silver	10.0	+/-10.0	U	5.00	10.0	P	05/20/2025	17:26	Lb135855
	Sodium	2000	+/-2000	U	1000	2000	P	05/20/2025	17:26	Lb135855
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	05/20/2025	17:26	Lb135855
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	05/20/2025	17:26	Lb135855
	Zinc	40.0	+/-40.0	U	15.0	40.0	P	05/20/2025	17:26	Lb135855
CCB07	Aluminum	100	+/-100	U	80.0	100	P	05/20/2025	18:13	Lb135855
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	05/20/2025	18:13	Lb135855
	Arsenic	20.0	+/-20.0	U	15.0	20.0	P	05/20/2025	18:13	Lb135855
	Barium	100	+/-100	U	25.0	100	P	05/20/2025	18:13	Lb135855
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	05/20/2025	18:13	Lb135855
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	05/20/2025	18:13	Lb135855
	Calcium	2000	+/-2000	U	500	2000	P	05/20/2025	18:13	Lb135855
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	05/20/2025	18:13	Lb135855
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	05/20/2025	18:13	Lb135855
	Copper	20.0	+/-20.0	U	16.0	20.0	P	05/20/2025	18:13	Lb135855
	Iron	100	+/-100	U	80.0	100	P	05/20/2025	18:13	Lb135855
	Lead	12.0	+/-12.0	U	9.60	12.0	P	05/20/2025	18:13	Lb135855
	Magnesium	2000	+/-2000	U	500	2000	P	05/20/2025	18:13	Lb135855
	Manganese	20.0	+/-20.0	U	15.0	20.0	P	05/20/2025	18:13	Lb135855
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	05/20/2025	18:13	Lb135855
	Potassium	2000	+/-2000	U	1600	2000	P	05/20/2025	18:13	Lb135855
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	05/20/2025	18:13	Lb135855
	Silver	10.0	+/-10.0	U	5.00	10.0	P	05/20/2025	18:13	Lb135855
	Sodium	2000	+/-2000	U	1000	2000	P	05/20/2025	18:13	Lb135855
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	05/20/2025	18:13	Lb135855
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	05/20/2025	18:13	Lb135855
	Zinc	40.0	+/-40.0	U	15.0	40.0	P	05/20/2025	18:13	Lb135855

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group **SDG No.:** Q1984
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1984 **SAS No.:** Q1984

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	80.0	100	P	05/21/2025	13:36	LB135868
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	05/21/2025	13:36	LB135868
	Arsenic	20.0	+/-20.0	U	15.0	20.0	P	05/21/2025	13:36	LB135868
	Barium	100	+/-100	U	25.0	100	P	05/21/2025	13:36	LB135868
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	05/21/2025	13:36	LB135868
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	05/21/2025	13:36	LB135868
	Calcium	2000	+/-2000	U	500	2000	P	05/21/2025	13:36	LB135868
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	05/21/2025	13:36	LB135868
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	05/21/2025	13:36	LB135868
	Copper	20.0	+/-20.0	U	16.0	20.0	P	05/21/2025	13:36	LB135868
	Iron	26.9	+/-100	J	80.0	100	P	05/21/2025	13:36	LB135868
	Lead	12.0	+/-12.0	U	9.60	12.0	P	05/21/2025	13:36	LB135868
	Magnesium	2000	+/-2000	U	500	2000	P	05/21/2025	13:36	LB135868
	Manganese	20.0	+/-20.0	U	15.0	20.0	P	05/21/2025	13:36	LB135868
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	05/21/2025	13:36	LB135868
	Potassium	2000	+/-2000	U	1600	2000	P	05/21/2025	13:36	LB135868
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	05/21/2025	13:36	LB135868
	Silver	10.0	+/-10.0	U	5.00	10.0	P	05/21/2025	13:36	LB135868
	Sodium	2000	+/-2000	U	1000	2000	P	05/21/2025	13:36	LB135868
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	05/21/2025	13:36	LB135868
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	05/21/2025	13:36	LB135868
	Zinc	40.0	+/-40.0	U	15.0	40.0	P	05/21/2025	13:36	LB135868

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>Q1984</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1984</u>	SAS No.: <u>Q1984</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	15.5	+/-100	J	80.0	100	P	05/21/2025	14:13	LB135868
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	05/21/2025	14:13	LB135868
	Arsenic	20.0	+/-20.0	U	15.0	20.0	P	05/21/2025	14:13	LB135868
	Barium	100	+/-100	U	25.0	100	P	05/21/2025	14:13	LB135868
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	05/21/2025	14:13	LB135868
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	05/21/2025	14:13	LB135868
	Calcium	2000	+/-2000	U	500	2000	P	05/21/2025	14:13	LB135868
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	05/21/2025	14:13	LB135868
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	05/21/2025	14:13	LB135868
	Copper	20.0	+/-20.0	U	16.0	20.0	P	05/21/2025	14:13	LB135868
	Iron	100	+/-100	U	80.0	100	P	05/21/2025	14:13	LB135868
	Lead	12.0	+/-12.0	U	9.60	12.0	P	05/21/2025	14:13	LB135868
	Magnesium	2000	+/-2000	U	500	2000	P	05/21/2025	14:13	LB135868
	Manganese	20.0	+/-20.0	U	15.0	20.0	P	05/21/2025	14:13	LB135868
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	05/21/2025	14:13	LB135868
	Potassium	2000	+/-2000	U	1600	2000	P	05/21/2025	14:13	LB135868
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	05/21/2025	14:13	LB135868
	Silver	10.0	+/-10.0	U	5.00	10.0	P	05/21/2025	14:13	LB135868
	Sodium	2000	+/-2000	U	1000	2000	P	05/21/2025	14:13	LB135868
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	05/21/2025	14:13	LB135868
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	05/21/2025	14:13	LB135868
	Zinc	40.0	+/-40.0	U	15.0	40.0	P	05/21/2025	14:13	LB135868
CCB02	Aluminum	100	+/-100	U	80.0	100	P	05/21/2025	15:04	LB135868
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	05/21/2025	15:04	LB135868
	Arsenic	20.0	+/-20.0	U	15.0	20.0	P	05/21/2025	15:04	LB135868
	Barium	100	+/-100	U	25.0	100	P	05/21/2025	15:04	LB135868
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	05/21/2025	15:04	LB135868
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	05/21/2025	15:04	LB135868
	Calcium	2000	+/-2000	U	500	2000	P	05/21/2025	15:04	LB135868
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	05/21/2025	15:04	LB135868
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	05/21/2025	15:04	LB135868
	Copper	20.0	+/-20.0	U	16.0	20.0	P	05/21/2025	15:04	LB135868
	Iron	100	+/-100	U	80.0	100	P	05/21/2025	15:04	LB135868
	Lead	12.0	+/-12.0	U	9.60	12.0	P	05/21/2025	15:04	LB135868
	Magnesium	2000	+/-2000	U	500	2000	P	05/21/2025	15:04	LB135868
	Manganese	20.0	+/-20.0	U	15.0	20.0	P	05/21/2025	15:04	LB135868
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	05/21/2025	15:04	LB135868
	Potassium	2000	+/-2000	U	1600	2000	P	05/21/2025	15:04	LB135868
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	05/21/2025	15:04	LB135868

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>Q1984</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1984</u>	SAS No.: <u>Q1984</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	5.00	10.0	P	05/21/2025	15:04	LB135868
	Sodium	2000	+/-2000	U	1000	2000	P	05/21/2025	15:04	LB135868
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	05/21/2025	15:04	LB135868
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	05/21/2025	15:04	LB135868
	Zinc	40.0	+/-40.0	U	15.0	40.0	P	05/21/2025	15:04	LB135868
CCB03	Aluminum	100	+/-100	U	80.0	100	P	05/21/2025	16:25	LB135868
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	05/21/2025	16:25	LB135868
	Arsenic	20.0	+/-20.0	U	15.0	20.0	P	05/21/2025	16:25	LB135868
	Barium	100	+/-100	U	25.0	100	P	05/21/2025	16:25	LB135868
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	05/21/2025	16:25	LB135868
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	05/21/2025	16:25	LB135868
	Calcium	2000	+/-2000	U	500	2000	P	05/21/2025	16:25	LB135868
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	05/21/2025	16:25	LB135868
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	05/21/2025	16:25	LB135868
	Copper	20.0	+/-20.0	U	16.0	20.0	P	05/21/2025	16:25	LB135868
	Iron	100	+/-100	U	80.0	100	P	05/21/2025	16:25	LB135868
	Lead	12.0	+/-12.0	U	9.60	12.0	P	05/21/2025	16:25	LB135868
	Magnesium	2000	+/-2000	U	500	2000	P	05/21/2025	16:25	LB135868
	Manganese	20.0	+/-20.0	U	15.0	20.0	P	05/21/2025	16:25	LB135868
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	05/21/2025	16:25	LB135868
	Potassium	2000	+/-2000	U	1600	2000	P	05/21/2025	16:25	LB135868
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	05/21/2025	16:25	LB135868
	Silver	10.0	+/-10.0	U	5.00	10.0	P	05/21/2025	16:25	LB135868
	Sodium	2000	+/-2000	U	1000	2000	P	05/21/2025	16:25	LB135868
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	05/21/2025	16:25	LB135868
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	05/21/2025	16:25	LB135868
	Zinc	40.0	+/-40.0	U	15.0	40.0	P	05/21/2025	16:25	LB135868
CCB04	Aluminum	13.8	+/-100	J	80.0	100	P	05/21/2025	17:32	LB135868
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	05/21/2025	17:32	LB135868
	Arsenic	20.0	+/-20.0	U	15.0	20.0	P	05/21/2025	17:32	LB135868
	Barium	100	+/-100	U	25.0	100	P	05/21/2025	17:32	LB135868
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	05/21/2025	17:32	LB135868
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	05/21/2025	17:32	LB135868
	Calcium	2000	+/-2000	U	500	2000	P	05/21/2025	17:32	LB135868
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	05/21/2025	17:32	LB135868
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	05/21/2025	17:32	LB135868
	Copper	20.0	+/-20.0	U	16.0	20.0	P	05/21/2025	17:32	LB135868
	Iron	100	+/-100	U	80.0	100	P	05/21/2025	17:32	LB135868
	Lead	12.0	+/-12.0	U	9.60	12.0	P	05/21/2025	17:32	LB135868



- 3a -

Client:	Nobis Group				SDG No.:	Q1984				
Contract:	NOBI03		Lab Code:	CHEM		Case No.:	Q1984		SAS No.:	Q1984
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	500	2000	P	05/21/2025	17:32	LB135868
	Manganese	20.0	+/-20.0	U	15.0	20.0	P	05/21/2025	17:32	LB135868
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	05/21/2025	17:32	LB135868
	Potassium	2000	+/-2000	U	1600	2000	P	05/21/2025	17:32	LB135868
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	05/21/2025	17:32	LB135868
	Silver	10.0	+/-10.0	U	5.00	10.0	P	05/21/2025	17:32	LB135868
	Sodium	2000	+/-2000	U	1000	2000	P	05/21/2025	17:32	LB135868
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	05/21/2025	17:32	LB135868
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	05/21/2025	17:32	LB135868
	Zinc	40.0	+/-40.0	U	15.0	40.0	P	05/21/2025	17:32	LB135868
CCB05	Aluminum	100	+/-100	U	80.0	100	P	05/21/2025	18:20	LB135868
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	05/21/2025	18:20	LB135868
	Arsenic	20.0	+/-20.0	U	15.0	20.0	P	05/21/2025	18:20	LB135868
	Barium	100	+/-100	U	25.0	100	P	05/21/2025	18:20	LB135868
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	05/21/2025	18:20	LB135868
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	05/21/2025	18:20	LB135868
	Calcium	2000	+/-2000	U	500	2000	P	05/21/2025	18:20	LB135868
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	05/21/2025	18:20	LB135868
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	05/21/2025	18:20	LB135868
	Copper	20.0	+/-20.0	U	16.0	20.0	P	05/21/2025	18:20	LB135868
CCB06	Iron	100	+/-100	U	80.0	100	P	05/21/2025	18:20	LB135868
	Lead	12.0	+/-12.0	U	9.60	12.0	P	05/21/2025	18:20	LB135868
	Magnesium	2000	+/-2000	U	500	2000	P	05/21/2025	18:20	LB135868
	Manganese	20.0	+/-20.0	U	15.0	20.0	P	05/21/2025	18:20	LB135868
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	05/21/2025	18:20	LB135868
	Potassium	2000	+/-2000	U	1600	2000	P	05/21/2025	18:20	LB135868
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	05/21/2025	18:20	LB135868
	Silver	10.0	+/-10.0	U	5.00	10.0	P	05/21/2025	18:20	LB135868
	Sodium	2000	+/-2000	U	1000	2000	P	05/21/2025	18:20	LB135868
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	05/21/2025	18:20	LB135868
CCB06	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	05/21/2025	18:20	LB135868
	Zinc	40.0	+/-40.0	U	15.0	40.0	P	05/21/2025	18:20	LB135868
	Aluminum	100	+/-100	U	80.0	100	P	05/21/2025	19:16	LB135868
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	05/21/2025	19:16	LB135868
	Arsenic	20.0	+/-20.0	U	15.0	20.0	P	05/21/2025	19:16	LB135868
	Barium	100	+/-100	U	25.0	100	P	05/21/2025	19:16	LB135868
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	05/21/2025	19:16	LB135868
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	05/21/2025	19:16	LB135868
	Calcium	2000	+/-2000	U	500	2000	P	05/21/2025	19:16	LB135868

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group		SDG No.: Q1984								
Contract:	NOBI03	Lab Code:	CHEM	Case No.:	Q1984	SAS No.:	Q1984			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	5.00	10.0	P	05/21/2025	19:16	LB135868
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	05/21/2025	19:16	LB135868
	Copper	20.0	+/-20.0	U	16.0	20.0	P	05/21/2025	19:16	LB135868
	Iron	100	+/-100	U	80.0	100	P	05/21/2025	19:16	LB135868
	Lead	12.0	+/-12.0	U	9.60	12.0	P	05/21/2025	19:16	LB135868
	Magnesium	2000	+/-2000	U	500	2000	P	05/21/2025	19:16	LB135868
	Manganese	20.0	+/-20.0	U	15.0	20.0	P	05/21/2025	19:16	LB135868
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	05/21/2025	19:16	LB135868
	Potassium	2000	+/-2000	U	1600	2000	P	05/21/2025	19:16	LB135868
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	05/21/2025	19:16	LB135868
	Silver	10.0	+/-10.0	U	5.00	10.0	P	05/21/2025	19:16	LB135868
	Sodium	2000	+/-2000	U	1000	2000	P	05/21/2025	19:16	LB135868
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	05/21/2025	19:16	LB135868
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	05/21/2025	19:16	LB135868
	Zinc	40.0	+/-40.0	U	15.0	40.0	P	05/21/2025	19:16	LB135868
CCB07	Aluminum	100	+/-100	U	80.0	100	P	05/21/2025	20:02	LB135868
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	05/21/2025	20:02	LB135868
	Arsenic	20.0	+/-20.0	U	15.0	20.0	P	05/21/2025	20:02	LB135868
	Barium	100	+/-100	U	25.0	100	P	05/21/2025	20:02	LB135868
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	05/21/2025	20:02	LB135868
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	05/21/2025	20:02	LB135868
	Calcium	2000	+/-2000	U	500	2000	P	05/21/2025	20:02	LB135868
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	05/21/2025	20:02	LB135868
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	05/21/2025	20:02	LB135868
	Copper	20.0	+/-20.0	U	16.0	20.0	P	05/21/2025	20:02	LB135868
	Iron	100	+/-100	U	80.0	100	P	05/21/2025	20:02	LB135868
	Lead	12.0	+/-12.0	U	9.60	12.0	P	05/21/2025	20:02	LB135868
	Magnesium	2000	+/-2000	U	500	2000	P	05/21/2025	20:02	LB135868
	Manganese	20.0	+/-20.0	U	15.0	20.0	P	05/21/2025	20:02	LB135868
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	05/21/2025	20:02	LB135868
	Potassium	2000	+/-2000	U	1600	2000	P	05/21/2025	20:02	LB135868
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	05/21/2025	20:02	LB135868
	Silver	10.0	+/-10.0	U	5.00	10.0	P	05/21/2025	20:02	LB135868
	Sodium	2000	+/-2000	U	1000	2000	P	05/21/2025	20:02	LB135868
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	05/21/2025	20:02	LB135868
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	05/21/2025	20:02	LB135868
	Zinc	40.0	+/-40.0	U	15.0	40.0	P	05/21/2025	20:02	LB135868
CCB08	Aluminum	100	+/-100	U	80.0	100	P	05/21/2025	20:48	LB135868
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	05/21/2025	20:48	LB135868

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>Q1984</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1984</u>	SAS No.: <u>Q1984</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	20.0	+/-20.0	U	15.0	20.0	P	05/21/2025	20:48	LB135868
	Barium	100	+/-100	U	25.0	100	P	05/21/2025	20:48	LB135868
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	05/21/2025	20:48	LB135868
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	05/21/2025	20:48	LB135868
	Calcium	2000	+/-2000	U	500	2000	P	05/21/2025	20:48	LB135868
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	05/21/2025	20:48	LB135868
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	05/21/2025	20:48	LB135868
	Copper	20.0	+/-20.0	U	16.0	20.0	P	05/21/2025	20:48	LB135868
	Iron	100	+/-100	U	80.0	100	P	05/21/2025	20:48	LB135868
	Lead	12.0	+/-12.0	U	9.60	12.0	P	05/21/2025	20:48	LB135868
	Magnesium	2000	+/-2000	U	500	2000	P	05/21/2025	20:48	LB135868
	Manganese	20.0	+/-20.0	U	15.0	20.0	P	05/21/2025	20:48	LB135868
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	05/21/2025	20:48	LB135868
	Potassium	2000	+/-2000	U	1600	2000	P	05/21/2025	20:48	LB135868
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	05/21/2025	20:48	LB135868
	Silver	10.0	+/-10.0	U	5.00	10.0	P	05/21/2025	20:48	LB135868
	Sodium	2000	+/-2000	U	1000	2000	P	05/21/2025	20:48	LB135868
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	05/21/2025	20:48	LB135868
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	05/21/2025	20:48	LB135868
	Zinc	40.0	+/-40.0	U	15.0	40.0	P	05/21/2025	20:48	LB135868
CCB09	Aluminum	100	+/-100	U	80.0	100	P	05/21/2025	21:34	LB135868
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	05/21/2025	21:34	LB135868
	Arsenic	20.0	+/-20.0	U	15.0	20.0	P	05/21/2025	21:34	LB135868
	Barium	100	+/-100	U	25.0	100	P	05/21/2025	21:34	LB135868
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	05/21/2025	21:34	LB135868
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	05/21/2025	21:34	LB135868
	Calcium	2000	+/-2000	U	500	2000	P	05/21/2025	21:34	LB135868
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	05/21/2025	21:34	LB135868
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	05/21/2025	21:34	LB135868
	Copper	20.0	+/-20.0	U	16.0	20.0	P	05/21/2025	21:34	LB135868
	Iron	100	+/-100	U	80.0	100	P	05/21/2025	21:34	LB135868
	Lead	12.0	+/-12.0	U	9.60	12.0	P	05/21/2025	21:34	LB135868
	Magnesium	2000	+/-2000	U	500	2000	P	05/21/2025	21:34	LB135868
	Manganese	20.0	+/-20.0	U	15.0	20.0	P	05/21/2025	21:34	LB135868
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	05/21/2025	21:34	LB135868
	Potassium	2000	+/-2000	U	1600	2000	P	05/21/2025	21:34	LB135868
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	05/21/2025	21:34	LB135868
	Silver	10.0	+/-10.0	U	5.00	10.0	P	05/21/2025	21:34	LB135868
	Sodium	2000	+/-2000	U	1000	2000	P	05/21/2025	21:34	LB135868

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>Q1984</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1984</u>	SAS No.: <u>Q1984</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	40.0	+/-40.0	U	20.0	40.0	P	05/21/2025	21:34	LB135868
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	05/21/2025	21:34	LB135868
	Zinc	40.0	+/-40.0	U	15.0	40.0	P	05/21/2025	21:34	LB135868
CCB10	Aluminum	100	+/-100	U	80.0	100	P	05/21/2025	21:54	LB135868
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	05/21/2025	21:54	LB135868
	Arsenic	20.0	+/-20.0	U	15.0	20.0	P	05/21/2025	21:54	LB135868
	Barium	100	+/-100	U	25.0	100	P	05/21/2025	21:54	LB135868
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	05/21/2025	21:54	LB135868
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	05/21/2025	21:54	LB135868
	Calcium	2000	+/-2000	U	500	2000	P	05/21/2025	21:54	LB135868
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	05/21/2025	21:54	LB135868
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	05/21/2025	21:54	LB135868
	Copper	20.0	+/-20.0	U	16.0	20.0	P	05/21/2025	21:54	LB135868
	Iron	100	+/-100	U	80.0	100	P	05/21/2025	21:54	LB135868
	Lead	12.0	+/-12.0	U	9.60	12.0	P	05/21/2025	21:54	LB135868
	Magnesium	2000	+/-2000	U	500	2000	P	05/21/2025	21:54	LB135868
	Manganese	20.0	+/-20.0	U	15.0	20.0	P	05/21/2025	21:54	LB135868
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	05/21/2025	21:54	LB135868
	Potassium	2000	+/-2000	U	1600	2000	P	05/21/2025	21:54	LB135868
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	05/21/2025	21:54	LB135868
	Silver	10.0	+/-10.0	U	5.00	10.0	P	05/21/2025	21:54	LB135868
	Sodium	2000	+/-2000	U	1000	2000	P	05/21/2025	21:54	LB135868
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	05/21/2025	21:54	LB135868
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	05/21/2025	21:54	LB135868
	Zinc	40.0	+/-40.0	U	15.0	40.0	P	05/21/2025	21:54	LB135868

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Nobis Group</u>				SDG No.:	<u>Q1984</u>				
Contract:	<u>NOBI03</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1984</u>	SAS No.:	<u>Q1984</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number