

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Nobis Group **SDG No.:** Q1984
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1984 **SAS No.:** Q1984
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.18	0.2	89	70 - 130	CV	05/14/2025	11:54	LB135763
CRI	Antimony	1.93	2.0	96	70 - 130	P	05/21/2025	15:00	LB135886
	Arsenic	0.95	1.0	95	70 - 130	P	05/21/2025	15:00	LB135886
	Barium	9.45	10.0	94	70 - 130	P	05/21/2025	15:00	LB135886
	Beryllium	1.13	1.0	113	70 - 130	P	05/21/2025	15:00	LB135886
	Cadmium	0.83	1.0	83	70 - 130	P	05/21/2025	15:00	LB135886
	Chromium	2.03	2.0	102	70 - 130	P	05/21/2025	15:00	LB135886
	Copper	2.16	2.0	108	70 - 130	P	05/21/2025	15:00	LB135886
	Lead	0.83	1.0	83	70 - 130	P	05/21/2025	15:00	LB135886
	Nickel	0.95	1.0	95	70 - 130	P	05/21/2025	15:00	LB135886
	Selenium	4.84	5.0	97	70 - 130	P	05/21/2025	15:00	LB135886
	Silver	1.05	1.0	105	70 - 130	P	05/21/2025	15:00	LB135886
	Thallium	0.93	1.0	93	70 - 130	P	05/21/2025	15:00	LB135886
	Vanadium	5.02	5.0	100	70 - 130	P	05/21/2025	15:00	LB135886
	Zinc	5.25	5.0	105	50 - 150	P	05/21/2025	15:00	LB135886
CRI	Antimony	1.79	2.0	90	70 - 130	P	05/22/2025	14:01	LB135893
	Arsenic	0.91	1.0	91	70 - 130	P	05/22/2025	14:01	LB135893
	Barium	8.99	10.0	90	70 - 130	P	05/22/2025	14:01	LB135893
	Beryllium	1.04	1.0	104	70 - 130	P	05/22/2025	14:01	LB135893
	Cadmium	0.70	1.0	70	70 - 130	P	05/22/2025	14:01	LB135893
	Chromium	1.77	2.0	88	70 - 130	P	05/22/2025	14:01	LB135893
	Copper	1.80	2.0	90	70 - 130	P	05/22/2025	14:01	LB135893
	Lead	0.77	1.0	77	70 - 130	P	05/22/2025	14:01	LB135893
	Nickel	0.84	1.0	84	70 - 130	P	05/22/2025	14:01	LB135893
	Selenium	4.57	5.0	91	70 - 130	P	05/22/2025	14:01	LB135893
	Silver	0.97	1.0	97	70 - 130	P	05/22/2025	14:01	LB135893
	Thallium	0.91	1.0	91	70 - 130	P	05/22/2025	14:01	LB135893
	Vanadium	4.60	5.0	92	70 - 130	P	05/22/2025	14:01	LB135893
	Zinc	4.30	5.0	86	50 - 150	P	05/22/2025	14:01	LB135893