



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Nobis Group				SDG No.:	Q1984				
Contract:	NOBI03	Lab Code:	CHEM	Case No.:	Q1984	SAS No.:	Q1984			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB08	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	05/14/2025	11:44	LB135763

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group **SDG No.:** Q1984
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1984 **SAS No.:** Q1984

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB22	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	05/14/2025	11:51	LB135763
CCB23	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	05/14/2025	12:26	LB135763
CCB24	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	05/14/2025	12:54	LB135763
CCB25	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	05/14/2025	13:30	LB135763
CCB26	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	05/14/2025	13:57	LB135763
CCB27	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	05/14/2025	14:25	LB135763
CCB28	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	05/14/2025	14:41	LB135763
CCB29	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	05/14/2025	14:59	LB135763

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group SDG No.: Q1984
Contract: NOBI03 Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Antimony	2.00	+/-2.00	U	0.25	2.00	P	05/21/2025	14:37	LB135886
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	05/21/2025	14:37	LB135886
	Barium	10.0	+/-10.0	U	1.25	10.0	P	05/21/2025	14:37	LB135886
	Beryllium	1.00	+/-1.00	U	0.75	1.00	P	05/21/2025	14:37	LB135886
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	05/21/2025	14:37	LB135886
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	05/21/2025	14:37	LB135886
	Copper	2.00	+/-2.00	U	1.50	2.00	P	05/21/2025	14:37	LB135886
	Lead	1.00	+/-1.00	U	0.75	1.00	P	05/21/2025	14:37	LB135886
	Nickel	1.00	+/-1.00	U	0.75	1.00	P	05/21/2025	14:37	LB135886
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	05/21/2025	14:37	LB135886
	Silver	1.00	+/-1.00	U	0.50	1.00	P	05/21/2025	14:37	LB135886
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	05/21/2025	14:37	LB135886
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	05/21/2025	14:37	LB135886
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	05/21/2025	14:37	LB135886

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>Q1984</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1984</u>	SAS No.: <u>Q1984</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Antimony	2.00	+/-2.00	U	0.25	2.00	P	05/21/2025	14:57	LB135886
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	05/21/2025	14:57	LB135886
	Barium	10.0	+/-10.0	U	1.25	10.0	P	05/21/2025	14:57	LB135886
	Beryllium	1.00	+/-1.00	U	0.75	1.00	P	05/21/2025	14:57	LB135886
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	05/21/2025	14:57	LB135886
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	05/21/2025	14:57	LB135886
	Copper	2.00	+/-2.00	U	1.50	2.00	P	05/21/2025	14:57	LB135886
	Lead	1.00	+/-1.00	U	0.75	1.00	P	05/21/2025	14:57	LB135886
	Nickel	1.00	+/-1.00	U	0.75	1.00	P	05/21/2025	14:57	LB135886
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	05/21/2025	14:57	LB135886
	Silver	0.080	+/-1.00	J	0.50	1.00	P	05/21/2025	14:57	LB135886
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	05/21/2025	14:57	LB135886
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	05/21/2025	14:57	LB135886
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	05/21/2025	14:57	LB135886
CCB02	Antimony	2.00	+/-2.00	U	0.25	2.00	P	05/21/2025	16:05	LB135886
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	05/21/2025	16:05	LB135886
	Barium	10.0	+/-10.0	U	1.25	10.0	P	05/21/2025	16:05	LB135886
	Beryllium	1.00	+/-1.00	U	0.75	1.00	P	05/21/2025	16:05	LB135886
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	05/21/2025	16:05	LB135886
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	05/21/2025	16:05	LB135886
	Copper	2.00	+/-2.00	U	1.50	2.00	P	05/21/2025	16:05	LB135886
	Lead	1.00	+/-1.00	U	0.75	1.00	P	05/21/2025	16:05	LB135886
	Nickel	1.00	+/-1.00	U	0.75	1.00	P	05/21/2025	16:05	LB135886
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	05/21/2025	16:05	LB135886
	Silver	1.00	+/-1.00	U	0.50	1.00	P	05/21/2025	16:05	LB135886
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	05/21/2025	16:05	LB135886
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	05/21/2025	16:05	LB135886
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	05/21/2025	16:05	LB135886
CCB03	Antimony	2.00	+/-2.00	U	0.25	2.00	P	05/21/2025	17:28	LB135886
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	05/21/2025	17:28	LB135886
	Barium	10.0	+/-10.0	U	1.25	10.0	P	05/21/2025	17:28	LB135886
	Beryllium	1.00	+/-1.00	U	0.75	1.00	P	05/21/2025	17:28	LB135886
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	05/21/2025	17:28	LB135886
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	05/21/2025	17:28	LB135886
	Copper	2.00	+/-2.00	U	1.50	2.00	P	05/21/2025	17:28	LB135886
	Lead	1.00	+/-1.00	U	0.75	1.00	P	05/21/2025	17:28	LB135886
	Nickel	1.00	+/-1.00	U	0.75	1.00	P	05/21/2025	17:28	LB135886
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	05/21/2025	17:28	LB135886
	Silver	1.00	+/-1.00	U	0.50	1.00	P	05/21/2025	17:28	LB135886

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>Q1984</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1984</u>	SAS No.: <u>Q1984</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB03	Thallium	1.00	+/-1.00	U	0.50	1.00	P	05/21/2025	17:28	LB135886
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	05/21/2025	17:28	LB135886
	Zinc	2.43	+/-5.00	J	1.50	5.00	P	05/21/2025	17:28	LB135886
CCB04	Antimony	2.00	+/-2.00	U	0.25	2.00	P	05/21/2025	18:16	LB135886
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	05/21/2025	18:16	LB135886
	Barium	10.0	+/-10.0	U	1.25	10.0	P	05/21/2025	18:16	LB135886
	Beryllium	1.00	+/-1.00	U	0.75	1.00	P	05/21/2025	18:16	LB135886
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	05/21/2025	18:16	LB135886
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	05/21/2025	18:16	LB135886
	Copper	2.00	+/-2.00	U	1.50	2.00	P	05/21/2025	18:16	LB135886
	Lead	1.00	+/-1.00	U	0.75	1.00	P	05/21/2025	18:16	LB135886
	Nickel	1.00	+/-1.00	U	0.75	1.00	P	05/21/2025	18:16	LB135886
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	05/21/2025	18:16	LB135886
	Silver	1.00	+/-1.00	U	0.50	1.00	P	05/21/2025	18:16	LB135886
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	05/21/2025	18:16	LB135886
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	05/21/2025	18:16	LB135886
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	05/21/2025	18:16	LB135886

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group **SDG No.:** Q1984
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1984 **SAS No.:** Q1984

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Antimony	2.00	+/-2.00	U	0.25	2.00	P	05/22/2025	13:40	LB135893
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	05/22/2025	13:40	LB135893
	Barium	10.0	+/-10.0	U	1.25	10.0	P	05/22/2025	13:40	LB135893
	Beryllium	1.00	+/-1.00	U	0.75	1.00	P	05/22/2025	13:40	LB135893
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	05/22/2025	13:40	LB135893
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	05/22/2025	13:40	LB135893
	Copper	2.00	+/-2.00	U	1.50	2.00	P	05/22/2025	13:40	LB135893
	Lead	1.00	+/-1.00	U	0.75	1.00	P	05/22/2025	13:40	LB135893
	Nickel	1.00	+/-1.00	U	0.75	1.00	P	05/22/2025	13:40	LB135893
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	05/22/2025	13:40	LB135893
	Silver	1.00	+/-1.00	U	0.50	1.00	P	05/22/2025	13:40	LB135893
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	05/22/2025	13:40	LB135893
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	05/22/2025	13:40	LB135893
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	05/22/2025	13:40	LB135893

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Nobis Group</u>		SDG No.: <u>Q1984</u>								
Contract: <u>NOBI03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1984</u>	SAS No.: <u>Q1984</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Antimony	0.12	+/-2.00	J	0.25	2.00	P	05/22/2025	13:54	LB135893
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	05/22/2025	13:54	LB135893
	Barium	10.0	+/-10.0	U	1.25	10.0	P	05/22/2025	13:54	LB135893
	Beryllium	1.00	+/-1.00	U	0.75	1.00	P	05/22/2025	13:54	LB135893
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	05/22/2025	13:54	LB135893
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	05/22/2025	13:54	LB135893
	Copper	2.00	+/-2.00	U	1.50	2.00	P	05/22/2025	13:54	LB135893
	Lead	1.00	+/-1.00	U	0.75	1.00	P	05/22/2025	13:54	LB135893
	Nickel	1.00	+/-1.00	U	0.75	1.00	P	05/22/2025	13:54	LB135893
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	05/22/2025	13:54	LB135893
	Silver	1.00	+/-1.00	U	0.50	1.00	P	05/22/2025	13:54	LB135893
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	05/22/2025	13:54	LB135893
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	05/22/2025	13:54	LB135893
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	05/22/2025	13:54	LB135893
CCB02	Antimony	2.00	+/-2.00	U	0.25	2.00	P	05/22/2025	14:32	LB135893
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	05/22/2025	14:32	LB135893
	Barium	10.0	+/-10.0	U	1.25	10.0	P	05/22/2025	14:32	LB135893
	Beryllium	1.00	+/-1.00	U	0.75	1.00	P	05/22/2025	14:32	LB135893
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	05/22/2025	14:32	LB135893
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	05/22/2025	14:32	LB135893
	Copper	2.00	+/-2.00	U	1.50	2.00	P	05/22/2025	14:32	LB135893
	Lead	1.00	+/-1.00	U	0.75	1.00	P	05/22/2025	14:32	LB135893
	Nickel	1.00	+/-1.00	U	0.75	1.00	P	05/22/2025	14:32	LB135893
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	05/22/2025	14:32	LB135893
	Silver	1.00	+/-1.00	U	0.50	1.00	P	05/22/2025	14:32	LB135893
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	05/22/2025	14:32	LB135893
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	05/22/2025	14:32	LB135893
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	05/22/2025	14:32	LB135893