

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Nobis Group **Level:** LOW **SDG No.:** Q1984
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1984 **SAS No.:** Q1984
Matrix: Water **Sample ID:** Q1984-16 **Client ID:** OU4-TS-28-050725DUP
Percent Solids for Sample: NA **Duplicate ID** Q1984-16DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	Duplicate		RPD	Qual	M
				C	Result			
Antimony	ug/L	20	10.0	UD	10.0	UD		P
Arsenic	ug/L	20	24.5	D	22.9	D	7	P
Barium	ug/L	20	114	D	114	D	0	P
Beryllium	ug/L	20	13.6	D	13.2	D	3	P
Cadmium	ug/L	20	9.85	D	9.70	D	2	P
Chromium	ug/L	20	41.9	D	42.8	D	2	P
Copper	ug/L	20	298	D	291	D	2	P
Lead	ug/L	20	35.6	D	34.9	D	2	P
Nickel	ug/L	20	234	D	231	D	1	P
Mercury	ug/L	20	0.20	U	0.20	U		CV
Selenium	ug/L	20	25.0	UD	25.0	UD		P
Silver	ug/L	20	5.00	UD	5.00	UD		P
Thallium	ug/L	20	0.85	JD	0.80	JD	6	P
Vanadium	ug/L	20	69.1	D	68.3	D	1	P
Zinc	ug/L	20	1810	D	1750	D	3	P

“A control limit of $\pm 20\%$ RPD for each matrix applies for sample values greater than 10 times Detection Limit”

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Nobis Group **Level:** LOW **SDG No.:** Q1984
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1984 **SAS No.:** Q1984
Matrix: Water **Sample ID:** Q1984-16MS **Client ID:** OU4-TS-28-050725MSD
Percent Solids for Sample: NA **Duplicate ID** Q1984-16MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	Duplicate		RPD	Qual	M
				C	Result			
Antimony	ug/L	20	471	D	463	D	2	P
Arsenic	ug/L	20	21.3	D	20.5	D	4	P
Barium	ug/L	20	2330	D	2320	D	0	P
Beryllium	ug/L	20	516	D	510	D	1	P
Cadmium	ug/L	20	486	D	477	D	2	P
Chromium	ug/L	20	499	D	494	D	1	P
Copper	ug/L	20	5390	D	5300	D	2	P
Lead	ug/L	20	2480	D	2450	D	1	P
Nickel	ug/L	20	709	D	699	D	1	P
Mercury	ug/L	20	3.76		3.21		16	CV
Selenium	ug/L	20	431	D	432	D	0	P
Silver	ug/L	20	81.2	D	79.8	D	2	P
Thallium	ug/L	20	490	D	475	D	3	P
Vanadium	ug/L	20	554	D	550	D	1	P
Zinc	ug/L	20	6590	D	6490	D	2	P

“A control limit of $\pm 20\%$ RPD for each matrix applies for sample values greater than 10 times Detection Limit”