

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group **SDG No.:** Q1984
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1984 **SAS No.:** Q1984

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167960BS							
Antimony	ug/L	500	516		103	85 - 117	P
Arsenic	ug/L	500	515		103	84 - 116	P
Barium	ug/L	2500	2580		103	86 - 114	P
Beryllium	ug/L	500	518		104	83 - 121	P
Cadmium	ug/L	500	528		106	87 - 115	P
Chromium	ug/L	500	524		105	85 - 116	P
Copper	ug/L	5000	5230		105	85 - 118	P
Lead	ug/L	2500	2500		100	88 - 115	P
Nickel	ug/L	500	495		99	85 - 117	P
Selenium	ug/L	500	505		101	80 - 120	P
Silver	ug/L	500	522		104	85 - 116	P
Thallium	ug/L	500	500		100	82 - 116	P
Vanadium	ug/L	500	516		103	86 - 115	P
Zinc	ug/L	5000	5300		106	83 - 119	P



Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group **SDG No.:** Q1984
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1984 **SAS No.:** Q1984

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167978BS Mercury	ug/L	4.0	3.69		92	82 - 119	CV
