

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB167996BS

Contract: NOBI03

Lab Code: CHEM **Case No.:** Q1984

SAS No.: Q1984 **SDG NO.:** Q1984

Lab Sample ID: PB167996BS

Date(s) Analyzed: 05/19/2025 05/19/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP **ID:** 0.32 (mm)

GC Column:(2): RTX-CLP2 **ID:** 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4,5-T	1	9.13	9.08	9.18	178	2.3
	2	9.92	9.87	9.97	174	
2,4,5-TP(Silvex)	1	8.85	8.80	8.90	177	1.1
	2	9.52	9.47	9.57	175	
2,4-D	1	8.01	7.96	8.06	174	1.2
	2	8.64	8.59	8.69	172	
2,4-DB	1	9.69	9.64	9.74	188	21.2
	2	10.48	10.43	10.53	152	
Dalapon	1	2.45	2.40	2.50	168	3
	2	2.52	2.47	2.57	163	
DICHLORPROP	1	7.79	7.74	7.84	172	2.9
	2	8.33	8.28	8.38	167	
Dinoseb	1	10.85	10.80	10.90	175	2.3
	2	10.85	10.80	10.90	171	
DICAMBA	1	7.11	7.06	7.16	173	2.3
	2	7.64	7.59	7.69	169	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

TP-9MS

Contract: NOBI03

Lab Code: CHEM Case No.: Q1984

SAS No.: Q1984 SDG NO.: Q1984

Lab Sample ID: Q1982-08MS

Date(s) Analyzed: 05/15/2025 05/15/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP ID: 0.32 (mm)

GC Column: (2): RTX-CLP2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Dalapon	1	2.44	2.39	2.49	140	64.1
	2	2.52	2.47	2.57	272	
DICHLORPROP	1	7.79	7.74	7.84	170	1.7
	2	8.33	8.28	8.38	173	
2,4-D	1	8.01	7.96	8.06	181	1.7
	2	8.64	8.59	8.69	178	
2,4,5-TP(Silvex)	1	8.85	8.80	8.90	161	4.3
	2	9.52	9.47	9.57	168	
2,4,5-T	1	9.13	9.08	9.18	154	5.7
	2	9.92	9.87	9.97	163	
2,4-DB	1	9.69	9.64	9.74	128	21
	2	10.48	10.43	10.53	158	
Dinoseb	1	10.85	10.80	10.90	14.3	142.9
	2	10.83	10.78	10.88	85.8	
DICAMBA	1	7.11	7.06	7.16	152	0.7
	2	7.64	7.59	7.69	153	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

TP-9MSD

Contract: NOBI03

Lab Code: CHEM Case No.: Q1984

SAS No.: Q1984

SDG NO.: Q1984

Lab Sample ID: Q1982-08MSD

Date(s) Analyzed: 05/15/2025 05/15/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP ID: 0.32 (mm)

GC Column: (2): RTX-CLP2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Dalapon	1	2.44	2.39	2.49	160	64.7
	2	2.52	2.47	2.57	313	
DICHLORPROP	1	7.79	7.74	7.84	198	1.5
	2	8.33	8.28	8.38	195	
2,4-D	1	8.01	7.96	8.06	207	1.5
	2	8.64	8.59	8.69	204	
2,4,5-TP(Silvex)	1	8.85	8.80	8.90	182	7.4
	2	9.52	9.47	9.57	196	
2,4,5-T	1	9.13	9.08	9.18	175	7.7
	2	9.92	9.87	9.97	189	
2,4-DB	1	9.69	9.64	9.74	145	25.3
	2	10.48	10.43	10.53	187	
DICAMBA	1	7.11	7.06	7.16	174	3.4
	2	7.64	7.59	7.69	180	