

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984 SDG NO.: Q1984

Continuing Calib Date: 05/15/2025 Initial Calibration Date(s): 05/12/2025 05/12/2025

Continuing Calib Time: 10:18 Initial Calibration Time(s): 12:30 14:06

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.11	7.11	7.01	7.21	0.00
2,4-DCAA	6.93	6.93	6.83	7.03	0.00
Dalapon	2.45	2.45	2.35	2.55	0.01
DICHLORPROP	7.79	7.79	7.69	7.89	0.00
2,4-D	8.01	8.01	7.91	8.11	0.00
2,4,5-TP(Silvex)	8.85	8.85	8.75	8.95	0.00
2,4,5-T	9.13	9.13	9.03	9.23	0.00
2,4-DB	9.69	9.69	9.59	9.79	0.00
Dinoseb	10.85	10.85	10.75	10.95	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984 SDG NO.: Q1984

Continuing Calib Date: 05/15/2025 Initial Calibration Date(s): 05/12/2025 05/12/2025

Continuing Calib Time: 10:18 Initial Calibration Time(s): 12:30 14:06

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.64	7.64	7.54	7.74	0.00
2,4-DCAA	7.45	7.45	7.35	7.55	0.00
Dalapon	2.52	2.52	2.42	2.62	0.00
DICHLORPROP	8.33	8.33	8.23	8.43	0.00
2,4-D	8.64	8.64	8.54	8.74	0.00
2,4,5-TP(Silvex)	9.52	9.52	9.42	9.62	0.00
2,4,5-T	9.92	9.93	9.83	10.03	0.01
2,4-DB	10.48	10.49	10.39	10.59	0.01
Dinoseb	10.85	10.86	10.76	10.96	0.01