

### CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984 SDG NO.: Q1984

Continuing Calib Date: 05/19/2025 Initial Calibration Date(s): 05/12/2025 05/12/2025

Continuing Calib Time: 21:26 Initial Calibration Time(s): 12:30 14:06

GC Column: RTX-CLP ID: 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|------------------|------------|-----------|-----------|-------|------------|
|                  |            |           | FROM      | TO    |            |
| DICAMBA          | 7.11       | 7.11      | 7.01      | 7.21  | 0.00       |
| 2,4-DCAA         | 6.93       | 6.93      | 6.83      | 7.03  | 0.00       |
| Dalapon          | 2.45       | 2.45      | 2.35      | 2.55  | 0.01       |
| DICHLORPROP      | 7.79       | 7.79      | 7.69      | 7.89  | 0.00       |
| 2,4-D            | 8.01       | 8.01      | 7.91      | 8.11  | 0.00       |
| 2,4,5-TP(Silvex) | 8.85       | 8.85      | 8.75      | 8.95  | 0.00       |
| 2,4,5-T          | 9.13       | 9.13      | 9.03      | 9.23  | 0.00       |
| 2,4-DB           | 9.69       | 9.69      | 9.59      | 9.79  | 0.00       |
| Dinoseb          | 10.85      | 10.85     | 10.75     | 10.95 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984 SDG NO.: Q1984

Continuing Calib Date: 05/19/2025 Initial Calibration Date(s): 05/12/2025 05/12/2025

Continuing Calib Time: 21:26 Initial Calibration Time(s): 12:30 14:06

GC Column: RTX-CLP2 ID: 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|------------------|------------|-----------|-----------|-------|------------|
|                  |            |           | FROM      | TO    |            |
| DICAMBA          | 7.64       | 7.64      | 7.54      | 7.74  | 0.00       |
| 2,4-DCAA         | 7.45       | 7.45      | 7.35      | 7.55  | 0.00       |
| Dalapon          | 2.52       | 2.52      | 2.42      | 2.62  | 0.00       |
| DICHLORPROP      | 8.33       | 8.33      | 8.23      | 8.43  | 0.00       |
| 2,4-D            | 8.64       | 8.64      | 8.54      | 8.74  | 0.00       |
| 2,4,5-TP(Silvex) | 9.52       | 9.52      | 9.42      | 9.62  | 0.00       |
| 2,4,5-T          | 9.92       | 9.93      | 9.83      | 10.03 | 0.01       |
| 2,4-DB           | 10.48      | 10.49     | 10.39     | 10.59 | 0.01       |
| Dinoseb          | 10.85      | 10.86     | 10.76     | 10.96 | 0.01       |