

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984 SDG NO.: Q1984

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL02 Date Analyzed: 05/15/2025

Lab Sample No.: HSTDCCC750 Data File : PS030248.D Time Analyzed: 20:55

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.131	9.033	9.233	717.030	712.500	0.6
2,4,5-TP(Silvex)	8.849	8.750	8.950	713.490	712.500	0.1
2,4-D	8.009	7.910	8.110	700.350	705.000	-0.7
2,4-DB	9.689	9.589	9.789	764.870	712.500	7.4
2,4-DCAA	6.933	6.834	7.034	740.230	750.000	-1.3
Dalapon	2.446	2.346	2.546	669.050	682.500	-2.0
DICAMBA	7.109	7.010	7.210	704.680	705.000	0.0
DICHLORPROP	7.789	7.689	7.889	694.060	705.000	-1.6
Dinoseb	10.849	10.751	10.951	699.880	705.000	-0.7

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984 SDG NO.: Q1984

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL02 Date Analyzed: 05/15/2025

Lab Sample No.: HSTDCCC750 Data File : PS030248.D Time Analyzed: 20:55

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.924	9.826	10.026	722.850	712.500	1.5
2,4,5-TP(Silvex)	9.520	9.423	9.623	726.910	712.500	2.0
2,4-D	8.642	8.544	8.744	719.140	705.000	2.0
2,4-DB	10.481	10.385	10.585	644.470	712.500	-9.5
2,4-DCAA	7.452	7.354	7.554	751.940	750.000	0.3
Dalapon	2.520	2.420	2.620	671.680	682.500	-1.6
DICAMBA	7.637	7.539	7.739	723.820	705.000	2.7
DICHLORPROP	8.330	8.232	8.432	696.240	705.000	-1.2
Dinoseb	10.855	10.758	10.958	712.020	705.000	1.0