

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984 SDG NO.: Q1984

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL03 Date Analyzed: 05/16/2025

Lab Sample No.: HSTDCCC750 Data File : PS030256.D Time Analyzed: 00:08

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.130	9.033	9.233	702.140	712.500	-1.5
2,4,5-TP(Silvex)	8.849	8.750	8.950	693.670	712.500	-2.6
2,4-D	8.008	7.910	8.110	691.170	705.000	-2.0
2,4-DB	9.688	9.589	9.789	759.510	712.500	6.6
2,4-DCAA	6.932	6.834	7.034	735.910	750.000	-1.9
Dalapon	2.445	2.346	2.546	680.410	682.500	-0.3
DICAMBA	7.109	7.010	7.210	702.910	705.000	-0.3
DICHLORPROP	7.788	7.689	7.889	685.810	705.000	-2.7
Dinoseb	10.850	10.751	10.951	669.500	705.000	-5.0

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984 SDG NO.: Q1984

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL03 Date Analyzed: 05/16/2025

Lab Sample No.: HSTDCCC750 Data File : PS030256.D Time Analyzed: 00:08

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.923	9.826	10.026	733.260	712.500	2.9
2,4,5-TP(Silvex)	9.520	9.423	9.623	736.130	712.500	3.3
2,4-D	8.641	8.544	8.744	721.030	705.000	2.3
2,4-DB	10.480	10.385	10.585	668.440	712.500	-6.2
2,4-DCAA	7.452	7.354	7.554	761.370	750.000	1.5
Dalapon	2.519	2.420	2.620	683.660	682.500	0.2
DICAMBA	7.637	7.539	7.739	736.130	705.000	4.4
DICHLORPROP	8.329	8.232	8.432	702.280	705.000	-0.4
Dinoseb	10.854	10.758	10.958	717.500	705.000	1.8