

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984 SDG NO.: Q1984

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL05 Date Analyzed: 05/19/2025

Lab Sample No.: HSTDCCC750 Data File : PS030288.D Time Analyzed: 21:26

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.128	9.033	9.233	720.910	712.500	1.2
2,4,5-TP(Silvex)	8.848	8.750	8.950	717.600	712.500	0.7
2,4-D	8.006	7.910	8.110	703.830	705.000	-0.2
2,4-DB	9.687	9.589	9.789	777.180	712.500	9.1
2,4-DCAA	6.931	6.834	7.034	740.950	750.000	-1.2
Dalapon	2.445	2.346	2.546	674.340	682.500	-1.2
DICAMBA	7.107	7.010	7.210	704.020	705.000	-0.1
DICHLORPROP	7.787	7.689	7.889	700.140	705.000	-0.7
Dinoseb	10.847	10.751	10.951	705.330	705.000	0.0

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984 SDG NO.: Q1984

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL05 Date Analyzed: 05/19/2025

Lab Sample No.: HSTDCCC750 Data File : PS030288.D Time Analyzed: 21:26

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.920	9.826	10.026	724.480	712.500	1.7
2,4,5-TP(Silvex)	9.517	9.423	9.623	729.750	712.500	2.4
2,4-D	8.639	8.544	8.744	708.970	705.000	0.6
2,4-DB	10.478	10.385	10.585	625.060	712.500	-12.3
2,4-DCAA	7.450	7.354	7.554	739.160	750.000	-1.4
Dalapon	2.519	2.420	2.620	670.310	682.500	-1.8
DICAMBA	7.635	7.539	7.739	717.500	705.000	1.8
DICHLORPROP	8.327	8.232	8.432	710.680	705.000	0.8
Dinoseb	10.852	10.758	10.958	713.460	705.000	1.2