

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1984

Analytical Method: 8151A

Client: Nobis Group

Datafile : PS030280.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167996BS (Column 1)	DICAMBA	166.5	173	ug/Kg	104				38	132	
	Dalapon	166.5	168	ug/Kg	101				70	130	
	DICHLORPROP	166.5	172	ug/Kg	103				28	155	
	2,4-D	166.5	174	ug/Kg	105				28	144	
	2,4,5-TP(Silvex)	166.5	177	ug/Kg	106				43	129	
	2,4,5-T	166.5	178	ug/Kg	107				31	138	
	2,4-DB	166.5	188	ug/Kg	113				34	142	
	Dinoseb	166.5	175	ug/Kg	105				57	152	
	DICAMBA	166.5	169	ug/Kg	102				38	132	
PB167996BS (Column 2)	Dalapon	166.5	163	ug/Kg	98				70	130	
	DICHLORPROP	166.5	167	ug/Kg	100				28	155	
	2,4-D	166.5	172	ug/Kg	103				28	144	
	2,4,5-TP(Silvex)	166.5	175	ug/Kg	105				43	129	
	2,4,5-T	166.5	174	ug/Kg	105				31	138	
	2,4-DB	166.5	152	ug/Kg	91				34	142	
	Dinoseb	166.5	171	ug/Kg	103				57	152	