

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984 SDG NO.: Q1984

Continuing Calib Date: 05/13/2025 Initial Calibration Date(s): 05/02/2025 05/02/2025

Continuing Calib Time: 09:08 Initial Calibration Time(s): 09:45 18:29

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.78	4.77	4.67	4.87	-0.01
Aroclor-1016-2	(2)	4.79	4.79	4.69	4.89	0.00
Aroclor-1016-3	(3)	4.85	4.85	4.75	4.95	0.00
Aroclor-1016-4	(4)	4.97	4.97	4.87	5.07	0.00
Aroclor-1016-5	(5)	5.23	5.23	5.13	5.33	0.00
Aroclor-1260-1	(1)	6.27	6.27	6.17	6.37	0.00
Aroclor-1260-2	(2)	6.46	6.46	6.36	6.56	0.00
Aroclor-1260-3	(3)	6.82	6.82	6.72	6.92	0.00
Aroclor-1260-4	(4)	7.08	7.08	6.98	7.18	0.00
Aroclor-1260-5	(5)	7.33	7.33	7.23	7.43	0.00
Tetrachloro-m-xylene		3.69	3.68	3.58	3.78	-0.01
Decachlorobiphenyl		8.73	8.73	8.63	8.83	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984 SDG NO.: Q1984

Continuing Calib Date: 05/13/2025 Initial Calibration Date(s): 05/02/2025 05/02/2025

Continuing Calib Time: 09:08 Initial Calibration Time(s): 09:45 18:29

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.76	4.76	4.66	4.86	0.00
Aroclor-1016-2	(2)	4.78	4.78	4.68	4.88	0.00
Aroclor-1016-3	(3)	4.95	4.95	4.85	5.05	0.00
Aroclor-1016-4	(4)	5.00	5.00	4.90	5.10	0.01
Aroclor-1016-5	(5)	5.21	5.21	5.11	5.31	0.00
Aroclor-1260-1	(1)	6.24	6.24	6.14	6.34	0.00
Aroclor-1260-2	(2)	6.43	6.43	6.33	6.53	0.00
Aroclor-1260-3	(3)	6.58	6.58	6.48	6.68	0.00
Aroclor-1260-4	(4)	7.05	7.05	6.95	7.15	0.00
Aroclor-1260-5	(5)	7.29	7.29	7.19	7.39	0.00
Tetrachloro-m-xylene		3.68	3.68	3.58	3.78	0.00
Decachlorobiphenyl		8.67	8.68	8.58	8.78	0.01