

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984 SDG NO.: Q1984

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/02/2025 05/02/2025

Client Sample No.: CCAL01 Date Analyzed: 05/12/2025

Lab Sample No.: AR1660CCC500 Data File : PO111005.D Time Analyzed: 09:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.779	4.673	4.873	436.510	500.000	-12.7
Aroclor-1016-2	4.797	4.692	4.892	443.860	500.000	-11.2
Aroclor-1016-3	4.855	4.748	4.948	419.940	500.000	-16.0
Aroclor-1016-4	4.974	4.868	5.068	451.560	500.000	-9.7
Aroclor-1016-5	5.232	5.125	5.325	439.380	500.000	-12.1
Aroclor-1260-1	6.271	6.165	6.365	410.080	500.000	-18.0
Aroclor-1260-2	6.460	6.355	6.555	405.250	500.000	-19.0
Aroclor-1260-3	6.828	6.722	6.922	427.730	500.000	-14.5
Aroclor-1260-4	7.087	6.982	7.182	425.430	500.000	-14.9
Aroclor-1260-5	7.330	7.225	7.425	423.880	500.000	-15.2
Decachlorobiphenyl	8.730	8.625	8.825	39.450	50.000	-21.1
Tetrachloro-m-xylene	3.689	3.583	3.783	49.390	50.000	-1.2

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984 SDG NO.: Q1984

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/02/2025 05/02/2025

Client Sample No.: CCAL01 Date Analyzed: 05/12/2025

Lab Sample No.: AR1660CCC500 Data File : PO111005.D Time Analyzed: 09:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.759	4.660	4.860	509.500	500.000	1.9
Aroclor-1016-2	4.777	4.679	4.879	513.910	500.000	2.8
Aroclor-1016-3	4.953	4.854	5.054	513.680	500.000	2.7
Aroclor-1016-4	4.995	4.896	5.096	516.840	500.000	3.4
Aroclor-1016-5	5.208	5.109	5.309	512.990	500.000	2.6
Aroclor-1260-1	6.238	6.139	6.339	503.590	500.000	0.7
Aroclor-1260-2	6.426	6.327	6.527	503.420	500.000	0.7
Aroclor-1260-3	6.579	6.480	6.680	507.610	500.000	1.5
Aroclor-1260-4	7.049	6.950	7.150	498.820	500.000	-0.2
Aroclor-1260-5	7.290	7.192	7.392	502.970	500.000	0.6
Decachlorobiphenyl	8.675	8.576	8.776	51.610	50.000	3.2
Tetrachloro-m-xylene	3.680	3.580	3.780	54.190	50.000	8.4