

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984 SDG NO.: Q1984

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/02/2025 05/02/2025

Client Sample No.: CCAL03 Date Analyzed: 05/13/2025

Lab Sample No.: AR1660CCC500 Data File : PO111026.D Time Analyzed: 09:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.775	4.673	4.873	450.880	500.000	-9.8
Aroclor-1016-2	4.794	4.692	4.892	462.430	500.000	-7.5
Aroclor-1016-3	4.851	4.748	4.948	435.760	500.000	-12.8
Aroclor-1016-4	4.971	4.868	5.068	471.860	500.000	-5.6
Aroclor-1016-5	5.228	5.125	5.325	458.360	500.000	-8.3
Aroclor-1260-1	6.268	6.165	6.365	421.560	500.000	-15.7
Aroclor-1260-2	6.457	6.355	6.555	419.450	500.000	-16.1
Aroclor-1260-3	6.824	6.722	6.922	436.640	500.000	-12.7
Aroclor-1260-4	7.083	6.982	7.182	429.160	500.000	-14.2
Aroclor-1260-5	7.326	7.225	7.425	425.180	500.000	-15.0
Decachlorobiphenyl	8.725	8.625	8.825	44.830	50.000	-10.3
Tetrachloro-m-xylene	3.686	3.583	3.783	50.700	50.000	1.4

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984 SDG NO.: Q1984

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/02/2025 05/02/2025

Client Sample No.: CCAL03 Date Analyzed: 05/13/2025

Lab Sample No.: AR1660CCC500 Data File : PO111026.D Time Analyzed: 09:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.759	4.660	4.860	523.220	500.000	4.6
Aroclor-1016-2	4.777	4.679	4.879	521.360	500.000	4.3
Aroclor-1016-3	4.953	4.854	5.054	525.860	500.000	5.2
Aroclor-1016-4	4.995	4.896	5.096	530.660	500.000	6.1
Aroclor-1016-5	5.208	5.109	5.309	522.190	500.000	4.4
Aroclor-1260-1	6.239	6.139	6.339	504.420	500.000	0.9
Aroclor-1260-2	6.426	6.327	6.527	503.080	500.000	0.6
Aroclor-1260-3	6.579	6.480	6.680	506.160	500.000	1.2
Aroclor-1260-4	7.049	6.950	7.150	491.110	500.000	-1.8
Aroclor-1260-5	7.290	7.192	7.392	485.360	500.000	-2.9
Decachlorobiphenyl	8.674	8.576	8.776	56.070	50.000	12.1
Tetrachloro-m-xylene	3.681	3.580	3.780	54.570	50.000	9.1