

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984 SDG NO.: Q1984

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/02/2025 05/02/2025

Client Sample No.: CCAL05 Date Analyzed: 05/14/2025

Lab Sample No.: AR1660CCC500 Data File : PO111047.D Time Analyzed: 09:14

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.772	4.673	4.873	480.640	500.000	-3.9
Aroclor-1016-2	4.791	4.692	4.892	488.430	500.000	-2.3
Aroclor-1016-3	4.849	4.748	4.948	461.470	500.000	-7.7
Aroclor-1016-4	4.969	4.868	5.068	498.870	500.000	-0.2
Aroclor-1016-5	5.225	5.125	5.325	482.510	500.000	-3.5
Aroclor-1260-1	6.264	6.165	6.365	446.320	500.000	-10.7
Aroclor-1260-2	6.455	6.355	6.555	445.410	500.000	-10.9
Aroclor-1260-3	6.821	6.722	6.922	469.000	500.000	-6.2
Aroclor-1260-4	7.081	6.982	7.182	460.370	500.000	-7.9
Aroclor-1260-5	7.324	7.225	7.425	457.360	500.000	-8.5
Decachlorobiphenyl	8.723	8.625	8.825	47.850	50.000	-4.3
Tetrachloro-m-xylene	3.683	3.583	3.783	55.660	50.000	11.3

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984 SDG NO.: Q1984

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/02/2025 05/02/2025

Client Sample No.: CCAL05 Date Analyzed: 05/14/2025

Lab Sample No.: AR1660CCC500 Data File : PO111047.D Time Analyzed: 09:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.758	4.660	4.860	552.490	500.000	10.5
Aroclor-1016-2	4.776	4.679	4.879	556.420	500.000	11.3
Aroclor-1016-3	4.951	4.854	5.054	556.580	500.000	11.3
Aroclor-1016-4	4.993	4.896	5.096	560.810	500.000	12.2
Aroclor-1016-5	5.206	5.109	5.309	551.850	500.000	10.4
Aroclor-1260-1	6.237	6.139	6.339	532.320	500.000	6.5
Aroclor-1260-2	6.425	6.327	6.527	529.310	500.000	5.9
Aroclor-1260-3	6.577	6.480	6.680	531.510	500.000	6.3
Aroclor-1260-4	7.047	6.950	7.150	514.110	500.000	2.8
Aroclor-1260-5	7.288	7.192	7.392	505.580	500.000	1.1
Decachlorobiphenyl	8.671	8.576	8.776	58.570	50.000	17.1
Tetrachloro-m-xylene	3.679	3.580	3.780	59.690	50.000	19.4