

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984 SDG NO.: Q1984

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/02/2025 05/02/2025

Client Sample No.: CCAL06 Date Analyzed: 05/14/2025

Lab Sample No.: AR1660CCC500 Data File : PO111054.D Time Analyzed: 12:33

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.773	4.673	4.873	467.820	500.000	-6.4
Aroclor-1016-2	4.793	4.692	4.892	462.680	500.000	-7.5
Aroclor-1016-3	4.849	4.748	4.948	434.400	500.000	-13.1
Aroclor-1016-4	4.969	4.868	5.068	474.450	500.000	-5.1
Aroclor-1016-5	5.226	5.125	5.325	466.230	500.000	-6.8
Aroclor-1260-1	6.265	6.165	6.365	427.130	500.000	-14.6
Aroclor-1260-2	6.455	6.355	6.555	423.330	500.000	-15.3
Aroclor-1260-3	6.822	6.722	6.922	437.730	500.000	-12.5
Aroclor-1260-4	7.081	6.982	7.182	430.380	500.000	-13.9
Aroclor-1260-5	7.324	7.225	7.425	443.120	500.000	-11.4
Decachlorobiphenyl	8.723	8.625	8.825	44.970	50.000	-10.1
Tetrachloro-m-xylene	3.684	3.583	3.783	52.750	50.000	5.5

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984 SDG NO.: Q1984

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/02/2025 05/02/2025

Client Sample No.: CCAL06 Date Analyzed: 05/14/2025

Lab Sample No.: AR1660CCC500 Data File : PO111054.D Time Analyzed: 12:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.757	4.660	4.860	536.700	500.000	7.3
Aroclor-1016-2	4.776	4.679	4.879	545.960	500.000	9.2
Aroclor-1016-3	4.952	4.854	5.054	541.050	500.000	8.2
Aroclor-1016-4	4.994	4.896	5.096	539.650	500.000	7.9
Aroclor-1016-5	5.207	5.109	5.309	534.440	500.000	6.9
Aroclor-1260-1	6.237	6.139	6.339	515.730	500.000	3.1
Aroclor-1260-2	6.425	6.327	6.527	509.970	500.000	2.0
Aroclor-1260-3	6.577	6.480	6.680	509.450	500.000	1.9
Aroclor-1260-4	7.048	6.950	7.150	486.660	500.000	-2.7
Aroclor-1260-5	7.289	7.192	7.392	480.350	500.000	-3.9
Decachlorobiphenyl	8.672	8.576	8.776	55.800	50.000	11.6
Tetrachloro-m-xylene	3.680	3.580	3.780	57.750	50.000	15.5