

### CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984 SDG NO.: Q1984

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL08 Date Analyzed: 05/13/2025

Lab Sample No.: AR1660CCC500 Data File : PP072018.D Time Analyzed: 16:35

| COMPOUND             | RT     | RT WINDOW<br>FROM TO |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|--------|----------------------|--------|--------------------|-------------------|------|
| Aroclor-1016-1       | 5.656  | 5.571                | 5.771  | 518.150            | 500.000           | 3.6  |
| Aroclor-1016-2       | 5.678  | 5.593                | 5.793  | 499.780            | 500.000           | 0.0  |
| Aroclor-1016-3       | 5.742  | 5.654                | 5.854  | 495.300            | 500.000           | -0.9 |
| Aroclor-1016-4       | 5.839  | 5.752                | 5.952  | 494.370            | 500.000           | -1.1 |
| Aroclor-1016-5       | 6.131  | 6.045                | 6.245  | 481.030            | 500.000           | -3.8 |
| Aroclor-1260-1       | 7.249  | 7.165                | 7.365  | 482.940            | 500.000           | -3.4 |
| Aroclor-1260-2       | 7.503  | 7.418                | 7.618  | 488.250            | 500.000           | -2.4 |
| Aroclor-1260-3       | 7.861  | 7.777                | 7.977  | 513.420            | 500.000           | 2.7  |
| Aroclor-1260-4       | 8.085  | 8.002                | 8.202  | 494.760            | 500.000           | -1.0 |
| Aroclor-1260-5       | 8.404  | 8.322                | 8.522  | 521.650            | 500.000           | 4.3  |
| Decachlorobiphenyl   | 10.212 | 10.137               | 10.337 | 52.800             | 50.000            | 5.6  |
| Tetrachloro-m-xylene | 4.505  | 4.417                | 4.617  | 50.480             | 50.000            | 1.0  |

### CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984 SDG NO.: Q1984

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL08 Date Analyzed: 05/13/2025

Lab Sample No.: AR1660CCC500 Data File : PP072018.D Time Analyzed: 16:35

| COMPOUND             | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|----------------------|-------|--------------------|-------------------|------|
| Aroclor-1016-1       | 4.885 | 4.798                | 4.998 | 536.020            | 500.000           | 7.2  |
| Aroclor-1016-2       | 4.904 | 4.816                | 5.016 | 582.240            | 500.000           | 16.4 |
| Aroclor-1016-3       | 5.081 | 4.993                | 5.193 | 526.540            | 500.000           | 5.3  |
| Aroclor-1016-4       | 5.122 | 5.036                | 5.236 | 533.710            | 500.000           | 6.7  |
| Aroclor-1016-5       | 5.337 | 5.250                | 5.450 | 561.140            | 500.000           | 12.2 |
| Aroclor-1260-1       | 6.370 | 6.285                | 6.485 | 573.210            | 500.000           | 14.6 |
| Aroclor-1260-2       | 6.558 | 6.473                | 6.673 | 564.270            | 500.000           | 12.9 |
| Aroclor-1260-3       | 6.711 | 6.627                | 6.827 | 559.120            | 500.000           | 11.8 |
| Aroclor-1260-4       | 7.181 | 7.098                | 7.298 | 568.790            | 500.000           | 13.8 |
| Aroclor-1260-5       | 7.422 | 7.339                | 7.539 | 570.140            | 500.000           | 14.0 |
| Decachlorobiphenyl   | 8.825 | 8.749                | 8.949 | 58.500             | 50.000            | 17.0 |
| Tetrachloro-m-xylene | 3.803 | 3.712                | 3.912 | 56.510             | 50.000            | 13.0 |