

Surrogate Summary

SDG No.: Q1984

Client: Nobis Group

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO110891.D	PIBLK-PO110891.D	Tetrachloro-m-xylene	1	20	16.1	81		60	140
		Decachlorobiphenyl	1	20	18.1	91		60	140
		Tetrachloro-m-xylene	2	20	15.2	76		60	140
		Decachlorobiphenyl	2	20	18.1	90		60	140
I.BLK-PO111009.D	PIBLK-PO111009.D	Tetrachloro-m-xylene	1	20	17.1	86		60	140
		Decachlorobiphenyl	1	20	14.3	72		60	140
		Tetrachloro-m-xylene	2	20	17.4	87		60	140
		Decachlorobiphenyl	2	20	17.9	89		60	140
PB167949BL	PB167949BL	Tetrachloro-m-xylene	1	20	22.2	111		44	130
		Decachlorobiphenyl	1	20	18.0	90		60	125
		Tetrachloro-m-xylene	2	20	22.3	111		44	130
		Decachlorobiphenyl	2	20	22.3	112		60	125
PB167949BS	PB167949BS	Tetrachloro-m-xylene	1	20	20.8	104		44	130
		Decachlorobiphenyl	1	20	18.4	92		60	125
		Tetrachloro-m-xylene	2	20	21.4	107		44	130
		Decachlorobiphenyl	2	20	22.5	112		60	125
Q1984-01	OU4-PCS-TC-33-050725	Tetrachloro-m-xylene	1	20	23.5	117		44	130
		Decachlorobiphenyl	1	20	17.3	87		60	125
		Tetrachloro-m-xylene	2	20	24.2	121		44	130
		Decachlorobiphenyl	2	20	21.6	108		60	125
Q1984-03	OU4-PCS-TC-34-050725	Tetrachloro-m-xylene	1	20	23.2	116		44	130
		Decachlorobiphenyl	1	20	16.2	81		60	125
		Tetrachloro-m-xylene	2	20	23.8	119		44	130
		Decachlorobiphenyl	2	20	20.8	104		60	125
Q1984-05	OU4-PCS-TC-35-050725	Tetrachloro-m-xylene	1	20	23.6	118		44	130
		Decachlorobiphenyl	1	20	16.7	84		60	125
		Tetrachloro-m-xylene	2	20	24.3	121		44	130
		Decachlorobiphenyl	2	20	21.6	108		60	125
Q1984-07	OU4-TS-24-050725	Tetrachloro-m-xylene	1	20	20.4	102		44	130
		Decachlorobiphenyl	1	20	12.4	62		60	125
		Tetrachloro-m-xylene	2	20	20.9	105		44	130
		Decachlorobiphenyl	2	20	16.1	80		60	125
Q1984-11	OU4-TS-26-050725	Tetrachloro-m-xylene	1	20	16.7	84		44	130
		Decachlorobiphenyl	1	20	6.77	34	*	60	125
		Tetrachloro-m-xylene	2	20	17.2	86		44	130
		Decachlorobiphenyl	2	20	9.69	48	*	60	125
Q1984-11MS	OU4-TS-26-050725MS	Tetrachloro-m-xylene	1	20	15.0	75		44	130
		Decachlorobiphenyl	1	20	7.63	38	*	60	125
		Tetrachloro-m-xylene	2	20	15.3	77		44	130
		Decachlorobiphenyl	2	20	10.8	54	*	60	125
Q1984-11MSD	OU4-TS-26-050725MSD	Tetrachloro-m-xylene	1	20	13.1	65		44	130

Surrogate Summary

SDG No.: Q1984

Client: Nobis Group

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q1984-11MSD	OU4-TS-26-050725MSD	Decachlorobiphenyl	1	20	6.71	34	*	60	125
		Tetrachloro-m-xylene	2	20	13.7	68		44	130
I.BLK-PO111024.D	PIBLK-PO111024.D	Decachlorobiphenyl	2	20	9.19	46	*	60	125
		Tetrachloro-m-xylene	1	20	17.9	89		60	140
		Decachlorobiphenyl	1	20	14.5	72		60	140
		Tetrachloro-m-xylene	2	20	17.9	89		60	140
I.BLK-PO111030.D	PIBLK-PO111030.D	Decachlorobiphenyl	2	20	18.6	93		60	140
		Tetrachloro-m-xylene	1	20	16.1	81		60	140
		Decachlorobiphenyl	1	20	13.2	66		60	140
		Tetrachloro-m-xylene	2	20	15.9	80		60	140
Q1984-09	OU4-TS-25-050725	Decachlorobiphenyl	2	20	16.3	82		60	140
		Tetrachloro-m-xylene	1	20	16.2	81		44	130
		Decachlorobiphenyl	1	20	14.2	71		60	125
		Tetrachloro-m-xylene	2	20	16.3	82		44	130
PB167972BL	PB167972BL	Decachlorobiphenyl	2	20	18.1	90		60	125
		Tetrachloro-m-xylene	1	20	20.6	103		44	130
		Decachlorobiphenyl	1	20	16.9	84		60	125
		Tetrachloro-m-xylene	2	20	20.4	102		44	130
PB167972BS	PB167972BS	Decachlorobiphenyl	2	20	20.4	102		60	125
		Tetrachloro-m-xylene	1	20	18.5	92		44	130
		Decachlorobiphenyl	1	20	16.4	82		60	125
		Tetrachloro-m-xylene	2	20	18.7	93		44	130
Q2017-01MS	MH-IMS	Decachlorobiphenyl	2	20	20.1	100		60	125
		Tetrachloro-m-xylene	1	20	19.8	99		44	130
		Decachlorobiphenyl	1	20	14.9	75		60	125
		Tetrachloro-m-xylene	2	20	20.6	103		44	130
Q2017-01MSD	MH-IMSD	Decachlorobiphenyl	2	20	19.0	95		60	125
		Tetrachloro-m-xylene	1	20	19.0	95		44	130
		Decachlorobiphenyl	1	20	14.2	71		60	125
		Tetrachloro-m-xylene	2	20	19.6	98		44	130
I.BLK-PO111045.D	PIBLK-PO111045.D	Decachlorobiphenyl	2	20	18.4	92		60	125
		Tetrachloro-m-xylene	1	20	16.4	82		60	140
		Decachlorobiphenyl	1	20	13.4	67		60	140
		Tetrachloro-m-xylene	2	20	16.2	81		60	140
I.BLK-PO111051.D	PIBLK-PO111051.D	Decachlorobiphenyl	2	20	16.5	83		60	140
		Tetrachloro-m-xylene	1	20	16.6	83		60	140
		Decachlorobiphenyl	1	20	13.8	69		60	140
		Tetrachloro-m-xylene	2	20	16.6	83		60	140
Q1984-15	OU4-TS-28-050725	Decachlorobiphenyl	2	20	16.9	84		60	140
		Tetrachloro-m-xylene	1	20	13.8	69		44	130
		Decachlorobiphenyl	1	20	18.3	91		60	125

Surrogate Summary

SDG No.: Q1984

Client: Nobis Group

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q1984-15	OU4-TS-28-050725	Tetrachloro-m-xylene	2	20	13.9	70		44	130
		Decachlorobiphenyl	2	20	22.9	115		60	125
I.BLK-PO111055.D	PIBLK-PO111055.D	Tetrachloro-m-xylene	1	20	16.8	84		60	140
		Decachlorobiphenyl	1	20	14.7	74		60	140
		Tetrachloro-m-xylene	2	20	17.2	86		60	140
		Decachlorobiphenyl	2	20	19.9	100		60	140
I.BLK-PP071388.D	PIBLK-PP071388.D	Tetrachloro-m-xylene	1	20	17.1	85		60	140
		Decachlorobiphenyl	1	20	17.6	88		60	140
		Tetrachloro-m-xylene	2	20	17.2	86		60	140
		Decachlorobiphenyl	2	20	17.6	88		60	140
I.BLK-PP072007.D	PIBLK-PP072007.D	Tetrachloro-m-xylene	1	20	17.1	85		60	140
		Decachlorobiphenyl	1	20	19.1	95		60	140
		Tetrachloro-m-xylene	2	20	19.7	98		60	140
		Decachlorobiphenyl	2	20	21.8	109		60	140
Q1984-13	OU4-TS-27-050725	Tetrachloro-m-xylene	1	20	14.6	73		44	130
		Decachlorobiphenyl	1	20	15.8	79		60	125
		Tetrachloro-m-xylene	2	20	18.0	90		44	130
		Decachlorobiphenyl	2	20	18.9	95		60	125
I.BLK-PP072022.D	PIBLK-PP072022.D	Tetrachloro-m-xylene	1	20	17.2	86		60	140
		Decachlorobiphenyl	1	20	19.0	95		60	140
		Tetrachloro-m-xylene	2	20	19.9	99		60	140
		Decachlorobiphenyl	2	20	21.2	106		60	140