

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: NOBI03

Lab Code: CHEM

SAS No.: Q1984

SDG NO.: Q1984

Lab File ID: BP024613.D

DFTPP Injection Date: 05/13/2025

Instrument ID: BNA_P

DFTPP Injection Time: 10:00

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	26.8
68	Less than 2.0% of mass 69	0.2 (0.6) 1
69	Mass 69 relative abundance	29
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	39.5
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.3
275	10.0 - 60.0% of mass 198	26.5
365	Greater than 1% of mass 198	4.3
441	Present, but less than mass 443	15.5
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.4 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BP024614.D	05/13/2025	10:41
SSTDICC005	SSTDICC005	BP024615.D	05/13/2025	11:22
SSTDICC010	SSTDICC010	BP024616.D	05/13/2025	12:03
SSTDICC020	SSTDICC020	BP024617.D	05/13/2025	12:43
SSTDICCC040	SSTDICCC040	BP024618.D	05/13/2025	13:24
SSTDICC050	SSTDICC050	BP024619.D	05/13/2025	14:05
SSTDICC060	SSTDICC060	BP024620.D	05/13/2025	14:45
SSTDICC080	SSTDICC080	BP024621.D	05/13/2025	15:26