

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: NOBI03
 Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984 SDG No.: Q1984
 Instrument ID: MSVOA_Y Calibration Date(s): 05/15/2025 05/15/2025
 Heated Purge: (Y/N) Y Calibration Time(s): 09:46 11:47
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VY022253.D		RRF010 = VY022254.D		RRF020 = VY022255.D		RRF050 = VY022256.D		RRF100 = VY022257.D		RRF150 = VY022258.D	
COMPOUND		RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD				
Dichlorodifluoromethane		0.546	0.587	0.541	0.507	0.470	0.479	0.522	8.5				
Chloromethane		1.212	1.237	1.153	1.042	1.046	0.894	1.097	11.7				
Vinyl Chloride		1.265	1.317	1.264	1.210	1.190	1.110	1.226	5.9				
Bromomethane		1.107	1.033	1.058	0.877	0.955	0.959	0.998	8.3				
Chloroethane		0.751	0.814	0.796	0.756	0.733	0.727	0.763	4.6				
Tetrahydrofuran		0.105	0.116	0.102	0.107	0.116	0.116	0.110	6				
Trichlorofluoromethane		1.254	1.292	1.260	1.214	1.211	1.217	1.241	2.6				
1,1,2-Trichlorotrifluoroethane		0.548	0.577	0.544	0.538	0.507	0.520	0.539	4.5				
1,1-Dichloroethene		0.545	0.544	0.540	0.539	0.518	0.523	0.535	2.2				
Acrylonitrile		0.123	0.126	0.115	0.120	0.127	0.126	0.123	3.8				
Acetone		0.126	0.114	0.096	0.093	0.094	0.092	0.102	13.6				
Carbon Disulfide		1.768	1.822	1.764	1.753	1.679	1.679	1.744	3.2				
Methyl tert-butyl Ether		1.463	1.523	1.393	1.459	1.504	1.507	1.475	3.2				
Methylene Chloride		0.797	0.811	0.646	0.597	0.570	0.552	0.662	17.3				
trans-1,2-Dichloroethene		0.577	0.609	0.603	0.592	0.580	0.581	0.590	2.2				
1,1-Dichloroethane		1.112	1.136	1.099	1.117	1.093	1.082	1.106	1.7				
2-Butanone		0.180	0.181	0.158	0.161	0.174	0.174	0.171	5.6				
Carbon Tetrachloride		0.485	0.524	0.494	0.516	0.509	0.524	0.509	3.2				
2,2-Dichloropropane		0.994	1.024	0.971	0.987	0.961	0.968	0.984	2.3				
cis-1,2-Dichloroethene		0.662	0.687	0.668	0.691	0.687	0.693	0.681	1.9				
Chloroform		1.032	1.079	1.067	1.079	1.054	1.057	1.061	1.7				
1,1,1-Trichloroethane		1.007	0.997	0.960	0.969	0.944	0.964	0.974	2.4				
1,1-Dichloropropene		0.474	0.502	0.488	0.505	0.484	0.496	0.492	2.4				
Benzene		1.369	1.462	1.435	1.499	1.482	1.496	1.457	3.4				
1,2-Dichloroethane		0.381	0.399	0.378	0.402	0.403	0.402	0.394	2.9				
Trichloroethene		0.358	0.366	0.361	0.371	0.356	0.363	0.362	1.5				
1,2-Dichloropropane		0.337	0.361	0.344	0.357	0.354	0.355	0.351	2.6				
Dibromomethane		0.189	0.192	0.180	0.189	0.193	0.194	0.189	2.6				
Bromodichloromethane		0.466	0.494	0.476	0.506	0.504	0.504	0.492	3.4				
4-Methyl-2-Pentanone		0.238	0.247	0.230	0.253	0.275	0.275	0.253	7.4				

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>CHEMTECH</u>	Contract:	<u>NOBI03</u>
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1984</u>
Instrument ID:	<u>MSVOA_Y</u>	SAS No.:	<u>Q1984</u>
Heated Purge: (Y/N)	<u>Y</u>	SDG No.:	<u>Q1984</u>
GC Column:	<u>RXI-624</u>	Calibration Date(s):	<u>05/15/2025</u>
ID:	<u>0.25 (mm)</u>	Calibration Time(s):	<u>09:46</u>
			<u>11:47</u>

LAB FILE ID:	RRF005 = VY022253.D	RRF010 = VY022254.D	RRF020 = VY022255.D					
	RRF050 = VY022256.D	RRF100 = VY022257.D	RRF150 = VY022258.D					
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Toluene	0.871	0.924	0.890	0.944	0.947	0.961	0.923	3.8
t-1,3-Dichloropropene	0.464	0.473	0.457	0.487	0.498	0.499	0.480	3.7
cis-1,3-Dichloropropene	0.523	0.548	0.535	0.562	0.562	0.568	0.550	3.2
1,1,2-Trichloroethane	0.241	0.243	0.232	0.243	0.248	0.249	0.243	2.4
1,3-Dichloropropane	0.418	0.436	0.413	0.438	0.442	0.440	0.431	2.8
2-Hexanone	0.167	0.167	0.151	0.166	0.183	0.185	0.170	7.4
Dibromochloromethane	0.296	0.314	0.298	0.323	0.329	0.330	0.315	4.9
1,2-Dibromoethane	0.221	0.228	0.214	0.227	0.234	0.232	0.226	3.2
Tetrachloroethene	0.474	0.481	0.470	0.474	0.446	0.429	0.462	4.4
Chlorobenzene	1.112	1.110	1.106	1.154	1.143	1.143	1.128	1.9
1,1,1,2-Tetrachloroethane	0.373	0.377	0.372	0.394	0.399	0.411	0.388	4.1
Ethyl Benzene	2.014	2.043	2.049	2.175	2.173	2.222	2.113	4.1
m/p-Xylenes	0.749	0.770	0.763	0.818	0.825	0.852	0.796	5.2
o-Xylene	0.704	0.741	0.723	0.768	0.782	0.802	0.753	4.9
Styrene	1.182	1.214	1.210	1.303	1.335	1.380	1.270	6.3
Bromoform	0.202	0.214	0.194	0.209	0.219	0.217	0.209	4.5
Isopropylbenzene	4.157	4.192	4.100	4.197	4.028	4.181	4.142	1.6
1,1,2,2-Tetrachloroethane	0.656	0.647	0.581	0.628	0.645	0.666	0.637	4.8
1,2,3-Trichloropropane	0.522	0.467	0.516	0.501	0.521	0.514	0.507	4.1
Bromobenzene	0.887	0.885	0.848	0.895	0.889	0.904	0.885	2.2
n-propylbenzene	5.031	5.047	4.904	5.127	4.865	5.006	4.996	1.9
2-Chlorotoluene	2.715	2.786	2.694	2.804	2.704	2.772	2.746	1.7
1,3,5-Trimethylbenzene	3.393	3.386	3.275	3.410	3.251	3.349	3.344	2
4-Chlorotoluene	2.801	2.829	2.744	2.907	2.797	2.867	2.824	2
tert-Butylbenzene	3.037	2.974	2.900	3.029	2.879	3.014	2.972	2.3
1,2,4-Trimethylbenzene	3.189	3.349	3.206	3.413	3.329	3.431	3.319	3.1
sec-Butylbenzene	4.390	4.461	4.318	4.573	4.307	4.421	4.412	2.2
p-Isopropyltoluene	3.598	3.610	3.573	3.818	3.754	3.911	3.710	3.7
1,3-Dichlorobenzene	1.757	1.774	1.715	1.838	1.839	1.914	1.806	4
1,4-Dichlorobenzene	1.759	1.765	1.706	1.750	1.723	1.720	1.737	1.4

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: NOBI03
 Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984 SDG No.: Q1984
 Instrument ID: MSVOA_Y Calibration Date(s): 05/15/2025 05/15/2025
 Heated Purge: (Y/N) Y Calibration Time(s): 09:46 11:47
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VY022253.D		RRF010 = VY022254.D		RRF020 = VY022255.D		
		RRF050 = VY022256.D		RRF100 = VY022257.D		RRF150 = VY022258.D		
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
n-Butylbenzene	3.419	3.500	3.491	3.668	3.452	3.533	3.510	2.5
1,2-Dichlorobenzene	1.526	1.520	1.479	1.534	1.520	1.523	1.517	1.3
1,2-Dibromo-3-Chloropropane	0.107	0.104	0.093	0.100	0.109	0.104	0.103	5.7
1,2,4-Trichlorobenzene	0.860	0.845	0.848	0.900	0.904	0.902	0.876	3.2
Hexachlorobutadiene	0.511	0.485	0.496	0.502	0.474	0.464	0.489	3.6
1,2,3-Trichlorobenzene	0.753	0.721	0.702	0.752	0.777	0.756	0.743	3.6
1,2-Dichloroethane-d4	0.547	0.559	0.527	0.541	0.545	0.560	0.546	2.2
Dibromofluoromethane	0.294	0.294	0.294	0.301	0.297	0.310	0.298	2.2
Toluene-d8	1.224	1.200	1.195	1.230	1.214	1.272	1.222	2.3
4-Bromofluorobenzene	0.390	0.386	0.368	0.387	0.387	0.407	0.387	3.3
trans-1,4-Dichloro-2-butene	0.269	0.260	0.241	0.237	0.246	0.248	0.250	4.8

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.