

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY051525

Review By	Mahesh Dadoda	Review On	5/16/2025 11:49:17 AM
Supervise By	Semsettin Yesilyurt	Supervise On	5/16/2025 11:51:38 AM
SubDirectory	VY051525	HP Acquire Method	MSVOA_Y HP Processing Method 82y051525s.m

STD. NAME	STD REF.#
Tune/Reschk	VP133916
Initial Calibration Stds	VP133918,VP133919,VP133920,VP133921,VP133922,VP133923
CCC	VP133925
Internal Standard/PEM	VP131783
ICV/I.BLK	VP133924
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022252.D	15 May 2025 07:52		SY/MD	Ok
2	VSTDIC005	VSTDIC005	VY022253.D	15 May 2025 09:46		SY/MD	Ok,M
3	VSTDIC010	VSTDIC010	VY022254.D	15 May 2025 10:16		SY/MD	Ok,M
4	VSTDIC020	VSTDIC020	VY022255.D	15 May 2025 10:39		SY/MD	Ok,M
5	VSTDIC050	VSTDIC050	VY022256.D	15 May 2025 11:02		SY/MD	Ok,M
6	VSTDIC100	VSTDIC100	VY022257.D	15 May 2025 11:24		SY/MD	Ok,M
7	VSTDIC150	VSTDIC150	VY022258.D	15 May 2025 11:47		SY/MD	Ok,M
8	VIBLK	VIBLK	VY022259.D	15 May 2025 12:34		SY/MD	Ok
9	VSTDICV050	ICVVY051525	VY022260.D	15 May 2025 13:20		SY/MD	Ok,M
10	VY0515SBL01	VY0515SBL01	VY022261.D	15 May 2025 13:44		SY/MD	Ok
11	VY0515SBS01	VY0515SBS01	VY022262.D	15 May 2025 14:07		SY/MD	Ok,M
12	VY0515SBSD01	VY0515SBSD01	VY022263.D	15 May 2025 14:29		SY/MD	Ok,M
13	Q1984-01	OU4-PCS-TC-33-05072	VY022264.D	15 May 2025 15:12	vial-A	SY/MD	Ok
14	Q1984-03	OU4-PCS-TC-34-05072	VY022265.D	15 May 2025 15:35	vial-A	SY/MD	Ok
15	Q1984-05	OU4-PCS-TC-35-05072	VY022266.D	15 May 2025 15:59	vial-A	SY/MD	Ok
16	Q1984-07	OU4-TS-24-050725	VY022267.D	15 May 2025 16:22	vial-A Surrogate fail	SY/MD	ReRun
17	Q1984-09	OU4-TS-25-050725	VY022268.D	15 May 2025 16:46	vial-A Surrogate fail	SY/MD	Not Ok
18	Q1984-11	OU4-TS-26-050725	VY022269.D	15 May 2025 17:09	vial-A Surrogate fail	SY/MD	Not Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY051525

Review By	Mahesh Dadoda	Review On	5/16/2025 11:49:17 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	5/16/2025 11:51:38 AM		
SubDirectory	VY051525	HP Acquire Method	MSVOA_Y	HP Processing Method	82y051525s.m
STD. NAME	STD REF.#				
Tune/Reschk	VP133916				
Initial Calibration Stds	VP133918,VP133919,VP133920,VP133921,VP133922,VP133923				
CCC	VP133925				
Internal Standard/PEM	VP131783				
ICV/I.BLK	VP133924				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q1984-13	OU4-TS-27-050725	VY022270.D	15 May 2025 17:33	vial-A Surrogate fail	SY/MD	Not Ok
20	Q1984-15	OU4-TS-28-050725	VY022271.D	15 May 2025 17:56	vial-A Surrogate fail	SY/MD	ReRun
21	Q1982-04	TP-4	VY022272.D	15 May 2025 18:19	vial-A	SY/MD	Ok
22	Q1982-05	TP-5	VY022273.D	15 May 2025 18:43	vial-A	SY/MD	Ok
23	Q1982-06	TP-6	VY022274.D	15 May 2025 19:06	vial-A	SY/MD	Ok
24	VSTDCCC050	VSTDCCC050EC	VY022275.D	15 May 2025 19:52		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY051625

Review By	Amit Patel	Review On	5/19/2025 1:35:43 PM
Supervise By	Maresh Dadoda	Supervise On	5/19/2025 2:39:18 PM
SubDirectory	VY051625	HP Acquire Method	MSVOA_Y HP Processing Method 82y051525s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133936
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133942,VP133943 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022276.D	16 May 2025 07:54		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022277.D	16 May 2025 08:24		SY/MD	Ok,M
3	VY0516SBL01	VY0516SBL01	VY022278.D	16 May 2025 09:37		SY/MD	Ok
4	VY0516SBS01	VY0516SBS01	VY022279.D	16 May 2025 10:09		SY/MD	Ok,M
5	VY0516SBSD01	VY0516SBSD01	VY022280.D	16 May 2025 10:32		SY/MD	Ok,M
6	Q1984-07RE	OU4-TS-24-050725RE	VY022281.D	16 May 2025 11:17	vial-B Surrogate fail	SY/MD	Confirms
7	Q1984-09	OU4-TS-25-050725	VY022282.D	16 May 2025 11:41	vial-B	SY/MD	Ok
8	Q1984-11	OU4-TS-26-050725	VY022283.D	16 May 2025 12:04	vial-B	SY/MD	Ok
9	Q1984-13	OU4-TS-27-050725	VY022284.D	16 May 2025 12:29	vial-B	SY/MD	Ok
10	Q1984-15RE	OU4-TS-28-050725RE	VY022285.D	16 May 2025 12:54	vial-B Surrogate fail	SY/MD	Confirms
11	Q1982-01	TP-1	VY022286.D	16 May 2025 13:17	vial-B	SY/MD	Ok
12	Q1982-02	TP-2B	VY022287.D	16 May 2025 13:41	vial-B	SY/MD	Ok
13	Q1982-03	TP-3	VY022288.D	16 May 2025 14:04	Vial A	SY/MD	Ok
14	Q1982-07	TP-8	VY022289.D	16 May 2025 14:28	Vial A	SY/MD	Ok
15	Q1982-08	TP-9	VY022290.D	16 May 2025 14:51	Vial A	SY/MD	Ok
16	Q2010-03	TP-14	VY022291.D	16 May 2025 15:15	vial-B	SY/MD	Ok
17	Q2010-04	TP-17	VY022292.D	16 May 2025 15:38	vial-B	SY/MD	Ok
18	Q2003-02	VOC	VY022293.D	16 May 2025 16:02	vial-A Internal standard fail	SY/MD	ReRun

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY051625

Review By	Amit Patel	Review On	5/19/2025 1:35:43 PM		
Supervise By	Mahesh Dadoda	Supervise On	5/19/2025 2:39:18 PM		
SubDirectory	VY051625	HP Acquire Method	MSVOA_Y	HP Processing Method	82y051525s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP133936			
CCC		VP133942,VP133943			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2032-03	TP-29-99	VY022294.D	16 May 2025 16:25	vial-A	SY/MD	Ok
20	Q2032-04	TP-24	VY022295.D	16 May 2025 16:48	vial-A	SY/MD	Ok
21	Q2032-07	TP-37	VY022296.D	16 May 2025 17:12	vial-A	SY/MD	Ok
22	Q2032-08	TP-32	VY022297.D	16 May 2025 17:35	vial-A	SY/MD	Ok
23	Q1984-19	OU4-TB01-050725	VY022298.D	16 May 2025 17:59	vial-A	SY/MD	Ok
24	Q2032-05MS	TP-24MS	VY022299.D	16 May 2025 18:21	vial-A Internal standard fail;Recovery Fail	SY/MD	Not Ok
25	Q2032-06MSD	TP-24MSD	VY022300.D	16 May 2025 18:44	vial-A Internal standard fail;Surrogate fail	SY/MD	Not Ok
26	VSTDCCC050	VSTDCCC050EC	VY022301.D	16 May 2025 19:29		SY/MD	Ok,M

M : Manual Integration