

Surrogate Summary

SDG No.: **Q1984**

Client: **Nobis Group**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1984-01	OU4-PCS-TC-33-050725	1,2-Dichloroethane-d4	50	44.4	89	71	136
		Dibromofluoromethane	50	48.9	98	78	119
		Toluene-d8	50	48.6	97	85	116
		4-Bromofluorobenzene	50	39.3	79	79	119
Q1984-03	OU4-PCS-TC-34-050725	1,2-Dichloroethane-d4	50	47.4	95	71	136
		Dibromofluoromethane	50	49.3	99	78	119
		Toluene-d8	50	49.7	99	85	116
		4-Bromofluorobenzene	50	43.3	87	79	119
Q1984-05	OU4-PCS-TC-35-050725	1,2-Dichloroethane-d4	50	48.0	96	71	136
		Dibromofluoromethane	50	49.2	98	78	119
		Toluene-d8	50	48.6	97	85	116
		4-Bromofluorobenzene	50	40.5	81	79	119
Q1984-07	OU4-TS-24-050725	1,2-Dichloroethane-d4	50	44.0	88	71	136
		Dibromofluoromethane	50	49.1	98	78	119
		Toluene-d8	50	48.3	96	85	116
		4-Bromofluorobenzene	50	31.9	64 *	79	119
Q1984-07RE	OU4-TS-24-050725RE	1,2-Dichloroethane-d4	50	44.7	89	71	136
		Dibromofluoromethane	50	48.5	97	78	119
		Toluene-d8	50	47.2	94	85	116
		4-Bromofluorobenzene	50	37.5	75 *	79	119
Q1984-09	OU4-TS-25-050725	1,2-Dichloroethane-d4	50	48.2	96	71	136
		Dibromofluoromethane	50	49.9	100	78	119
		Toluene-d8	50	49.0	98	85	116
		4-Bromofluorobenzene	50	39.7	79	79	119
Q1984-11	OU4-TS-26-050725	1,2-Dichloroethane-d4	50	54.0	108	71	136
		Dibromofluoromethane	50	51.6	103	78	119
		Toluene-d8	50	49.7	99	85	116
		4-Bromofluorobenzene	50	41.4	83	79	119
Q1984-13	OU4-TS-27-050725	1,2-Dichloroethane-d4	50	49.1	98	71	136
		Dibromofluoromethane	50	49.8	100	78	119
		Toluene-d8	50	48.9	98	85	116
		4-Bromofluorobenzene	50	40.1	80	79	119
Q1984-15	OU4-TS-28-050725	1,2-Dichloroethane-d4	50	48.5	97	71	136
		Dibromofluoromethane	50	49.9	100	78	119
		Toluene-d8	50	47.1	94	85	116
		4-Bromofluorobenzene	50	30.7	61 *	79	119
Q1984-15RE	OU4-TS-28-050725RE	1,2-Dichloroethane-d4	50	52.3	105	71	136
		Dibromofluoromethane	50	49.9	100	78	119
		Toluene-d8	50	48.9	98	85	116
		4-Bromofluorobenzene	50	38.8	78 *	79	119
Q1984-19	OU4-TB01-050725	1,2-Dichloroethane-d4	50	48.6	97	71	136
		Dibromofluoromethane	50	49.7	99	78	119
		Toluene-d8	50	48.9	98	85	116
		4-Bromofluorobenzene	50	39.9	80	79	119
VY0515SBL01	VY0515SBL01	1,2-Dichloroethane-d4	50	42.8	86	71	136
		Dibromofluoromethane	50	48.3	97	78	119
		Toluene-d8	50	48.7	97	85	116
		4-Bromofluorobenzene	50	39.8	80	79	119
VY0515SBS01	VY0515SBS01	1,2-Dichloroethane-d4	50	51.2	102	71	136
		Dibromofluoromethane	50	51.1	102	78	119

Surrogate Summary

SDG No.: Q1984

Client: Nobis Group

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VY0515SBS01	VY0515SBS01	Toluene-d8	50	51.3	103	85	116
		4-Bromofluorobenzene	50	50.2	100	79	119
VY0515SBSD01	VY0515SBSD01	1,2-Dichloroethane-d4	50	51.9	104	71	136
		Dibromofluoromethane	50	51.4	103	78	119
		Toluene-d8	50	51.2	102	85	116
		4-Bromofluorobenzene	50	50.4	101	79	119
VY0516SBL01	VY0516SBL01	1,2-Dichloroethane-d4	50	43.9	88	71	136
		Dibromofluoromethane	50	48.0	96	78	119
		Toluene-d8	50	48.8	98	85	116
		4-Bromofluorobenzene	50	48.5	97	79	119
VY0516SBS01	VY0516SBS01	1,2-Dichloroethane-d4	50	51.1	102	71	136
		Dibromofluoromethane	50	50.8	102	78	119
		Toluene-d8	50	50.9	102	85	116
		4-Bromofluorobenzene	50	49.4	99	79	119