

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: NOBI03
 Lab Code: CHEM Case No.: Q1984 SAS No.: Q1984 SDG NO.: Q1984
 Lab File ID: VY022277.D Date Analyzed: 05/16/2025
 Instrument ID: MSVOA_Y Time Analyzed: 08:24
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	144477	13.346				
UPPER LIMIT	288954	13.846				
LOWER LIMIT	72238.5	12.846				
EPA SAMPLE NO.						
OU4-TS-24-050725RE	87017	13.35				
OU4-TS-25-050725	145583	13.35				
OU4-TS-26-050725	157695	13.35				
OU4-TS-27-050725	152557	13.35				
OU4-TS-28-050725RE	125365	13.35				
OU4-TB01-050725	137120	13.35				
VY0516SBL01	151389	13.35				
VY0516SBS01	141210	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.