

Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB135711

Analyst By : jignesh
Supervisor Review By : rubina
Slope : 99.2
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/08/2025	16:00
2	CAL2	1	Water	NA	NA	20.3	7.00	05/08/2025	16:01
3	CAL3	1	Water	NA	NA	20.3	10.02	05/08/2025	16:05
4	ICV	1	Water	NA	NA	20.4	7.02	05/08/2025	16:11
5	CCV1	1	Water	NA	NA	20.3	2.01	05/08/2025	16:15
6	Q1985-01	1	Water	NA	NA	21.6	8.30	05/08/2025	16:20
7	Q1985-01DUP	1	Water	NA	NA	21.7	8.32	05/08/2025	16:22
8	Q1988-01	1	Water	NA	NA	20.1	5.57	05/08/2025	16:30
9	Q1989-01	1	Water	NA	NA	20.7	7.09	05/08/2025	16:35
10	Q1992-01	1	Water	NA	NA	22.1	5.20	05/08/2025	16:40
11	CCV2	1	Water	NA	NA	20.3	12.02	05/08/2025	16:45

WORKLIST(Hardcopy Internal Chain)

8135711

WorkList Name : ph q1992

WorkList ID : 189396

Department : Wet-Chemistry

Date : 05-08-2025 15:36:48

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1985-01	F RW8-BW-20250507	Water	pH	Cool 4 deg C	TETR06	L41	05/07/2025	9040C
Q1988-01	A 5725	Water	pH	Cool 4 deg C	PSEG03	L41	05/08/2025	9040C
Q1989-01	C 50225	Water	pH	Cool 4 deg C	PSEG03	L51	05/08/2025	9040C
Q1992-01	A 60453	Water	pH	Cool 4 deg C	PSEG03	L41	05/08/2025	9040C

Date/Time 05/08/25 15:45
Raw Sample Received by: SP WOL
Raw Sample Relinquished by: ap sm

Date/Time 05/08/25 18:00
Raw Sample Received by: ap sm
Raw Sample Relinquished by: SP WOL