

Duplicate Sample Summary

Client:	Tetra Tech NUS, Inc.	SDG No.:	Q1985
Project:	NWIRP Bethpage 112G08005-WE13	Sample ID:	Q1981-02
Client ID:	COMPDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	256		262		1	2.32		05/13/2025

Duplicate Sample Summary

Client:	Tetra Tech NUS, Inc.	SDG No.:	Q1985
Project:	NWIRP Bethpage 112G08005-WE13	Sample ID:	Q1985-01
Client ID:	RW8-BW-20250507DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
pH	pH	+/-20	8.30		8.32		1	0.24		05/08/2025