

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1985
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1985 **SAS No.:** Q1985

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167955BS							
Aluminum	ug/L	1000	914		91	86 - 115	P
Antimony	ug/L	400	382		96	88 - 113	P
Arsenic	ug/L	400	382		96	87 - 113	P
Barium	ug/L	100	107		107	88 - 113	P
Beryllium	ug/L	100	96.5		96	89 - 112	P
Cadmium	ug/L	100	97.5		98	88 - 113	P
Calcium	ug/L	500	461	J	92	87 - 113	P
Chromium	ug/L	200	190		95	90 - 113	P
Cobalt	ug/L	100	94.5		94	89 - 114	P
Copper	ug/L	150	145		97	86 - 114	P
Iron	ug/L	1500	1400		93	87 - 115	P
Lead	ug/L	500	473		95	86 - 113	P
Magnesium	ug/L	1000	872	J	87	85 - 113	P
Manganese	ug/L	100	92.3		92	90 - 114	P
Nickel	ug/L	250	240		96	88 - 113	P
Potassium	ug/L	5000	4550		91	86 - 114	P
Selenium	ug/L	1000	973		97	83 - 114	P
Silver	ug/L	37.5	34.3		92	84 - 115	P
Sodium	ug/L	1500	1330		89	87 - 115	P
Thallium	ug/L	1000	968		97	85 - 114	P
Vanadium	ug/L	150	139		93	90 - 111	P
Zinc	ug/L	100	94.6		95	87 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q1985

Contract: TETR06

Lab Code: CHEM

Case No.: Q1985

SAS No.: Q1985

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167980BS Mercury	ug/L	4.0	3.71		93	82 - 119	CV