

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1985 SAS No.: Q1985 SDG No.: Q1985

Instrument ID: BNA_N Calibration Date/Time: 05/14/2025 10:31

Lab File ID: BN037009.D Init. Calib. Date(s): 05/13/2025 05/13/2025

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 17:41 21:17

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.563	0.552		-2.0	20.0
Fluoranthene-d10	1.097	1.065		-2.9	20.0
2-Fluorophenol	1.048	1.078		2.9	20.0
Phenol-d6	1.311	1.303		-0.6	20.0
Nitrobenzene-d5	0.436	0.407		-6.7	20.0
2-Fluorobiphenyl	1.832	1.850		1.0	20.0
2,4,6-Tribromophenol	0.176	0.176		0.0	20.0
Terphenyl-d14	0.856	0.892		4.2	20.0
1,4-Dioxane	0.491	0.448		-8.8	20.0

All other compounds must meet a minimum RRF of 0.010.