

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q1985

SDG NO.: Q1985

Lab File ID: BF142383.D

DFTPP Injection Date: 05/15/2025

Instrument ID: BNA_F

DFTPP Injection Time: 09:09

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	31.9
68	Less than 2.0% of mass 69	0.5 (1.7) 1
69	Mass 69 relative abundance	26.6
70	Less than 2.0% of mass 69	0.1 (0.5) 1
127	10.0 - 80.0% of mass 198	36
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.4
275	10.0 - 60.0% of mass 198	23.5
365	Greater than 1% of mass 198	3
441	Present, but less than mass 443	15.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.1 (19.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF142384.D	05/15/2025	10:08
PB167951BL	PB167951BL	BF142385.D	05/15/2025	10:37
PB167951BS	PB167951BS	BF142386.D	05/15/2025	11:06
RW8-BW-20250507	Q1985-01	BF142387.D	05/15/2025	11:49
MW-3-20250508MS	Q1993-02MS	BF142388.D	05/15/2025	12:18
MW-3-20250508MSD	Q1993-03MSD	BF142389.D	05/15/2025	12:47
SSTDCCC040EC	SSTDCCC040	BF142391.D	05/15/2025	13:58