

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q1985 SAS No.: Q1985 SDG No.: Q1985
 Instrument ID: MSVOA_X Calibration Date(s): 05/05/2025 05/05/2025
 Heated Purge: (Y/N) N Calibration Time(s): 11:35 16:27
 GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:	RRF020 = VX046041.D		RRF050 = VX046042.D		RRF100 = VX046043.D		RRF150 = VX046044.D		RRF005 = VX046046.D		RRF001 = VX046047.D	
COMPOUND	RRF020	RRF050	RRF100	RRF150	RRF005	RRF001	RRF	% RSD				
Chloromethane	0.727	0.775	0.787	0.791	0.679	0.694	0.742	6.6				
Vinyl Chloride	0.660	0.710	0.727	0.755	0.619	0.673	0.691	7.2				
Bromomethane	0.296	0.326	0.340	0.334	0.305		0.320	5.8				
Chloroethane	0.354	0.378	0.329	0.317	0.368	0.467	0.369	14.4				
Trichlorofluoromethane	1.035	1.068	0.983	0.985	0.990	1.064	1.021	3.9				
1,1,2-Trichlorotrifluoroethane	0.628	0.641	0.629	0.648	0.610	0.633	0.632	2.1				
1,1-Dichloroethene	0.565	0.601	0.607	0.625	0.567	0.594	0.593	3.9				
Acetone	0.361	0.362	0.361	0.370	0.408	0.380	0.374	4.9				
Carbon Disulfide	1.295	1.455	1.522	1.597	1.141	1.423	1.406	11.7				
Methyl tert-butyl Ether	2.044	2.160	2.172	2.239	1.908	1.949	2.079	6.4				
Methylene Chloride	0.689	0.684	0.691	0.691	0.689	0.853	0.716	9.4				
trans-1,2-Dichloroethene	0.573	0.610	0.612	0.622	0.557	0.604	0.596	4.3				
1,1-Dichloroethane	1.233	1.263	1.263	1.286	1.154	1.116	1.219	5.6				
2-Butanone	0.540	0.555	0.558	0.569	0.539	0.495	0.543	4.8				
Carbon Tetrachloride	0.528	0.558	0.552	0.577	0.505	0.541	0.544	4.6				
cis-1,2-Dichloroethene	0.716	0.737	0.738	0.755	0.642	0.719	0.718	5.5				
Chloroform	1.287	1.296	1.277	1.300	1.199	1.265	1.271	3				
1,1,1-Trichloroethane	1.106	1.131	1.155	1.188	1.013	1.015	1.101	6.6				
Methylcyclohexane	0.596	0.641	0.627	0.658	0.587	0.627	0.623	4.3				
Benzene	1.426	1.474	1.441	1.477	1.337	1.348	1.417	4.3				
1,2-Dichloroethane	0.632	0.627	0.611	0.625	0.594	0.579	0.612	3.5				
Trichloroethene	0.344	0.355	0.345	0.362	0.315	0.324	0.341	5.3				
1,2-Dichloropropane	0.356	0.371	0.368	0.378	0.324	0.317	0.352	7.4				
Bromodichloromethane	0.557	0.577	0.573	0.594	0.498	0.485	0.547	8.2				
4-Methyl-2-Pentanone	0.620	0.634	0.630	0.631	0.555	0.561	0.605	6				
Toluene	0.884	0.898	0.885	0.904	0.838	0.803	0.869	4.5				
t-1,3-Dichloropropene	0.468	0.528	0.555	0.591	0.406	0.371	0.487	17.9				
cis-1,3-Dichloropropene	0.531	0.578	0.602	0.623	0.469	0.423	0.538	14.6				
1,1,2-Trichloroethane	0.349	0.354	0.351	0.356	0.337	0.308	0.343	5.3				
2-Hexanone	0.466	0.473	0.477	0.473	0.414	0.385	0.448	8.7				

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q1985 SAS No.: Q1985 SDG No.: Q1985
 Instrument ID: MSVOA_X Calibration Date(s): 05/05/2025 05/05/2025
 Heated Purge: (Y/N) N Calibration Time(s): 11:35 16:27
 GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF020 = VX046041.D		RRF050 = VX046042.D		RRF100 = VX046043.D			
		RRF150 = VX046044.D		RRF005 = VX046046.D		RRF001 = VX046047.D			
COMPOUND	RRF020	RRF050	RRF100	RRF150	RRF005	RRF001	RRF	% RSD	
Dibromochloromethane	0.378	0.400	0.415	0.431	0.326	0.306	0.376	13.3	
Tetrachloroethene	0.390	0.375	0.345	0.344	0.323	0.347	0.354	6.8	
Chlorobenzene	1.093	1.098	1.085	1.114	1.046	1.131	1.094	2.7	
Ethyl Benzene	1.919	2.022	1.979	2.036	1.816	1.803	1.929	5.2	
m/p-Xylenes	0.706	0.740	0.721	0.740	0.678	0.648	0.706	5.2	
o-Xylene	0.688	0.727	0.706	0.726	0.639	0.642	0.688	5.7	
Styrene	1.135	1.219	1.214	1.230	1.012	0.951	1.127	10.6	
Bromoform	0.270	0.304	0.312	0.327	0.236	0.234	0.281	14.2	
Isopropylbenzene	3.843	4.130	3.876	4.156	3.562	3.789	3.893	5.7	
1,1,2,2-Tetrachloroethane	1.315	1.338	1.284	1.345	1.350	1.552	1.364	7	
1,3-Dichlorobenzene	1.633	1.701	1.656	1.730	1.558	1.619	1.649	3.7	
1,4-Dichlorobenzene	1.629	1.693	1.639	1.722	1.606	1.817	1.684	4.6	
1,2-Dichlorobenzene	1.613	1.696	1.634	1.702	1.577	1.710	1.655	3.3	
1,2-Dichloroethane-d4	0.953	0.910	0.930	0.932	0.935		0.932	1.6	
Dibromofluoromethane	0.359	0.355	0.364	0.368	0.354		0.360	1.7	
Toluene-d8	1.246	1.223	1.266	1.275	1.221		1.246	2	
4-Bromofluorobenzene	0.455	0.470	0.500	0.500	0.464		0.478	4.4	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.